

A Prospective, Single-Arm Multi-Center Study of the HARMONIC® 1100 Shears and Generator G11 in Pediatric Surgical Procedures (General) and Adult Surgical Procedures (General, Gynecological, Urological, and Thoracic)

Trial Number: ENG_2020_06

Document	Date
Original	16 FEB 2021
Amendment 1	23 SEP 2021
Amendment 2	10 MAY 2022
Amendment 3	17 MAY 2023

Sponsor: Ethicon Endo-Surgery, Inc.
4545 Creek Road
Cincinnati, Ohio 45242

Name of Finished Product(s): HARMONIC® 1100 Shears
Generator G11 (GEN11)

Regulatory Classifications: HARMONIC 1100 Shears: Class II in the United States and Canada, Class IIb in the European Union and United Kingdom

GEN11: Class II in the United States and Class III in Canada, Class IIb in the European Union and United Kingdom

Sponsor's Medical Monitor: Gustavo Scapini, M.D.
Director, Medical Affairs
4545 Creek Rd.
Cincinnati, OH 45242

This study will be performed in compliance with Good Clinical Practice (GCP), and in accordance with the Declaration of Helsinki, as well as any other applicable local and country regulatory requirements.

CONFIDENTIALITY STATEMENT

The information in this document contains trade secrets and commercial information that are privileged or confidential and may not be disclosed unless such disclosure is required by federal or state law or regulations. Subject to the foregoing, this information may be disclosed only to those persons involved in the study who have a need to know, but all such persons must be instructed not to further disseminate this information to others. These restrictions on disclosure will apply equally to all future information supplied to you which is indicated as privileged or confidential.

PROTOCOL SIGNATURE PAGE

A Prospective, Single-Arm Multi-Center Study of the HARMONIC® 1100 Shears and Generator G11 in Pediatric Surgical Procedures (General) and Adult Surgical Procedures (General, Gynecological, Urological, and Thoracic)

Trial Number: ENG_2020_06

Document	Date
Original	16 FEB 2021
Amendment 1	23 SEP 2021
Amendment 2	10 MAY 2022
Amendment 3	17 MAY 2023

Approvals:

KATHRINE LORENZ

Digitally signed by KATHRINE LORENZ
DN: c=US, o=JNJ, ou=Subscribers,
0.9.2342.19200300.100.1.1=7354, cn=KATHRINE LORENZ
Reason: I am approving this document.
Date: 2023.06.16 11:07:38 -04'00'
Adobe Acrobat version: 2020.013.20064

Franchise Clinical Head
Ethicon Endo-Surgery, Inc.

Date

This study will be performed in compliance with GCP, and in accordance with the Declaration of Helsinki, as well as any other applicable local and country regulatory requirements.

INVESTIGATOR SIGNATURE PAGE

I have read, understood, and agree to:

- Ensure that the requirements for obtaining informed consent are met;
- Conduct the study in accordance with this protocol, including applicable local laws and regulations;
- Maintain the confidentiality of all information received or developed in connection with this protocol;
- Report all serious adverse events (SAEs) as soon as possible, but no later than 72 hours after becoming aware of the event;
- Adhere to the publication policy, as stated in the Clinical Study Agreement, for data collected during this study;
- Ensure that all associates, colleagues, and employees assisting in the conduct of the study are informed of their obligations in meeting the above commitments; and
- Provide copies of the protocol and all pertinent information to all individuals responsible to me who assist in the conduct of this study. I will discuss this material with them to ensure they are fully informed regarding the conduct of the study.

I will ensure that the Institutional Review Board (IRB) / Ethics Committee (EC) review complies with governmental requirements and will be responsible for the initial and continuing review and approval of the clinical investigation. I also agree to promptly report to the IRB/EC all changes in the research activity and all unanticipated problems involving risks to human subjects or others. Additionally, I will not make any changes in the research without Sponsor and IRB/EC approval of an amended protocol, except where necessary to eliminate apparent immediate hazards to human subjects.

I agree to comply with all other requirements regarding the obligation of clinical Investigators and all other pertinent requirements of the Sponsor and government agencies.

Signature of Principal Investigator (PI)

Date

Printed Name of Principal Investigator

Site Identification Number

CLINICAL STUDY PROTOCOL SYNOPSIS

Regulatory Classification:	<p>HARMONIC 1100 Shears: Class II in the United States and Canada, Class IIb in the European Union and United Kingdom</p> <p>GEN11: Class II in the United States and Class III in Canada, Class IIb in the European Union and United Kingdom</p>
Indication:	<p><u>HARMONIC® 1100 Shears</u></p> <p>The HARMONIC 1100 Shears are indicated for soft tissue incisions when bleeding control and minimal thermal injury are desired. The instruments can be used as an adjunct to or substitute for electrosurgery, lasers and steel scalpels in general, pediatric, gynecologic, urologic, thoracic procedures, and sealing and transection of lymphatic vessels. The instruments allow for the coagulation of vessels up to and including 7 mm in diameter, using the Energy button with Advanced Hemostasis.</p> <p>The instruments are not indicated for incising bone and are not intended for contraceptive tubal occlusion.</p> <p><u>Generator G11 (GEN11)</u></p> <p>The GEN11 provides radiofrequency power to drive ENSEAL electrosurgical instruments that are used during open or laparoscopic general and gynecological surgery to cut and seal vessels and to cut, grasp, and dissect tissues. In addition, the generator provides power to drive HARMONIC ultrasonic surgical instruments that are indicated for soft tissue incisions when bleeding control and minimal thermal injury are desired.</p> <p>ENSEAL and HARMONIC instruments, when used with the GEN11, have not been shown to be effective for sterilization procedures or tubal coagulation.</p>
Objective(s):	<p>The primary objective of this study is to demonstrate the acceptable performance and safety of the HARMONIC 1100 Shears and GEN11 devices when used per the instructions for use (IFU) in pediatric surgical procedures (general) or adult surgical procedures (general, gynecological, urological, and thoracic) where at least one single vessel is planned to be transected by the HARMONIC 1100 Shears.</p> <p>The secondary objective of this study is to demonstrate the acceptable performance and safety of the HARMONIC 1100 Shears in the transection of > 5 to 7 mm diameter size vessels.</p>
Overview of Study Design:	<p>This prospective, single-arm, multi-center study will collect clinical data in a post-market setting for the pediatric population (general surgical procedures) and adult population (general, gynecological, urological, and thoracic surgical procedures). Investigators will perform each procedure using the device in compliance with their standard surgical approach and the HARMONIC 1100 Shears and GEN11 IFUs.</p> <p>Subjects will be consented and screened anytime during a period of 8 weeks prior to the date of surgery. Subjects will be considered enrolled</p>

	when the HARMONIC 1100 Shears has been attempted to be used for at least one single vessel transection during pediatric surgical procedures (general) and adult surgical procedures (general, gynecological, urological, and thoracic). All enrolled subjects will be followed post-operatively through discharge and again at 28 days (± 14 days) post-surgery; therefore, from the surgery date to study exit, the duration will be approximately 6 weeks.																											
Number of Subjects:	A minimum of 165 subjects enrolled (maximum of 328 subjects enrolled) will be included in the study from up to 15 surgery centers globally with the following procedure targets:																											
	<table><tr><th>Procedure Group</th><th>Procedures</th><th>Estimated Initial Enrollment Range (minimum is 33 per procedure group)</th></tr><tr><td>1</td><td>Pediatric</td><td>33 – 40</td></tr><tr><td>2</td><td>Adult General (contain the following subgroups)</td><td>33 – 119</td></tr><tr><td>2a</td><td>Adult hepato-pancreato-biliary (HPB) (e.g. liver resection)</td><td>11 – 15</td></tr><tr><td>2b</td><td>Adult lower gastrointestinal (GI) (e.g., large intestine resection)</td><td>11 – 15</td></tr><tr><td>2c</td><td>Adult gastric (gastrectomy [e.g., subtotal, total], esophagectomy, Roux-en-Y gastric bypass [RYGB], sleeve gastrectomy, or biliopancreatic diversion with duodenal switch [BPD/DS])</td><td>11 – 89</td></tr><tr><td>3</td><td>Adult gynecological (e.g. hysterectomy, oophorectomy, myomectomy or endometriosis resection)</td><td>33 – 40</td></tr><tr><td>4</td><td>Adult urological (e.g., nephrectomy, cystectomy or prostatectomy)</td><td>33 – 40</td></tr><tr><td>5</td><td>Adult thoracic (lung resection [branches, excluding main pulmonary artery {PA}/pulmonary vein {PV}])</td><td>33 – 89</td></tr></table>	Procedure Group	Procedures	Estimated Initial Enrollment Range (minimum is 33 per procedure group)	1	Pediatric	33 – 40	2	Adult General (contain the following subgroups)	33 – 119	2a	Adult hepato-pancreato-biliary (HPB) (e.g. liver resection)	11 – 15	2b	Adult lower gastrointestinal (GI) (e.g., large intestine resection)	11 – 15	2c	Adult gastric (gastrectomy [e.g., subtotal, total], esophagectomy, Roux-en-Y gastric bypass [RYGB], sleeve gastrectomy, or biliopancreatic diversion with duodenal switch [BPD/DS])	11 – 89	3	Adult gynecological (e.g. hysterectomy, oophorectomy, myomectomy or endometriosis resection)	33 – 40	4	Adult urological (e.g., nephrectomy, cystectomy or prostatectomy)	33 – 40	5	Adult thoracic (lung resection [branches, excluding main pulmonary artery {PA}/pulmonary vein {PV}])	33 – 89
	Procedure Group	Procedures	Estimated Initial Enrollment Range (minimum is 33 per procedure group)																									
	1	Pediatric	33 – 40																									
	2	Adult General (contain the following subgroups)	33 – 119																									
	2a	Adult hepato-pancreato-biliary (HPB) (e.g. liver resection)	11 – 15																									
	2b	Adult lower gastrointestinal (GI) (e.g., large intestine resection)	11 – 15																									
	2c	Adult gastric (gastrectomy [e.g., subtotal, total], esophagectomy, Roux-en-Y gastric bypass [RYGB], sleeve gastrectomy, or biliopancreatic diversion with duodenal switch [BPD/DS])	11 – 89																									
	3	Adult gynecological (e.g. hysterectomy, oophorectomy, myomectomy or endometriosis resection)	33 – 40																									
	4	Adult urological (e.g., nephrectomy, cystectomy or prostatectomy)	33 – 40																									
5	Adult thoracic (lung resection [branches, excluding main pulmonary artery {PA}/pulmonary vein {PV}])	33 – 89																										
Every subject enrolled in the study would count toward the primary enrollment goal of 165 subjects. In addition, a secondary enrollment goal for the study is to enroll a sufficient number of subjects to achieve a minimum of 165 vessels transected with a diameter size of > 5 to 7 mm based on surgeon determination. Every subject enrolled in the study that also had a > 5 to 7 mm vessel transected would count toward the secondary enrollment goal in addition to the primary enrollment goal. The initial assumption is that PA branches (thoracic procedure group) and left gastric veins (gastric procedure group) will have the greatest likelihood of having a vessel with a diameter size of > 5 to 7 mm. The number of vessels transected with a diameter of > 5 to 7 mm will be continuously monitored by the Sponsor throughout the course of the study.																												

	<p>The table above provides the minimum number of subjects to be enrolled per procedure group and the Sponsor will determine the maximum number of subjects to be enrolled per procedure group based on fulfillment of secondary endpoint. Once the minimum number has been met for a procedure group, enrollment for that group could be closed to further enrollment.</p>
Criteria for Inclusion:	<p>Subjects satisfying the following criteria will be considered eligible for enrollment in this study:</p> <p>Pediatric population:</p> <ol style="list-style-type: none"> 1. Non-emergent procedure (general) where at least one vessel is planned to be transected by the HARMONIC 1100 Shears per the IFU; 2. Less than 18 years of age at the time of procedure; and 3. The subject's parent/legal guardian must be willing to give permission for the subject to participate in the study and provide written informed consent for the subject. In addition, assent must be obtained from pediatric subjects who possess the intellectual and emotional ability to comprehend the concepts involved in the study. If the pediatric subject is not able to provide assent (due to age, maturity and/or inability to intellectually and/or emotionally comprehend the study), the parent/legal guardian's written informed consent for the subject will be acceptable for the subject to be included in the study. <p>Adult population:</p> <ol style="list-style-type: none"> 1. Elective procedure (general, gynecological, urological, or thoracic) where at least one vessel is planned to be transected by the HARMONIC 1100 Shears per the IFU; 2. Willingness to give consent and comply with all study-related evaluations and visit schedule; and 3. At least 18 years of age.
Criteria for Exclusion:	<p>Subjects meeting any of the following criteria will be considered ineligible for enrollment in this study:</p> <p><u>Preoperative</u></p> <ol style="list-style-type: none"> 1. Physical or psychological condition which would impair study participation; 2. Female subjects, of childbearing age, who are pregnant; or 3. Enrollment in a concurrent interventional clinical study that could impact the study endpoints. <p><u>Intraoperative</u></p> <ol style="list-style-type: none"> 1. The HARMONIC 1100 Shears were not attempted to be used for at least one single vessel transection during the surgical procedure.

Test Product:	HARMONIC® 1100 Shears (Product Codes: HAR1120 or HAR1136) and Generator G11 (GEN11).
Reference Therapy/Product:	No comparator product is being used in this study.
Duration of Treatment:	Subjects will be exposed to the HARMONIC 1100 Shears and GEN11 during the procedure and while the devices are used as indicated in the IFU. Neither the devices nor any device components will remain in the patient after use.
Criteria for Evaluation:	<p>The primary performance endpoint will be:</p> <ul style="list-style-type: none"> Achievement of Grade 3 or lower hemostasis for each vessel transection (per the grading scale below): <ul style="list-style-type: none"> Grade 1: no bleeding at transection site; Grade 2: minor bleeding at transection site, no intervention needed; Grade 3: minor bleeding at transection site, mild intervention needed, use of compression, basic energy devices (monopolar and/or bipolar) and/or touch-ups with HARMONIC 1100 Shears; or Grade 4: significant bleeding (e.g., pulsatile blood flow, venous pooling) requiring intervention such as extensive coagulation or ligation with use of additional hemostatic measures (e.g., hemoclips, staples, sutures, fibrin sealants, other advanced energy products); <p>The secondary performance endpoints will be:</p> <ul style="list-style-type: none"> Distribution of 5-point scale (very dissatisfied, dissatisfied, neither satisfied or dissatisfied, satisfied, or very satisfied) scores for sealing and transection of lymphatic vessels completed by the HARMONIC 1100 Shears; Distribution of hemostasis grading scale (1-4) for each vessel transected; and Type, name, and number of additional hemostasis measures required to achieve hemostasis for Grade 4 vessel transections. <p>The primary safety endpoint will be:</p> <ul style="list-style-type: none"> Occurrence of device-related (possible, probable or causal) adverse events (AEs).

Statistical Analysis:

Analysis Sets

The summary of all performance and safety endpoints will be performed on the set of subjects in whom the HARMONIC 1100 Shears is utilized during the surgical procedure. The summary of all performance and safety endpoints will be performed by surgical population (pediatric and adult), surgical procedure group, procedure subgroup, and on the entire pooled set of subjects.

Sample Size Determination

Primary performance analysis: A sample size of at least 165 subjects is required to have a minimum of 80% power for rejecting the null hypothesis that the rate of Grade 3 or lower transections is less than or equal to 87.5% when the expected rate of Grade 3 hemostatic transections is at least 94.0% using exact binomial test and a one-sided significance level of 0.025. The sample size estimation makes the assumption of one transection per subject which is likely very conservative as most subjects may have 2 or more transections during their procedure. At the time of study protocol amendment 3 it was found that on average there were 2 vessel transections per subject. A minimum of 33 subjects will be enrolled into each of the five individual procedure groups.

Secondary performance analysis: A sample size of at least 165 vessels is required to have a minimum of 80% power for rejecting the null hypothesis that the rate of Grade 3 or lower transections is less than or equal to 87.5% when the expected rate of Grade 3 hemostatic transections is at least 94.0% for vessels with diameter size of > 5 to 7 mm. Sample size estimation considered an exact binomial test and a one-sided level of significance of 0.025. The number of vessels transected with a diameter size of > 5 to 7 mm will be continuously monitored by the Sponsor throughout the course of the study to achieve the required number of vessels.

Performance Analyses

The number and percentage of vessels where hemostasis is achieved (\leq Grade 3) will be summarized and an exact 95% confidence interval based on Clopper-Pearson's method will be estimated for each procedure group (pediatrics [general surgical procedures] and adult [general, gynecological, urological, and thoracic surgical procedures]), the three individual procedure subgroups, the > 5 to 7 mm diameter vessel population, and in total. The performance goal hypothesis will be evaluated on the total population and the > 5 to 7 mm diameter vessel population using the lower bound of the exact two-sided 95% confidence interval based on Clopper-Pearson's method and a p-value from exact binomial test. Counts and percentages will be provided for type, size, and number of vessels transected, grading scale distribution for all vessels transected, number of times HARMONIC 1100 Shears touch-ups were required, and incidence of requirement for additional measures to obtain hemostasis on vessels (other advanced energy devices or hemostatic measures).

Safety Analyses

All device-related and procedure-related AEs reported during the study will be coded to the Medical Dictionary for Regulatory Activities (MedDRA). All AEs will be summarized by MedDRA system organ class and preferred term by subject population (pediatric and adult), procedure group, procedure subgroup and in total. Separate summaries will be provided for device-related and procedure-related AEs. Serious AEs will be summarized in a similar manner.

TABLE OF CONTENTS

	Page
PROTOCOL SIGNATURE PAGE	2
INVESTIGATOR SIGNATURE PAGE	3
CLINICAL STUDY PROTOCOL SYNOPSIS	4
PROTOCOL-SPECIFIC ACRONYMS AND ABBREVIATIONS	12
ETHICS	13
ADMINISTRATIVE REQUIREMENTS	14
1.0 INTRODUCTION	15
2.0 STUDY OBJECTIVES	15
2.1 Primary Performance Endpoint	16
2.2 Secondary Performance Endpoints	16
2.3 Safety Endpoint	16
2.4 Additional Key Data Collected	16
3.0 INVESTIGATIONAL PLAN	17
3.1 OVERALL STUDY DESIGN AND PLAN - DESCRIPTION	17
3.2 STUDY POPULATION	17
3.2.1 Enrollment	17
3.2.2 Screening Failures	18
3.2.3 Inclusion Criteria	18
3.2.4 Exclusion Criteria	19
3.2.5 Removal of Subjects from Study	19
3.2.6 Subject Replacement in Study	20
3.3 STUDY PROCEDURES	20
3.3.1 Procedure Description	20
3.3.2 Identity of Study Products	20
3.3.2.1 Device Description	21
3.3.2.2 Indications	21
3.3.2.3 Contraindications	21
3.3.2.4 Labeling of Study Products	22
3.3.2.5 Accountability of Study Products	22
3.3.3 Prior and Concomitant Therapy	22
3.4 STUDY VARIABLES	22
3.4.1 Demographic and Baseline Characteristics	22
3.4.2 Study Variables	23
3.4.3 Questionnaires	24
3.4.4 Safety Data Variables	25
3.5 SCHEDULE OF EVENTS	25
3.6 STUDY PROCEDURES	26
3.6.1 Visit 1 – Screening (may occur over visits within 8 weeks of Visit 2 including up to the day of the procedure)	26
3.6.2 Visit 2 – Procedure Through Discharge	26
3.6.3 Visit 3 – Post-procedure Follow-up	26
3.6.4 Unscheduled Visit(s)	26
3.6.5 Vessel Image Capture	26
4.0 DATA MANAGEMENT AND INTEGRITY	27
4.1 DATA COMPLETION AND RECORD KEEPING	27
4.1.1 Source Documents	27
4.1.2 Electronic Data Capture	27
4.1.3 Data Collection	27

4.1.4	Data Correction	28
4.1.5	Data Privacy	28
4.1.6	Record Retention, Inspection, and Custody	28
4.2	MEDICAL DICTIONARY CODING	28
4.3	DATA QUALITY ASSURANCE	28
4.3.1	Site Personnel Training	28
4.3.2	Monitoring	28
4.3.3	Quality Assurance	29
4.4	PROTOCOL DEVIATIONS	29
5.0	STATISTICAL METHODS PLANNED IN THE PROTOCOL AND DETERMINATION OF SAMPLE SIZE	29
5.1	STUDY DESIGN	29
5.2	TREATMENT ASSIGNMENT	29
5.3	INTERVAL WINDOWS	29
5.4	PRIMARY AND SECONDARY ENDPOINTS AND ASSOCIATED HYPOTHESES	30
5.5	LEVELS OF SIGNIFICANCE	30
5.6	ANALYSIS SETS	30
5.7	SAMPLE SIZE JUSTIFICATION	30
5.8	DATA MONITORING COMMITTEE	31
5.9	ANALYSES TO BE CONDUCTED	31
5.9.1	General Conventions	31
5.9.2	Disposition of Study Subjects	32
5.9.3	Demographic, Baseline, and Surgical Characteristics	32
5.9.4	Primary and Secondary Endpoint Analyses	32
5.9.5	Plans for Interim Analyses	32
5.9.6	Handling of Missing Data	32
5.9.7	Subgroup Analyses	32
6.0	ADVERSE EVENTS	33
6.1	DEFINITIONS	33
6.1.1	Adverse Event	33
6.1.2	Serious Adverse Event	33
6.1.3	Pre-existing Condition	33
6.2	SEVERITY OF ADVERSE EVENTS	33
6.3	RELATIONSHIP OF ADVERSE EVENTS	34
6.4	REPORTING PROCEDURES FOR ADVERSE EVENTS	35
6.4.1	Recording Adverse Events	35
6.4.2	Reporting Adverse Events	35
6.4.3	Reporting Serious Adverse Events	35
7.0	PRODUCT COMPLAINTS	36
7.1	REPORTING PRODUCT COMPLAINTS FOR MARKETED DEVICES	36
7.1.1	Reporting of Approved EES Marketed Devices	36
7.1.2	Reporting of Approved Marketed Devices and Commercially Available Devices (not Ethicon)	36
8.0	REFERENCES	36

PROTOCOL-SPECIFIC ACRONYMS AND ABBREVIATIONS

Acronyms/Abbreviations	Terms
AE	Adverse event
ASA	American Society of Anesthesiologists Physical Status Classification System
BPD/DS	Biliopancreatic diversion with duodenal switch
EC	Ethics committee
eCRF	Electronic case report form
EDC	Electronic Data Capture
EEPROM	Electrically erasable programmable read-only memory
EES	Ethicon Endo-Surgery, Inc.
GEN11	Ethicon Endo-Surgery Generator G11
FDA	Food and Drug Administration
GCP	Good clinical practice
GI	Gastrointestinal
HPB	Hepato-pancreato-biliary
ICD	Informed consent document
IFU	Instructions for use
IRB	Institutional review board
MedDRA	Medical Dictionary for Regulatory Activities
OR	Operating room
PA	Pulmonary artery
PI	Principal Investigator
PV	Pulmonary vein
PCF	Product complaint form
PHI	Protected health information
PMCF	Post-market clinical follow-up
RYGB	Roux-en-Y gastric bypass
SAE	Serious adverse event
SAP	Statistical analysis plan
SOC	Standard of care

ETHICS

Institutional Review Board/Ethics Committee

Participating investigators will ensure that this protocol, Informed Consent Document (ICD), ICD or protocol amendments, and if applicable, any other written information provided to the subjects that assist in the decision to participate are reviewed by an Institutional Review Board (IRB) or Ethics Committee (EC) that complies with governmental requirements. The approving IRB/EC will be responsible for the initial and continuing review and approval of this clinical investigation. Participating investigators will be required to promptly report to the IRB/EC as required by the IRB/EC's policies. Additionally, investigators will be required to refrain from making any changes in the clinical investigation plan without Sponsor and IRB/EC approval of an amended protocol, except where necessary to eliminate apparent immediate hazards to study subjects or others.

Applicable Regulations

This study will be conducted in compliance with GCP and in accordance with the Declaration of Helsinki, as well as any other applicable local and country regulatory requirements.

Subject Information and Consent

Regulations concerning the protection of subjects require that informed consent be obtained before a subject may participate in any clinical investigation. An IRB/EC approved informed consent must be sought from each subject and must be appropriately documented in the subject's medical record prior to initiating the study. It is the Investigator's responsibility to obtain written informed consent from the subject, the Investigator may delegate this responsibility if appropriately documented. In addition, assent must be obtained from pediatric subjects who possess the intellectual and emotional ability to comprehend the concepts involved in the study. If the pediatric subject is not able to provide assent (due to age, maturity and/or inability to intellectually and/or emotionally comprehend the trial), the parent/legal guardian's written informed consent for the subject will be acceptable for the subject to be included in the study.

The informed consent process involves the following: giving a subject adequate information concerning the study, providing adequate time for the subject to consider all available options, responding to the subject's questions, ensuring that the subject has comprehended this information and finally, obtaining the subject's written consent to participate in this study. All subjects in this study should be completely informed about the purpose, risks, benefits, and other pertinent details of this study. The informed consent process is careful to avoid the perception of any coercion or undue influence on, or inducement of, the subject to participate, and does not waive or appear to waive the subject's legal rights. The ICD is presented in native, non-technical language that is understandable to the subject.

Prior to a subject's participation in this study, an ICD will be signed and dated by the subject and person who conducted the consent discussion. The subject will be provided a copy of the signed ICD. The ICD and any other written materials provided to the subject to assist in the decision to participate must be revised whenever new information becomes available that may be relevant to their willingness to participate or continue participation in this study. Revision to the ICD and other written materials will receive IRB/EC approval before implementation. Each subject will be required to sign any amended ICD (as required by the IRB/EC) and will receive a copy of the signed ICD.

ADMINISTRATIVE REQUIREMENTS

This study is sponsored by Ethicon Endo-Surgery, Inc. (EES, Cincinnati, OH, USA) and will be conducted in up to fifteen surgery centers globally under a single protocol approved by each participating site's IRB/EC prior to implementation. The PI at each study site is a surgeon qualified by education, experience, and training to perform the study procedure and to assume responsibility for the conduct of this study.

The Data Management and Biostatistics groups of the Sponsor will be responsible for the analysis of data from this protocol. An Electronic Data Capture (EDC) system will be utilized by study site personnel to transfer study data from source records (the first point of clinical data capture) onto common electronic case report forms (eCRFs).

Protocol Modifications

All protocol amendments must be issued by the Sponsor, signed and dated by the Investigator, and should not be implemented without prior IRB/EC approval, except where necessary to eliminate immediate hazards to the subjects or when the change(s) involves only logistical or administrative aspects of the study (e.g., change in monitor(s), change of telephone number(s)). The Investigator reports the protocol amendments to the IRB/EC as per their local requirements.

1.0 INTRODUCTION

The HARMONIC 1100 Shears is a sterile, single-patient-use instrument used for dissection, grasping, coagulation, and cutting between the blade and clamp arm. It can be used in general, pediatric, gynecologic, urologic, thoracic, and sealing and transection of lymphatic vessels. The instruments allow for the coagulation of vessels up to and including 7 mm in diameter.

The HARMONIC 1100 Shears differs primarily in two ways from the earlier version of the device, the HARMONIC HD 1000i Shears:

1. An electrically erasable programmable read-only memory (EEPROM) with an improved algorithm that actively controls blade heat for lower maximum blade temperature compared to previous HARMONIC devices; and
2. Branding updates to the device handle and distal shaft.

The HARMONIC 1100 Shears have undergone extensive preclinical testing to ensure that they are capable of producing strong and durable seals for vessels up to 7mm in diameter. Figure 1 shows a picture of the HARMONIC 1100 Shears. The device will be available in shaft lengths of 20 and 36 cm. The instrument shaft can be rotated continuously to facilitate visualization and access to targeted tissue.



Figure 1: Image of the HARMONIC 1100 Shears.

The GEN11 supplies energy to the HARMONIC and ENSEAL surgical instruments. It also has a touchscreen for easy setup and operation, and a high-resolution display with wider viewing angles. Software updates to ensure that the system is operating optimally are available via USB memory stick. The generator uses a touchscreen display interface and has a unique receptable port that accepts either a HARMONIC or an ENSEAL instrument. Connectors (HGA11 for HARMONIC and EGA11 for ENSEAL) are used to enable the generator to power legacy instruments.

This study is being conducted to evaluate the devices when used per their instructions for use (IFU) in the following types of procedures: pediatric surgical procedures (general) and adult surgical procedures (general, gynecological, urological, and thoracic).

2.0 STUDY OBJECTIVES

The primary objective of this study is to demonstrate the acceptable performance and safety of the HARMONIC 1100 Shears and GEN11 devices when used per the IFUs in pediatric surgical

procedures (general) or adult surgical procedures (general, gynecological, urological, and thoracic) where at least one single vessel is planned to be transected by the HARMONIC 1100 Shears.

The secondary objective of this study is to demonstrate the acceptable performance and safety of the HARMONIC 1100 Shears in the transection of > 5 to 7 mm diameter size vessels.

2.1 PRIMARY PERFORMANCE ENDPOINT

A hemostasis grading scale was adapted from Siegel et al.¹ The primary performance endpoint will be:

- Achievement of Grade 3 or lower hemostasis for each vessel transection (per the grading scale below):
 - Grade 1: no bleeding at transection site;
 - Grade 2: minor bleeding at transection site, no intervention needed;
 - Grade 3: minor bleeding at transection site, mild intervention needed, use of compression, basic energy devices (monopolar and/or bipolar) and/or touch-ups with HARMONIC 1100 Shears; or
 - Grade 4: significant bleeding (e.g., pulsatile blood flow, venous pooling) requiring intervention such as extensive coagulation or ligation with use of additional hemostatic measures (e.g., hemoclips, staples, sutures, fibrin sealants, other advanced energy products).

2.2 SECONDARY PERFORMANCE ENDPOINTS

The secondary performance endpoints will be:

- Distribution of 5-point scale scores for sealing and transection of lymphatic vessels completed by the HARMONIC 1100 Shears. The scale will be a 5-point scale (very dissatisfied, dissatisfied, neither satisfied or dissatisfied, satisfied, or very satisfied) asking how satisfied the surgeon was with the use of the device for the task;
- Distribution of hemostasis grading scale (1-4) for each vessel transected; and
- Type, name, and number of additional hemostasis measures required to achieve hemostasis for Grade 4 vessel transections.

2.3 SAFETY ENDPOINT

The primary safety endpoint will be occurrence of device-related (possible, probable or causal) adverse events (AEs).

2.4 ADDITIONAL KEY DATA COLLECTED

Additional key data collected for evaluation in this study are:

- Procedure duration;
- Use of any other energy device (monopolar, traditional bipolar, advanced bipolar, ultrasonic) in primary procedure (type, name, and number of device and reason for use), if applicable;
- Surgeon questionnaire administered once per investigator;
- Generator questionnaire after each procedure for each GEN11 used;
- Surgical procedure conducted;
- Hospital stay duration;

- Name and number of vessels that were transected, surgeon determination of diameter size range (< 3 mm, 3 to 5 mm, and > 5 to 7 mm);
- Occurrence and location of cancer and occurrence of pre-surgical radiation/chemotherapy within 90 days prior to surgery, if applicable;
- Occurrence of vessel skeletonization;
- Presence of inflamed tissue, calcified tissues/vessels, atherosclerotic tissue, fibrotic tissue, or presence of adhesions intraoperatively, if applicable;
- Volume of estimated intra-operative blood loss;
- Occurrence of blood transfusion, if applicable (record the total required units of blood and rationale);
- Type of additional mild interventions (use of compression, monopolar device, bipolar device, and/or touch-ups with HARMONIC 1100 Shears) required to achieve hemostasis for Grade 3 vessel transections including number of times when HARMONIC 1100 touch-ups were used; and
- Procedure-related AEs (possible, probable or causal).

3.0 INVESTIGATIONAL PLAN

3.1 OVERALL STUDY DESIGN AND PLAN - DESCRIPTION

This prospective, single-arm, multi-center study will collect clinical data in a post-market setting for the pediatric population (general surgical procedures) and adult population (general, gynecological, urological, and thoracic surgical procedures). Investigators will perform each procedure using the device in compliance with their standard surgical approach and the HARMONIC 1100 Shears and GEN11 IFUs.

Subjects will be consented and screened anytime during a period of 8 weeks prior to the date of surgery. Subjects will be considered enrolled when the HARMONIC 1100 Shears has been attempted to be used for at least one single vessel transection during pediatric surgical procedures (general) and adult surgical procedures (general, gynecological, urological, and thoracic). All subjects enrolled will be followed post-operatively through discharge and again at 28 days (\pm 14 days) post-surgery; therefore, from the surgery date to study exit, the duration will be approximately 6 weeks.

3.2 STUDY POPULATION

3.2.1 Enrollment

Subjects will be recruited from the investigators' existing patient population who plan to have a procedure from the proposed procedure groups utilizing consecutive screening and enrollment in an effort to generate a random and representative patient population sample. All eligible subjects (Sections 3.2.3 and 3.2.4) will be considered enrolled at the time when the HARMONIC 1100 Shears has been attempted to be used for at least one single vessel transection during one of the proposed procedure groups. Up to 15 surgery centers globally will be selected as study sites. Each participating investigator is expected to perform a minimum of two procedures using the HARMONIC 1100 Shears. A minimum of 165 subjects enrolled (maximum of 328 subjects enrolled) will be included in the study with the following procedure targets:

Table 1: Distribution of Targeted Procedures

Procedure Group	Procedures	Estimated Initial Enrollment Range (minimum is 33 per procedure group)
1	Pediatric	33 – 40
2	Adult General (contain the following subgroups)	33 – 119
2a	Adult HPB (e.g. liver resection)	11 – 15
2b	Adult lower GI (e.g., large intestine resection)	11 – 15
2c	Adult gastric (gastrectomy [e.g., subtotal, total], esophagectomy, RYGB, sleeve gastrectomy, or BPD/DS)	11 – 89
3	Adult gynecological (e.g., hysterectomy, oophorectomy, myomectomy or endometriosis resection)	33 – 40
4	Adult urological (e.g., nephrectomy, cystectomy or prostatectomy)	33 – 40
5	Adult thoracic (lung resection [branches, excluding main PA/PV])	33 – 89

Every subject enrolled in the study would count toward the primary enrollment goal of 165 subjects. In addition, a secondary enrollment goal for the study is to enroll a sufficient number of subjects to achieve a minimum of 165 vessels transected with a diameter size of > 5 to 7 mm based on surgeon determination. Each discreet vessel transected measuring > 5 to 7 mm in diameter (per surgeon assessment) would count toward the secondary enrollment goal in addition to the primary enrollment. The initial assumption is that PA branches (thoracic procedure group) and left gastric veins (gastric procedure group) will have the greatest likelihood of having a vessel with a diameter size of > 5 to 7 mm. The number of vessels transected with a diameter of > 5 to 7 mm will be continuously checked throughout the course of the study by surgeon determination (see Section 3.4.2).

The table above provides the minimum number of subjects to be enrolled per procedure group and the Sponsor will determine the maximum number of subjects to be enrolled per procedure group. Once the minimum number has been met for a procedure group, enrollment for that group could be closed to further enrollment.

3.2.2 Screening Failures

All subjects signing consent who do not have at least one single vessel transection attempted by HARMONIC 1100 Shears during their procedures will be recorded as screen failures. For subjects who are determined to be screen failures, only the relevant eCRF pages (inclusion/exclusion criteria, demographics, AEs, subject completion/discontinuation) will be completed.

3.2.3 Inclusion Criteria

Subjects satisfying the following criteria will be considered eligible for enrollment in this study:

Pediatric population:

1. Non-emergent procedure (general) where at least one vessel is planned to be

transected by the HARMONIC 1100 Shears per the IFU;

2. Less than 18 years of age at the time of procedure; and
3. The subject's parent/legal guardian must be willing to give permission for the subject to participate in the study and provide written informed consent for the subject. In addition, assent must be obtained from pediatric subjects who possess the intellectual and emotional ability to comprehend the concepts involved in the study. If the pediatric subject is not able to provide assent (due to age, maturity and/or inability to intellectually and/or emotionally comprehend the study), the parent/legal guardian's written informed consent for the subject will be acceptable for the subject to be included in the study.

Adult population:

1. Elective procedure (general, gynecological, urological, or thoracic) where at least one vessel is planned to be transected by the HARMONIC 1100 Shears per the IFU;
2. Willingness to give consent and comply with all study-related evaluations and visit schedule; and
3. At least 18 years of age.

3.2.4 Exclusion Criteria

Subjects meeting the following criteria will be considered ineligible for enrollment in this study:

Preoperative

1. Physical or psychological condition which would impair study participation; or
2. Female subjects, of childbearing age, who are pregnant; or
3. Enrollment in a concurrent interventional clinical study that could impact the study endpoints.

Intraoperative

1. The HARMONIC 1100 Shears were not attempted to be used for at least one single vessel transection during the surgical procedure.

3.2.5 Removal of Subjects from Study

A subject has the right to withdraw from the study at any time for any reason without prejudice to his/her future medical care by the physician or the institution. Should a subject (or subject's legally authorized guardian/representative) decide to withdraw, 1) all data collected up to the point of withdrawal will be considered for analysis; and 2) all efforts will be made to collect and report the final visit observations as thoroughly and timely as possible. Participation may be terminated prior to completing the study for any of the reasons listed below (reasons that do not fit the categories below will be documented as "other").

Adverse Event

When the subject experiences an AE and the Investigator or Medical Monitor believes it is in the best interest of the subject to discontinue participation in the study, the subject will be withdrawn from the study.

Lost to follow-up

When contact with the subject has been lost without completing a final contact assessment, and every attempt to contact the subject has failed, the subject will be considered lost to follow-up. All attempts to contact the subject requesting his/her return for the final visit must be documented.

Withdrawal of consent

Any method of contact with the subject (or subject's legally authorized guardian/representative) in which he/she state they no longer want to participate in the study specific activities constitutes withdrawal of consent for participation in the study. When possible, the reason for withdrawal will be documented.

Site Termination or Study Termination

A study site or the entire study may be terminated. When this occurs, all subjects currently enrolled at the site will be withdrawn and documented as early terminations. Reasons for site or study termination may include, but are not limited to the following:

- Administrative concerns (e.g., inadequate subject enrollment, Investigator/institution non-compliance, change of business strategy, etc.);
- Safety Issues, including those due to non-compliance, which substantially affect the risk to benefit ratio of the study subjects at a site or for the study as a whole; or
- Regulatory Body Mandate(s).

The Investigator has the right to terminate the subject's participation at any time. Should this be necessary, procedures for termination will be provided by the Sponsor.

Death

- When possible, the cause of death will be documented.

Other (which may include)

- Investigator recommendation.

3.2.6 Subject Replacement in Study

Subjects who withdraw or are terminated early from the study will not be replaced.

3.3 STUDY PROCEDURES

3.3.1 Procedure Description

Elective pediatric surgical procedures (general) or adult surgical procedures (general, gynecological, urological, and thoracic) where HARMONIC 1100 is indicated for use will be performed per the institution's SOC. Only the PI or Sub-Investigator can use the device during the vessel transections. See Section 3.2.1 for a list of possible specific procedures.

3.3.2 Identity of Study Products

The HARMONIC 1100 Shears and the GEN11 are the study devices for this study.

3.3.2.1 Device Description**HARMONIC 1100 Shears**

The HARMONIC 1100 Shears instrument is a sterile, single-patient-use instrument used for dissection, grasping, coagulation, and cutting between the blade and clamp arm. It consists of an ergonomic handle with an integrated hand piece and two energy delivery buttons. The instrument shaft can be rotated continuously to facilitate visualization and access to targeted tissue. The Energy button is indicated for vessels up to 5 mm in diameter. When the Energy button is used, cutting speed is the fastest. The Energy button with Advanced Hemostasis is designed for larger vessels and is indicated for vessels up to 7 mm in diameter. When the Energy button with Advanced Hemostasis is used, cutting speed is reduced and hemostasis is maximized. The instrument utilizes Adaptive Tissue Technology that is specific to the HARMONIC 1100 Shears. This provides the generator with the ability to identify and monitor the instrument during use, which enables the generator to modulate and adjust its power output as well as provide audible feedback to the user as appropriate. The HARMONIC 1100 Shears is designed for use exclusively with the GEN11 software version 2018-1 or later, packaged separately.

GEN11

The GEN11 device supplies energy to the HARMONIC and ENSEAL surgical instruments. The generator uses a touchscreen display interface and has a unique receptacle port that accepts either a HARMONIC or an ENSEAL instrument. Connectors (HGA11 for HARMONIC and EGA11 for ENSEAL) are used to enable the generator to power legacy instruments.

3.3.2.2 Indications**HARMONIC 1100 Shears**

The HARMONIC 1100 Shears instrument is indicated for soft tissue incisions when bleeding control and minimal thermal injury are desired. The instruments can be used as an adjunct to or substitute for electrosurgery, lasers and steel scalpels in general, pediatric, gynecologic, urologic, thoracic, and sealing and transection of lymphatic vessels. The instruments allow for the coagulation of vessels up to and including 7 mm in diameter, using the Energy button with Advanced Hemostasis.

GEN11

The GEN11 provides radiofrequency power to drive ENSEAL electrosurgical instruments that are used during open or laparoscopic general and gynecological surgery to cut and seal vessels and to cut, grasp, and dissect tissues. In addition, the generator provides power to drive HARMONIC ultrasonic surgical instruments that are indicated for soft tissue incisions when bleeding control and minimal thermal injury are desired.

ENSEAL and HARMONIC instruments, when used with the GEN11, have not been shown to be effective for sterilization procedures or tubal coagulation.

3.3.2.3 Contraindications**HARMONIC 1100 Shears**

The HARMONIC 1100 Shears are not indicated for incising bone and are not intended for contraceptive tubal occlusion.

GEN11

The use of the GEN11 and the attached instruments are contraindicated when, in the judgement of the physician, radiofrequency or ultrasonic surgery would be contrary to the best interest of the patient. The instruments are not indicated for incising bone.

3.3.2.4 Labeling of Study Products

The HARMONIC 1100 Shears has received 510(k) clearance in the United States by the Food and Drug Administration (FDA), has clearance in Canada by Health Canada and has CE Mark in the European Union and United Kingdom. The GEN11 has received 510(k) clearance in the United States by the FDA, has clearance in Canada by Health Canada and has CE Mark in the European Union and United Kingdom. The study devices will use the cleared labeling. The Sponsor requires no additional labeling in this study.

3.3.2.5 Accountability of Study Products

The HARMONIC 1100 Shears may be provided to the participating institutions. The GEN11 (including footswitches) may be loaned to any institution in the study. Each institution not provided with HARMONIC 1100 Shears will use devices acquired through normal procurement process. If institutions are provided HARMONIC 1100 Shears, the devices will be tracked using shipping receipts and device accountability logs and all device returns will be managed by contacting the Sponsor study team. Devices provided for the study must be kept in a secure area and used only for treating subjects participating in the study, in accordance with the protocol. If applicable, the study device inventory/accountability logs must be available for routine monitoring purposes as part of required Sponsor oversight, or inspection.

3.3.3 Prior and Concomitant Therapy

Subjects will be assessed for certain prior medication therapy used up to 30 days prior to surgery. These medications include aspirin, antiplatelet agents, nonsteroidal anti-inflammatory drugs, and anticoagulants. Review will occur during Visit 1 (Screening Visit) and Visit 2 (Procedure Visit).

Subjects will be assessed for occurrence of pre-surgical radiation/chemotherapy used up to 90 days prior to surgery. Review will occur during Visit 1 (Screening Visit) and Visit 2 (Procedure Visit).

Subjects may continue with their current medical care while in the study and the following medications will be collected for this study:

- Certain prior medication therapy used up to 30 days prior to surgery (listed above); and
- Any medications administered during the study due to a device-related or procedure-related AE.

3.4 STUDY VARIABLES

Specific variables assessed in the study are provided in the following sections. Refer to Section 3.5 Schedule of Events for the time when the study variables will be collected throughout the course of the study.

3.4.1 Demographic and Baseline Characteristics

The following will be collected preoperatively:

- Age (years);
- Gender;
- Race;
- Ethnicity;
- Administer pregnancy test as per standard of care;
- Body weight (kg or lbs; no shoes);
- Body height (cm or in; no shoes);
- American Society of Anesthesiologists Physical Status Classification System (ASA) score;
- Occurrence and location of cancer, if applicable;
- Review and collection of relevant medical history and surgical history including the following:
 - For the gynecological procedure group, gravidity and parity;
 - Surgical procedure to be conducted, including primary indication for surgery; and
 - Smoking history (current smoking status, if stopped, length of time [months and years]).

3.4.2 Study Variables

The following variables will be recorded preoperatively, intraoperatively, or postoperatively:

- Date/time of hospital admission;
- Procedure duration, defined as first skin incision to final skin closure;
- Did vessel skeletonization occur (Yes/No; if Yes, describe vessel skeletonized and device used);
- Was there presence of inflamed tissue, calcified tissues/vessels, atherosclerotic tissue, fibrotic tissue, or presence of adhesions intraoperatively, if applicable (Yes/No; if Yes, describe the tissue along with location)?;
- Was there sealing and transection of lymphatic vessels by HARMONIC 1100 Shears (Yes/No; if Yes, describe lymphatics and tissue bundles along with location and complete 5-point scale)?;
- Conversion to open with reason, if applicable;
- Volume of estimated intra-operative blood loss;
- Use of any other energy device (monopolar, traditional bipolar, advanced bipolar, ultrasonic) in primary procedure (type, name, and number of device and reason for use), if applicable;
- Name, product code, lot #, and number used of trocars used with HARMONIC 1100 Shears, if applicable;
- Was there prophylactic use of clips or sutures as standard of surgical care before vessel transection (Yes/No; if Yes, name vessel transected, and number of clips or

sutures prophylactically applied);

- For each vessel transected with HARMONIC 1100 Shears:
 - Name of each vessel;
 - Hemostasis grading scale assessment (1-4);
 - Surgeon determination of diameter size range (< 3 mm, 3 to 5 mm, and > 5 to 7 mm);
 - Image captured where the vessel is perpendicular and below the open jaw of the HARMONIC 1100 Shears;
 - For every Grade 3 vessel transection:
 - Compression used as a mild intervention, if applicable (Yes/No);
 - Touch-ups with a monopolar device, if applicable (Yes/No; if Yes, record number of touch-ups);
 - Touch-ups with a bipolar device, if applicable (Yes/No; if Yes, record number of touch-ups);
 - Touch-ups with HARMONIC 1100 (Yes/No; if Yes, record number of touch-ups);
 - For every Grade 4 vessel transection intervention:
 - Type, name, and number of times used for each additional hemostatic measure used along with a description of why it was considered Grade 4;
- For the gynecological procedure group, if applicable:
 - Estimated uterine size in grams;
- Occurrence of blood transfusion, if applicable (record the total required units of blood and rationale);
- Concomitant procedures, if applicable (defined as any medical or surgical procedure beyond those associated with primary study procedure). If a concomitant procedure requires vessel transection by HARMONIC 1100 Shears, each vessel transected will have the same information collected as stated above for vessel transections, but these will not be included as part of the primary endpoint;
- Concomitant medication usage associated with device-related or procedure-related AEs and specific drugs used up to 30 days prior to surgery (aspirin, antiplatelet agents, nonsteroidal anti-inflammatory drugs, anticoagulants) and 90 days prior to surgery (pre-surgical radiation/chemotherapy); and
- Date/time of hospital discharge.

3.4.3 Questionnaires

Data related to the following study variables will be obtained from the Surgeon Questionnaire for every investigator after they have completed his/her second procedure and a GEN11 Questionnaire after every completed procedure. Both Surgeon and Generator Questionnaire responses will then be transcribed onto eCRFs, with the questionnaire serving as the source document.

Surgeon Questionnaire

Investigators will be asked to answer a non-validated surgeon questionnaire related to his/her experience using the HARMONIC 1100 Shears. The survey will be completed by the Investigators as soon as possible after completion of his/her second procedure in any of the seven procedure groups.

Generator Questionnaire

Investigators will be asked to answer a non-validated generator questionnaire related to his/her experience using the GEN11 and any HARMONIC and/or ENSEAL device. The survey will be completed by the Investigators as soon as possible after each procedure, preferably on the same day for each GEN11 used.

3.4.4 Safety Data Variables

All device-related and procedure-related AEs will be collected as outlined in Section 6.0.

3.5 SCHEDULE OF EVENTS

Table 2: Schedule of Events

Activity	Visit 1 ^a	Visit 2 ^a	Visit 3	Unscheduled Visit, if applicable ^b
	Screening Visit (Day -56 to Day 0)	Procedure Day (0) Through Discharge Visit	Post Procedure Follow Up Visit (Day 28 ± 14) ^b	
Informed consent/assent, as applicable	X			
Demographics (age, gender, race, and ethnicity)	X			
Administer pregnancy test as per standard of care	X	X		
Height and weight (both with no shoes)		X		
Medical and surgical history (including diagnosis)	X	X		
Review of inclusion/exclusion criteria	X	X		
Concomitant medications	X	X	X	X
Surgical data collected for evaluation		X		
Collect vessel image ^c		X		
Collect generator questionnaire ^d		X		
Collect surgeon questionnaire ^e		X		
Concomitant procedures conducted besides primary procedure		X		
Re-operation(s) associated with the primary procedure			X	X
Assess for device-related and procedure-related AEs		X	X	X
Subject completion/discontinuation		X	X	X

^a Visit 1 and Visit 2 may be combined;

^b This can be either an office visit or telephone follow-up;

^c The image captured where the vessel is below the open jaw of the HARMONIC 1100 Shears perpendicular to the vessel during the primary procedure at Visit 2;

^d The generator questionnaire will be completed after each Visit 2 procedure;

^e The surgeon questionnaire will be completed after the investigator has completed his/her second procedure in any of the seven procedure groups.

AE = adverse events.

3.6 STUDY PROCEDURES

3.6.1 Visit 1 – Screening (may occur over visits within 8 weeks of Visit 2 including up to the day of the procedure)

Prospective subjects will be provided with the study information including the ICD. The subject must be given ample time to review and sign the ICD. The screening activities noted in Section 3.5 will occur prior to the study procedure.

3.6.2 Visit 2 – Procedure Through Discharge

Pre-procedure

The following must be obtained prior to the surgical procedure:

- Pregnancy test as per standard of care;
- Height and weight (both with no shoes);
- Updates to concomitant medication and medical/surgical/smoking history; and
- Confirm inclusion and exclusion criteria.

Intra-operative and Post-operative

Data collected during and after the procedure as defined in Section 3.4.

3.6.3 Visit 3 – Post-procedure Follow-up

After the date of surgery, subjects will have a follow-up visit approximately 28 days (\pm 14 days) later. This visit can be either an office visit or a telephone follow-up. Data to be collected is defined in Section 3.5.

3.6.4 Unscheduled Visit(s)

Any unscheduled visit between Visit 2 and Visit 3 will be documented including the reason for the visit. This visit can be either an office visit or a telephone follow-up. Data to be collected is listed in Section 3.5.

3.6.5 Vessel Image Capture

The preferred image should be captured during the primary procedure at Visit 2 where the vessel (or the tissue bundle containing the vessel) is below the open jaw of the HARMONIC 1100 Shears. The image should show the device perpendicular to the vessel. The image should have the entire length of the jaw of the HARMONIC 1100 Shears in view. The image will be uploaded to the electronic database capture (EDC) system as either a jpeg or pdf file. The Sponsor will determine if the image is analyzable and then each image will be reviewed and have a vessel size determined (if measurement is feasible) to the nearest quarter decimal (e.g., 5.0 mm, 5.25 mm, 5.50 mm, etc.) by the Sponsor utilizing computer software with the measurement captured in the EDC. No reconciliation will be performed between the surgeon determination of vessel size (see Section 3.4.2) and the vessel size determined by the image analysis. The primary vessel size determination will be by the surgeon determination.

4.0 DATA MANAGEMENT AND INTEGRITY

4.1 DATA COMPLETION AND RECORD KEEPING

4.1.1 Source Documents

Source documents are documents on which information regarding subjects is first recorded, including printed, optical, or electronic documents. Investigator subject files or hospital records generally are the basis of source document information. This includes but is not limited to, original subject files; hospital/clinic records; original recordings /tracing; radiographs; device accountability records; photographic negatives; and records kept at the investigation site, at the laboratories and at other departments involved in the clinical investigation. Source document worksheets may also be used to facilitate data collection and entry into eCRFs.

Source documents must be retained by the Investigator as part of the subject's study record. The information in the source documents is used to complete the eCRFs. All information captured on the eCRFs should be completely and accurately supported in source documentation. Any additional information relevant to the study should be included in the source documents. Any deviations from the study protocol or procedures should be recorded in the source documents. The Investigator will retain originals of all source documents, subject consent forms, and study data per site policy.

4.1.2 Electronic Data Capture

An EDC system will be utilized by study site personnel to transfer study data from source records (medical records and/or source document worksheets) onto common eCRFs. This system is a web-based, secure electronic software application (Medidata® Rave, 350 Hudson St. 9th Floor, NY, NY 10014). This system was designed and is developed and maintained by Medidata in a manner that is compliant with national and international GCP data protection/data privacy and electronic record/electronic signature (e.g., 21 Code of Federal Regulations Part 11) regulatory requirements. The EDC system will be used to facilitate the collection of all study data at the site. Designated site personnel will be responsible for entering subject data into the EDC system. All external and Sponsor internal users will be trained on the EDC application at a level dependent on their planned function. An EDC digital User Manual will be available under the help menu within the Medidata® Rave website to assist in the collection and entry of source data into the electronic casebook.

A 24/7/365 Help Desk Support line (telephone: +001 973-659-6780; fax: 973-954-5621; toll free: +001 866-633-4328; Email: helpdesk@mdsol.com) staffed by the outsourced vendor will also be available to respond to site and monitor questions.

4.1.3 Data Collection

Each EDC eCRF will be completed by the PI or PI's designee. Every effort should be made to respond to all monitoring and/or data management questions on each eCRF as completion of the data is required by the protocol. A unique ID number will identify each subject and will be visible on each eCRF. At no time should the subject name or protected health information appear on the eCRFs.

All data should be recorded accurately and completely. The PI is responsible for reviewing and approving each completed eCRF. Assurance of overall review and approval will be documented by the PI electronically signing each subject's electronic casebook at the end of the study.

4.1.4 Data Correction

Required data corrections to eCRFs will be prompted via automated electronic edit checks and/or queries manually created by Sponsor reviewers. The change(s), individual making the change(s), and time the change(s) was made to the eCRFs will be automatically captured in the audit trail within Medidata® Rave.

4.1.5 Data Privacy

The collection, use, and disclosure of all personal data, including subject health and medical information, are to be maintained in compliance with applicable personal data protection and security laws and regulations that govern protected health information and the informed consent given by each study subject. When collecting and processing such personal data, appropriate measures are to be taken to maintain the confidentiality of subject health and medical information and to prevent access by unauthorized persons.

4.1.6 Record Retention, Inspection, and Custody

The PI must maintain all documentation related to the study until notified by the Sponsor. The PI will allow representatives of the Sponsor or other government regulatory agencies to inspect all study records, eCRFs, and corresponding portions of the subject's office and/or hospital medical records at regular intervals during the study. These inspections are to verify adherence to the protocol, integrity of the data being captured on the eCRFs, and compliance with applicable regulations.

Subject study records will be maintained in a confidential manner. Study reports will not identify subjects by name. These reports may be submitted to regulatory authorities.

If custody of the records is transferred, notice of such a transfer should be given to the Sponsor no later than 10 working days before the transfer occurs.

4.2 MEDICAL DICTIONARY CODING

Medical dictionary coding of medical history and verbatim AEs captured on eCRFs will be performed using a coding thesaurus algorithm. The Medical Dictionary for Regulatory Activities (MedDRA) will be used after data entry and query resolution, via auto-encoding and interactive coding processes.

4.3 DATA QUALITY ASSURANCE

Steps to be taken to assure the accuracy and reliability of data include the selection of qualified investigators and appropriate sites, review of protocol procedures with the Investigator and associated personnel prior to the study, and periodic monitoring visits by the Sponsor. The Sponsor will review eCRFs for accuracy and completeness during either onsite or offsite monitoring visits; any discrepancies will be resolved with the Investigator or designees, as appropriate.

4.3.1 Site Personnel Training

Prior to screening subjects for this study, any relevant study staff will be trained on study execution, data collection, and procedures specific to this clinical protocol, as applicable.

4.3.2 Monitoring

This study will be monitored by the Sponsor or its representative to ensure:

- The rights and well-being of the subjects are protected;
- The reported study data is accurate, complete, and verifiable from source documents; and

- The conduct of the study is in compliance with the currently approved protocol/amendment(s), applicable GCPs, and with applicable local/regional regulatory requirements.

The extent and nature of monitoring will be predetermined and agreed to by the Sponsor and investigators. Monitors will comply with established written standard operating procedures as well as procedures specified by the Sponsor in the monitoring plan for this study.

4.3.3 Quality Assurance

The extent and nature of quality assurance audits will be predetermined and based on considerations such as the regulatory classification, objective, and complexity of the study. Any audits performed will comply with established written standard operating procedures and the audit plan for this study.

4.4 PROTOCOL DEVIATIONS

A deviation (any activity conducted outside the parameters established by the protocol) can be identified from a number of sources. Potential sources for identification of deviations include, but are not limited to: a member of the Investigator's staff, a Sponsor representative during monitoring visits, or a member of the data management or statistical groups when entering or analyzing data. Regardless of the source, it is crucial to document the deviation within source documentation. Protocol deviations will be captured in the eCRFs. The PI will report protocol deviations to the IRB/EC as required by the IRB/EC procedures.

5.0 STATISTICAL METHODS PLANNED IN THE PROTOCOL AND DETERMINATION OF SAMPLE SIZE

The Sponsor will be responsible for the analysis of data from this protocol. A detailed Statistical Analysis Plan (SAP) will be written and approved prior to final database lock. The SAP will describe all planned analyses based on the statistical design of this study and the subsequent data collected. A brief overview of key statistical analyses is provided below.

5.1 STUDY DESIGN

Study design is described in Section 3.1.

5.2 TREATMENT ASSIGNMENT

This is a single-arm study where all enrolled subjects will have the HARMONIC 1100 Shears utilized for transection of at least one single vessel.

5.3 INTERVAL WINDOWS

Interval windows for the purpose of analysis in this study will not be defined outside of those already specified in the protocol for visit scheduling as the collection of data for the primary and secondary performance endpoints occurs intra-operatively. The final visit occurs approximately 4 weeks after surgery, thus no interval windows need to be defined given the absence of long-term follow-up in this study. The Schedule of Events specifies a window of 14 days around the scheduling of the 4-week follow-up visit, and any information entered in the eCRFs at this visit will correspond to the 4-week visit. There will be no assigning of observations to time points outside of the visit to which they are recorded in the eCRFs.

5.4 PRIMARY AND SECONDARY ENDPOINTS AND ASSOCIATED HYPOTHESES

All study endpoints are described in Section 2.1 through Section 2.4. The following hypothesis will be evaluated for the primary endpoint:

$$H_0: p \leq 87.5\%$$

$$H_1: p > 87.5\%$$

Where p is the percentage of transections achieving a Grade 3 or lower hemostasis rating and 87.5% is set as a performance goal for the lower bound of acceptable hemostasis. A two-sided 95% confidence interval will be calculated for p based on the sample proportion of transections where Grade 3 or lower hemostasis was achieved using exact Binomial methods proposed by Clopper-Pearson and the lower limit of this confidence interval will be compared to 87.5% to evaluate the above hypotheses. A p-value will be determined based on an exact binomial test. The null hypothesis will be rejected if the lower limit of the confidence interval is greater than 87.5%.

Given that multiple transections per subject may occur, a sensitivity analysis will be performed using an appropriate method to account for the potential correlation among the outcomes of multiple transections within the same subject. Complete details of this approach will be provided in the SAP.

The hypothesis and evaluation of hypothesis for the secondary endpoint for the proportion of vessels transected of diameter size > 5 to 7 mm with hemostasis grade 3 or lower will be similar to the primary endpoint as described above.

5.5 LEVELS OF SIGNIFICANCE

The hypotheses above will be evaluated using a one-sided significance level of 0.025. Estimation of all additional endpoints will be performed using two-sided 95% confidence intervals.

5.6 ANALYSIS SETS

The summary of all performance and safety endpoints will be performed on the set of subjects in whom the HARMONIC 1100 Shears is utilized during the surgical procedure to transect at least one vessel. The summary of all primary and secondary performance and safety endpoints will be performed by surgical population (pediatric and adult), surgical procedure group, procedure subgroup, and on the entire pooled set of subjects. The hypotheses described in Section 5.4 will be evaluated on the pooled set of subjects and the > 5 to 7 mm diameter vessel population. The > 5 to 7 mm diameter vessel population will be based off the surgeon determination of the vessel size.

5.7 SAMPLE SIZE JUSTIFICATION

Primary performance analysis: A sample size of at least 165 subjects is required to have a minimum of 80% power for rejecting the null hypothesis that the rate of Grade 3 or lower transections is less than or equal to 87.5% when the expected rate of Grade 3 hemostatic transections is at least 94.0% using an exact binomial test and an one-sided significance level of 0.025. This sample size estimation makes the assumption of one transection per subject which is likely very conservative as most subjects may have 2 or more transections during their procedure. Therefore, it is expected that greater than 165 transections will be observed and thus the study will be well powered for evaluating the performance goal hypothesis under the assumption of independence of observations. At the time of study protocol amendment 3 it was found that on average there were 2 vessel transections per subject.

The assumption of independence and its impact on the overall study conclusion will be evaluated through sensitivity analyses. Given that the primary endpoint is being evaluated intra-operatively, it is not anticipated that there will be subject dropout prior to evaluating this endpoint, thus no adjustment or increase in subject enrollment is planned to account for dropout. A minimum of 33 subjects will be enrolled into each of the five individual procedure subgroups.

The performance goal of 87.5% was selected to represent a minimum acceptable level of clinical performance across all procedure groups. A recent energy study² with the same hemostasis endpoint in similar procedure groups showed point estimates of the primary endpoint ranging from 94.2% to 100.0%. The 95% confidence interval for the lowest performing group had a lower bound of 89.3%, thus selection of 87.5% represents a less than 2% region of indifference for maintaining average acceptable performance. In the recent study, average performance across all groups was 96.2% and selection of 87.5% as the performance goal corresponds to a noninferiority margin of 8.7% which is less than the commonly chosen value of 10%. Finally, consideration of the performance goal of 87.5% in the context of the relative reduction from the overall average performance of 96.2%, this can be interpreted as preserving at least 91.0% of the original average treatment effect ($87.5/96.2 \times 100\% = 91.0\%$). Maintaining at least 91.0% of the expected effect of similar therapies represents a clinically meaningful metric for maintaining minimum acceptable performance and is actually considerably larger than the standard limits used for bioequivalence where the lower bound for relative therapeutic efficacy is set at 80%, considerably lower than the 91.0% of therapeutic efficacy that is represented by a performance goal of 87.5% chosen for this trial.

Secondary performance analysis: A sample size of at least 165 vessels is required to have a minimum of 80% power for rejecting the null hypothesis that the rate of Grade 3 or lower transections is less than or equal to 87.5% when the expected rate of Grade 3 hemostatic transections is at least 94.0% for vessels with diameter size of > 5 to 7 mm. Sample size estimation considered an exact binomial test and a one-sided level of significance of 0.025. The number of vessels transected with a diameter size of > 5 to 7 mm will be continuously checked throughout the course of the study to achieve the required number of vessels.

Safety analysis: From a safety perspective on the pooled analysis of 165 subjects and in consideration of rare AEs that may occur (e.g., bleeding requiring blood product transfusion), for an event that has an incidence rate of, for example, 1%, then in a sample of 165 subjects, the probability of observing at least 1 event is 81.0% under a binomial probability model. Thus, this sample size provides a high probability of observing rare events if they do occur, and provides reasonable assurance to conclude that the likelihood of such AEs is less than 2.3% if they do not occur based on the upper limit of an exact 95% confidence interval when 0 events out of 165 subjects are observed.

5.8 DATA MONITORING COMMITTEE

There are no plans to utilize a data monitoring committee during this study.

5.9 ANALYSES TO BE CONDUCTED

5.9.1 General Conventions

Categorical variables will be summarized descriptively by frequencies and associated percentages. Continuous variables will be summarized descriptively by number of subjects, mean, standard deviation, median, minimum, and maximum. Confidence intervals will also be provided for procedure-related variables.

5.9.2 Disposition of Study Subjects

Subject disposition will be summarized in total, by surgical population (pediatric and adult), surgical procedure group, and procedure subgroups using counts and percentages. The number and percentage of subjects completed and discontinued will be tabulated along with the specific reasons for discontinuation.

5.9.3 Demographic, Baseline, and Surgical Characteristics

Summary statistics of subject demographics (age, gender, and race) will be presented in total, and by surgical population (pediatric and adult), surgical procedure group, and procedure subgroups. Surgical characteristics including, at minimum, estimated blood loss, requirement for blood transfusion, and procedure duration will be summarized by surgical population (pediatric and adult), surgical procedure group, and procedure subgroups and in total.

5.9.4 Primary and Secondary Endpoint Analyses

The number and percentage of vessels where hemostasis is achieved (\leq Grade 3) will be summarized and an exact two-sided 95% confidence interval based on Clopper-Pearson's method will be estimated for each surgical population (pediatric and adult), surgical procedure group, procedure subgroup, the > 5 to 7 mm diameter vessel population, and in total. The performance goal hypothesis will be tested as described in Section 5.4 on the total population and the > 5 to 7 mm diameter vessel population.

Summary statistics appropriated for categorical or continuous variable will be provided for type, size, and number of vessels transected, as well as for the grading scale distribution for all vessels transected, number of times HARMONIC 1100 Shears touch-ups were required, incidence of requirement for additional measures to obtain hemostasis on vessels (other advanced energy devices or hemostatic measures), and 5-point scale scores for sealing and transection of lymphatic vessels completed by the HARMONIC 1100 Shears. In addition, these analyses will also be completed for the vessels in the > 5 to 7 mm diameter size vessel range.

All device-related and procedure-related AEs reported during the study will be coded to MedDRA. All AEs will be summarized by MedDRA system organ class and preferred term by surgical population (pediatric and adult), surgical procedure group, procedure subgroup, and in total. Separate summaries will be provided for device-related and procedure-related AEs. Serious AEs will be summarized in a similar manner. In addition, these analyses will also be completed for the vessels in the > 5 to 7 mm diameter size vessel range.

5.9.5 Plans for Interim Analyses

No interim analysis is planned for this study.

5.9.6 Handling of Missing Data

All summaries will be performed only on subjects undergoing the scheduled procedure and only observed data will be summarized. There will be no imputation of data for early terminated subjects or for missing data within the database.

5.9.7 Subgroup Analyses

Subgroup analyses are planned to be performed for the subgroup of subjects who have a medical history of treatment for cancer (e.g. chemotherapy or radiation) and may be performed for additional groups pending the distributions of baseline demographic or clinical characteristics. These analyses will be exploratory and summary statistics for the

procedure-related parameters will be provided for each subgroup.

6.0 ADVERSE EVENTS

Only AEs that have a relationship considered possible, probable, or causal to the devices or the procedure will be collected during this study.

6.1 DEFINITIONS

6.1.1 Adverse Event

An AE is defined as any untoward medical occurrence, regardless of its relationship to the study device or the study procedure. An untoward medical occurrence includes any new, undesirable medical experience or worsening of a pre-existing condition, which occurs throughout the duration of the clinical study.

6.1.2 Serious Adverse Event

It is the Investigator's responsibility to determine the "seriousness" of a reportable AE.

A SAE is defined as an AE (as defined in Section 6.1.1) that results in any of the following:

- Death;
- A life-threatening illness or injury;
- A permanent impairment of a body structure or a body function;
- Required in-patient hospitalization or prolongation of existing hospitalization;
- Resulted in medical or surgical intervention to prevent life-threatening illness or injury or permanent impairment to a body structure or a body function;
- Led to a fetal distress, fetal death, or a congenital abnormality or birth defect.

Note: "Death" should not be reported as an AE. The cause of death should be reported as the AE. The only exception is "Sudden Death" when the cause is unknown.

Note: Planned hospitalization for a pre-existing condition, or a procedure required by the protocol, without serious deterioration in health, is not considered a SAE.

6.1.3 Pre-existing Condition

A pre-existing condition is defined as a medical condition that is present at the initiation of the study and is to be reported as part of the subject's medical history with a listed start date. Any pre-existing condition that has worsened in intensity, frequency, or the character of the condition, if related to the devices or the procedure, should be recorded as a new AE on the AE eCRF, as an exacerbation of the pre-existing condition and the start date will be recorded as the time when the exacerbation occurred.

6.2 SEVERITY OF ADVERSE EVENTS

It is the Investigator's responsibility to assess the severity of an AE. A change in severity may constitute a new reportable AE.

The following guideline should be used to determine the severity of each AE:

- **MILD:** A type of AE that is usually transient and may require only minimal treatment or therapeutic intervention. The event does not generally interfere with usual activities of daily living.
- **MODERATE:** A type of AE that is usually alleviated with additional specific therapeutic intervention. The event interferes with usual activities of daily living, causing discomfort but poses no significant or permanent risk of harm to the

research participant.

- **SEVERE:** A type of AE that interrupts usual activities of daily living, or significantly affects clinical status, or may require intensive therapeutic intervention.

6.3 RELATIONSHIP OF ADVERSE EVENTS

It is the Investigator's responsibility to assess the relationship of a reportable AE (as defined in Section 6.1.1). Only AEs that have a relationship considered possible, probable, or causal to the study procedures or either of the HARMONIC 1100 Shears or GEN11 devices are to be recorded in the eCRF and reported to the Sponsor.

The following guidelines should be used in determining the relationship of an AE in the study:

- **Not related** – Relationship to the procedures or device can be excluded when:
 - The event is not a known side effect of the product category the device belongs to or of similar devices and procedures;
 - The event has no temporal relationship with the use of the device or the procedures;
 - The event does not follow a known response pattern to the device (if the response pattern is previously known) and is biologically implausible;
 - The discontinuation of the device application or the reduction of the level activation/exposure (when clinically feasible) and reintroduction of its use (or increase of the level of activation/exposure), does not impact on the event;
 - The event involves a body-site or an organ not expected to be affected by the device or the procedure;
 - The event can be attributed to another cause (e.g. an underlying or concurrent illness/clinical condition, an effect of another device, drug, treatment, or other risk factors);
 - Harms to the subject are not clearly due to use error; or
 - To establish the non-relatedness, not all the criteria listed above might be met at the same time, depending on the type of device/procedure and the event.
- **Possible** – The relationship with the use of the device is weak but cannot be ruled out completely. Alternative causes are also possible (e.g. an underlying or concurrent illness/condition and/or an effect of another device, drug, or treatment). Cases where relatedness cannot be assessed or no information has been obtained should also be classified as possible.
- **Probable** – The relationship with the use of the device seems relevant and/or the event cannot reasonably be explained by another cause, but additional information may be obtained.
- **Causal relationship** – The event is associated with the device or with procedures beyond reasonable doubt when:
 - The event is a known side effect of the product category the device belongs to or of similar devices and procedures;
 - The event has a temporal relationship with the device uses/application or

procedures;

- The event involves a body-site or organ that:
 - The device or procedures are applied to;
 - The device or procedures have an effect on;
- The event follows a known response pattern to the medical device (if the response pattern is previously known);
- The discontinuation of medical device application (or reduction of the level of activation/exposure) and reintroduction of its use (or increase of the level of activation/exposure), impact on the event (when clinically feasible);
- Other possible causes (e.g. an underlying or concurrent illness/clinical condition and/or an effect of another device, drug, or treatment) have been adequately ruled out;
- Harm to the subject is due to error in use; or
- To establish the relatedness, not all the criteria listed above might be met at the same time, depending on the type of device/procedure and the event.

6.4 REPORTING PROCEDURES FOR ADVERSE EVENTS

6.4.1 Recording Adverse Events

The Investigator will record all AEs (both AEs and SAEs) considered attributable (relationship of possible, probable, causal relationship) to the study procedure or either of the HARMONIC 1100 Shears or GEN11 devices in the source documents and eCRF.

Standard medical terminology should be used when recording AEs. In addition, the following information should be recorded:

- Onset date;
- Resolution date or date of death;
- Severity of the event;
- Indication of whether the event is serious;
- Relationship of AE to the study devices;
- Relationship of AE to the study procedure;
- Action taken;
- Event status; and
- Was AE anticipated or not (only for SAEs).

Data related to AEs will be collected until the event resolution, or until the event is considered stable, or until all attempts to determine the resolution of the event are exhausted. All AEs that are unresolved at study completion (or at the last visit completed for early withdrawal subjects) will be recorded as ongoing at study end.

6.4.2 Reporting Adverse Events

The Investigator is required to report AEs to the Sponsor within 2 weeks from when the study site becomes aware of the event. All AEs must be reported on the AE eCRF and be documented in a timely manner throughout the study duration, with reporting of AEs starting from the time the subject is enrolled.

6.4.3 Reporting Serious Adverse Events

Any AE that the Investigator determines to be a SAE must be reported by the study site

within 72 hours of becoming aware of the event to the Sponsor by completing the AE eCRF. A notification containing the pertinent data will be automatically generated by the EDC system and forwarded to the Sponsor. If applicable, supporting SAE documentation should be de-identified and provided to the Clinical Trial Leader. The report of an SAE by a study site does not constitute an admission that study personnel or the user facility (hospital/clinic) caused or contributed to the event. The study site is also responsible for submitting SAEs to the reviewing IRB/EC per their IRB/EC procedures.

7.0 PRODUCT COMPLAINTS

A product complaint is any written, electronic, or oral communication that alleges deficiencies related to the identity, labeling, quality, durability, reliability, safety, effectiveness, or performance of a device after it is released for distribution. A product complaint may or may not be associated with an AE/SAE.

7.1 REPORTING PRODUCT COMPLAINTS FOR MARKETED DEVICES

7.1.1 Reporting of Approved EES Marketed Devices

All product complaints related to devices in the procedure shall be documented throughout the clinical investigation. Product complaints related to Ethicon-manufactured HARMONIC 1100 Shears or GEN11 devices must be reported to the Sponsor in a timely manner and no later than 24 hours after becoming aware of the event. When a sponsor representative becomes aware of a product complaint, the Product Complaint Team must be notified within 24 hours after becoming aware of the event. The Product Complaint Form must be emailed to the Sponsor Customer Complaint team at the following email address:

Productcomplaint1@its.jnj.com

The sites will report Product Complaints by completing the Product Complaint Form. If a product complaint is related to a safety event, the site must complete an AE eCRF as well. One copy of the processed form should be kept on-site, and the device should be retained. Sponsor representatives will organize collection of the device for evaluation as needed.

7.1.2 Reporting of Approved Marketed Devices and Commercially Available Devices (not Ethicon)

The site should report Product Complaints of non-Ethicon devices that are approved marketed devices or commercially available medical devices according to the specific product's manufacturer specifications/guidelines.

8.0 REFERENCES

1. Siegel JM, Cummings JF, Clymer JW. Reproducible, repeatable and clinically-relevant hemostasis scoring. J Adv Med Pharm Sci 2014;1:30-39.
2. Schilder J, Anderson D, Shah F. et al. Hemostatic efficacy of an advanced bipolar sealer in open gynecologic, thoracic, and colectomy procedure: a prospective cohort study. Int J Surg Open 2020;24:57-63.