

Title

**Clinical Performance of a Silicone Hydrogel Following One Night of  
Extended Wear**

Protocol Number: CLL949-C004 / NCT03560141

Development Stage of  
Project:

Sponsor Name and  
Address: Alcon Research [REDACTED]  
6201 South Freeway  
Fort Worth, Texas 76134-2099

Test Product: [REDACTED] soft contact lenses

Investigator Agreement: I have read the clinical study described herein, and recognize its confidentiality. I agree to conduct this study in accordance with the ethical principles contained within the Declaration of Helsinki, and the described study in compliance with the protocol and any conditions of approval imposed by the Institutional Review Board and/or the FDA, Good Clinical Practice (GCP), ISO 14155, and all applicable regulatory requirements including 21 CFR Parts 50 (Protection of Human Subjects), 54 (Financial Disclosure), 56 (Institutional Review Boards), and 812 (Investigational Device Exemptions). Additionally, I will comply with all procedures for data recording and reporting, will permit monitoring, auditing, and inspection of my research center, and will retain all records until notified by the Study Sponsor.

Principal Investigator:

---

Signature

Date

Name and professional  
position:

Address:

[REDACTED]

## Table of Contents

Table of Contents .....	2
List of Tables .....	4
List of Figures .....	5
1 GLOSSARY OF TERMS .....	6
2 LIST OF ACRONYMS AND ABBREVIATIONS .....	10
3 PROTOCOL SUMMARY .....	11
4 [REDACTED]	
5 [REDACTED]	
5 INTRODUCTION .....	20
5.1 Rationale and Background .....	20
5.2 Purpose of the Study .....	20
5.3 Risks and Benefits .....	21
6 STUDY OBJECTIVES .....	21
6.1 Primary Objective .....	21
6.2 Secondary Objective .....	21
6.4 Safety Objective .....	22
7 INVESTIGATIONAL PLAN .....	22
7.1 Study Design .....	22
7.2 Rationale for Study Design .....	23
7.3 Rationale for Duration of Treatment/Follow-Up .....	23
7.4 Rationale for Choice of Control Product .....	23
7.5 Data Monitoring Committee .....	23
8 STUDY POPULATION .....	24
8.1 Inclusion Criteria .....	24
8.2 Exclusion Criteria .....	25
8.3 Rescreening of Subjects .....	26
9 TREATMENTS ADMINISTERED .....	26
9.1 Investigational Product(s) .....	26
9.2 Other Medical Device or Medication Specified for Use During the Study .....	29

9.3	Treatment Assignment / Randomization .....	29
9.4	Treatment masking .....	30
9.5	Accountability Procedures.....	31
9.6	Changes to concomitant medications, treatments/ procedures.....	32
<b>10</b>	<b>STUDY PROCEDURES AND ASSESSMENTS .....</b>	<b>32</b>
10.1	Informed Consent and Screening .....	33
10.2	Description of Study Procedures and Assessments .....	33
10.2.1	Demographics .....	33
10.2.2	Medical History .....	34
10.2.3	Investigational Product compliance .....	34
10.2.4	Adverse Event Collection: Safety Assessment .....	34
10.2.5	Slit-Lamp Biomicroscopy: Safety Assessment .....	34
10.2.6	Device Deficiencies: Safety Assessment.....	34
10.3	Unscheduled Visits .....	35
10.4	Discontinued Subjects .....	35
10.4.1	Screen Failures .....	35
10.4.2	Discontinuations .....	35
10.4.3	Schedule of Procedures and Assessments for Subjects Discontinued from Investigational Product .....	36
10.5	Clinical Study Termination.....	36
10.5.1	Follow-up of subjects after study participation has ended .....	37
<b>11</b>	<b>ADVERSE EVENTS AND DEVICE DEFICIENCIES .....</b>	<b>37</b>
11.1	General Information .....	37
11.2	Monitoring for Adverse Events .....	40
11.3	Procedures for Recording and Reporting .....	40
11.4	Return product analysis .....	42
11.5	Unmasking of the Study Treatment.....	42
11.6	Follow-Up of Subjects with Adverse Events.....	43
11.7	Pregnancy in the Clinical Study .....	43
<b>12</b>	<b>ANALYSIS PLAN .....</b>	<b>43</b>
12.1	Subject Evaluability.....	44
12.2	Analysis Sets .....	44

12.3	Demographic and Baseline Characteristics .....	44
12.4	Effectiveness Analyses .....	44
12.4.1	Analysis of Primary Effectiveness Endpoint(s).....	44
12.4.1.1	Statistical Hypotheses .....	44
12.4.1.2	Analysis Methods .....	44
12.5	Handling of Missing Data .....	45
12.6	Safety Analyses .....	45
13	DATA HANDLING AND ADMINISTRATIVE REQUIREMENTS .....	46
13.1	Subject Confidentiality .....	46
13.2	Completion of Source Documents and Case Report Forms .....	47
13.3	Data Review and Clarifications .....	48
13.4	Sponsor and Monitoring Responsibilities .....	48
13.5	Regulatory Documentation and Records Retention .....	49
13.6	Quality Assurance and Quality Control.....	49
14	ETHICS .....	49
15	REFERENCES .....	51
15.1	References applicable for all clinical studies .....	51
15.1.1	US references applicable for clinical studies.....	51
15.2	References for this clinical study .....	51

## List of Tables

Table 2-1	List of Acronyms and Abbreviations Used in This Protocol .....	10
Table 3-1	Schedule of Study Procedures and Assessments .....	15
Table 6-2	Safety Objective.....	22
Table 9-1	Test Product .....	26

Table 9-2	Control Product.....	28
-----------	----------------------	----



## **List of Figures**

Figure 11-1	Categorization of All Adverse Events.....	37
Figure 11-2	Categorization of All Serious Adverse Events.....	38

## 1 GLOSSARY OF TERMS

Names of test product	Throughout this document, the test product will be referred to as [REDACTED] soft contact lenses.
Name of Control Product	Biofinity® (comfilcon A) soft contact lenses, referred to as Biofinity.
Adverse Device Effect	Adverse event related to the use of an investigational medical device (test product) or control product. <i>Note: This definition includes adverse events resulting from insufficient or inadequate instructions for use, deployment, implantation, installation, or operation; any malfunction; and use error or intentional misuse of the test product or control product.</i>
Adverse Event	Any untoward medical occurrence, unintended disease or injury, or untoward clinical signs (including abnormal laboratory findings) in subjects, users or other persons, whether or not related to the investigational medical device (test product). <i>Note: For subjects, this definition includes events related to the test product, the control product, or the procedures involved. For users or other persons, this definition is restricted to events related to the test product.</i>  Requirements for reporting Adverse Events in the study can be found in Section 11.
Anticipated Serious Adverse Device Effect	Serious adverse device effect which by its nature, incidence, severity, or outcome has been identified in the risk management file.

Device Deficiency	<p>Inadequacy of a medical device with respect to its identity, quality, durability, reliability, safety, or performance.</p> <p><i>Note: This definition includes malfunctions, use errors, and inadequate labeling.</i></p> <p>Requirements for reporting Device Deficiencies in the study can be found in Section 11.</p>
Enrolled Subject	Any subject who signs an informed consent form for participation in the study.
Interventional Clinical Trial	A research trial that prospectively assigns, whether randomly or not, human participants or groups of humans to one or more health-related interventions to evaluate the effects on health outcomes, and/or a research trial in which diagnostic or monitoring procedures beyond standard of care are conducted and generate outcomes for use in analysis of data.
Investigational Product	Is defined as a preventative (vaccine), a therapeutic (drug or biologic), device, diagnostic, or palliative used as a test or control product in a clinical trial, including a product with a marketing authorization when used or assembled (formulated or packaged) in a way different from the authorized form, or when used for an unauthorized indication, or when used to gain further information about the authorized form.
Malfunction	Failure of a medical device to perform in accordance with its intended purpose when used in accordance with the instructions for use or clinical investigation plan.
Non-serious Adverse Event	Adverse event that does not meet the criteria for a serious adverse event.
Randomized Subjects	Any subject who is assigned a randomized treatment.
Serious Adverse Device Effect (SADE)	Adverse device effect that has resulted in any of the consequences characteristic of a serious adverse event.

Serious Adverse Event (SAE)	<p>Adverse event that led to any of the following:</p> <ul style="list-style-type: none"><li>• Death.</li><li>• A serious deterioration in the health of the subject that either resulted in:<ol style="list-style-type: none"><li>a life-threatening illness or injury. <i>Note: Life-threatening means that the individual was at immediate risk of death from the event as it occurred, ie, it does not include an event which hypothetically might have caused death had it occurred in a more severe form.</i></li><li>b. any potentially sight-threatening event or permanent impairment to a body structure or a body function.</li><li>c. in-patient hospitalization or prolonged hospitalization. <i>Note: Planned hospitalization for a pre-existing condition, without serious deterioration in health, is not considered a serious adverse event. In general, hospitalization signifies that the individual remained at the hospital or emergency ward for observation and/or treatment (usually involving an overnight stay) that would not have been appropriate in the physician's office or an out-patient setting. Complications that occur during hospitalization are adverse events. If a complication prolongs hospitalization or fulfills any other serious criteria, the event is serious. When in doubt as to whether "hospitalization" occurred, the event should be considered serious.</i></li><li>d. a medical or surgical intervention to prevent a) or b).</li></ol></li></ul>
-----------------------------	---

	<p>e. any indirect harm as a consequence of incorrect diagnostic test results when used within manufacturer's instructions for use.</p> <ul style="list-style-type: none"> <li>• Fetal distress, fetal death, or a congenital abnormality or birth defect.</li> </ul> <p><i>Refer to Section 11 for additional SAEs.</i></p>
Significant Non-Serious Adverse Event	<p>Is a symptomatic, device-related, non-sight threatening adverse event that warrants discontinuation of any contact lens wear for greater than or equal to 2 weeks.</p> <p><i>Refer to Section 11 for additional Significant Non-Serious AEs.</i></p>
Unanticipated Serious Adverse Device Effect (USADE)	<p>Serious adverse device effect which by its nature, incidence, severity or outcome has not been identified in the risk management file.</p>
Use Error	<p>Act or omission of an act that results in a different medical device response than intended by manufacturer or expected by user.</p> <p><i>Note: This definition includes slips, lapses, and mistakes. An unexpected physiological response of the subject does not in itself constitute a use error.</i></p>

## 2 LIST OF ACRONYMS AND ABBREVIATIONS

Table 2-1

List of Acronyms and Abbreviations Used in This Protocol

Abbreviation	Definition
ADE	Adverse device effect
AE	Adverse event
[REDACTED]	[REDACTED]
Biofinity	Biofinity (comfilcon A) soft contact lenses
CFR	Code of Federal Regulations
CRF	Case report form
[REDACTED]	[REDACTED]
CRF	Case report form
CSM	Clinical site manager
CTT	Clinical trial team
D	Diopter(s)
eCRF	Electronic case report form
EDC	Electronic data capture
FDA	US Food and Drug Administration
GCP	Good Clinical Practice
IB	Investigator's brochure
ICF	Informed consent form
ICH	International Council for Harmonisation of Technical Requirements for Pharmaceuticals for Human Use
IDE	Investigational Device Exemptions
IEC	Independent ethics committee
IP	Investigational product
IRB	Institutional review board
IRT	Interactive response technology
ISO	International Organization for Standardization
LCSM	Lead clinical site manager
LID	Lens identification
LogMAR	Logarithm of the minimum angle of resolution
mm	Millimeter
MOP	Manual of procedures
N/A	Not applicable
OD	Right eye
OS	Left eye
SAE	Serious adverse event
SADE	Serious adverse device effect
SD	Standard deviation
SiHy	Silicone hydrogel
SLE	Slit-lamp examination
SOP	Standard operating procedure
Unsched	Unscheduled

Abbreviation	Definition
US / USA	United States of America
USADE	Unanticipated serious adverse device effect
VA	Visual acuity



### 3 PROTOCOL SUMMARY

This will be a prospective, randomized, controlled, double-masked, contralateral wear clinical trial.

Approximately 1 site in the US will enroll approximately 12 subjects. Subjects will be randomized to wear the test [REDACTED] lens in 1 eye and the control Biofinity lens in the other eye. Subjects will be expected to attend 4 visits: Baseline/Dispense, Follow-up prior to sleeping, Follow-up upon awakening/eye opening the morning following overnight lens wear, and Follow-up/Exit approximately 1 hour after awakening.

Following randomization during the Baseline/Dispense Visit, study lenses will be dispensed to the subject. All study contact lenses are to be worn overnight for 1 night only. Subjects will be required to sleep in the study lenses approximately 8 hours (7-10 hours) of closed-eye sleep.

<b>Investigational product type</b>	Device
<b>Study type</b>	Interventional
<b>Investigational products</b>	Test Product: [REDACTED] soft contact lenses Control Product: Biofinity (comfilcon A) soft contact lenses
<b>Purpose and rationale</b>	The purpose of this clinical is to evaluate the on-eye performance of the investigational [REDACTED] lens compared to the commercially available Biofinity lens following 1 night of extended wear.
<b>Objective</b>	The primary objective is to obtain [REDACTED] insights to safety and performance of [REDACTED] soft contact lenses when worn in an extended wear modality compared to the Biofinity soft contact lenses.
<b>Endpoints</b>	Primary Effectiveness

	<ul style="list-style-type: none"> <li>• Distance visual acuity (Snellen)</li> </ul> 
	<p>Safety</p> <ul style="list-style-type: none"> <li>• AEs</li> <li>• Biomicroscopy findings</li> <li>• Device deficiencies</li> </ul>
<b>Assessments</b>	<p>Effectiveness</p> <ul style="list-style-type: none"> <li>• VA (Snellen distance) with IP</li> </ul> 

	<p>Safety</p> <ul style="list-style-type: none"> <li>• AEs</li> <li>• Biomicroscopy</li> <li>• Device deficiencies</li> </ul>
<b>Study Design</b>	This will be a prospective, randomized, controlled, double-masked, contralateral clinical trial. Subject participation in the study will be approximately 2 days with 1 night of extended wear.
<b>Subject population</b>	<p>Volunteer subjects aged 18 or over who are soft contact lens wearers, excluding Biofinity habitual wearers, have at least 3 months of contact lens wearing experience, and who wear their habitual lenses at least 5 days per week in either a daily wear or extended wear modality. Subjects who wear their habitual lenses in a daily wear modality (ie, not extended wear) should wear their habitual lenses at least 12 hours per day.</p> <p>Pregnant and breastfeeding women are excluded from this study.</p>
<p><b>Key inclusion criteria</b> (See Section 8.1 for a complete list of inclusion criteria)</p>	<ul style="list-style-type: none"> <li>• Successful wear of spherical soft contact lenses in both eyes during the past 3 months for a minimum of 5 days per week in either a daily wear or extended wear modality. Subjects who wear their habitual lenses in a daily wear modality (ie, not extended wear) should wear their habitual lenses at least 12 hours per day. Subjects who wear their habitual lenses in an extended wear modality should wear their habitual lenses a minimum of 1 night per week.</li> <li>• BCVA 20/25 or better in each eye.</li> </ul>

<b>Key exclusion criteria</b>  (See Section 8.2 for a complete list of exclusion criteria)	<ul style="list-style-type: none"><li>• Current Biofinity lens wearer.</li></ul>
<b>Data analysis and sample size justification</b>	No formal hypotheses are formulated for the primary effectiveness endpoint of VA; hence, no inferential testing will be performed.  [REDACTED]  [REDACTED]  [REDACTED]  [REDACTED]  [REDACTED]
<b>Key words</b>	[REDACTED] Biofinity, extended wear, 1 night
<b>Associated materials</b>	Lubrication/re-wetting drops will not be permitted during this trial.

Table 3-1

Schedule of Study Procedures and Assessments

Procedure/ Assessment	Visit 1, Day 1: Screening/ Baseline/ Dispense	Visit 2, Day 1: Follow-up prior to sleeping ≥ 4 hours after Dispense	Visit 3, Day 2: Follow-up 0 to 10 mins upon eye opening* ~7:30am	Visit 4, Day 2: 1 Hour Follow- up/ Exit (-30/+60 mins after Visit 3)	Unsched Visit
		Instruct the subject to retire on Day 1 so that he/she sleeps in the lenses ~7-10 hours.	Visit 3 to occur immediately upon eye opening on Day 2. Subjects will be escorted to the slit-lamp with eyes closed.		
Informed Consent	✓	-	-	-	-
Demographics	✓	-	-	-	-
Medical History	✓	-	-	-	-
Concomitant Medications	✓	(✓)	(✓)	(✓)	(✓)
Inclusion/ Exclusion	✓	-	-	-	-
Habitual lens (brand, power)	✓	-	-	-	-
VA w/ habitual correction (Snellen distance, OD,OS)*	✓	-	-	✓	(✓)
███████████	█	█	█	█	██████████
███████████	█	█	█	█	██████████
███████████	█	█	█	█	██████████
Biomicroscopy	✓	✓ <sup>†</sup>	✓ <sup>†</sup>	✓ <sup>†</sup> ✓ for Exit	✓
Dispense study lenses	✓	-	-	-	-
VA w/ study lenses (Snellen distance, OD, OS)	✓	✓	✓	✓	✓
███████████	█	█	█	█	██████████

Procedure/ Assessment	Visit 1, Day 1: Screening/ Baseline/ Dispense	Visit 2, Day 1: Follow-up prior to sleeping ≥ 4 hours after Dispense	Visit 3, Day 2: Follow-up 0 to 10 mins upon eye opening* ~7:30am	Visit 4, Day 2: 1 Hour Follow- up/ Exit (-30/+60 mins after Visit 3)	Unsched Visit
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
[REDACTED]	■	■	■	■	■
AEs	✓	✓	✓	✓	✓
Device deficiencies	✓	✓	✓	✓	✓
Exit Form	(✓)	(✓)	(✓)	(✓)	(✓)

\*source only

\*Biomicroscopy with contact lenses still on-eye (not to include corneal staining)

(✓) assessment performed as necessary.

## 4 PROTOCOL AMENDMENTS

Modification of the protocol is prohibited without prior written agreement in the form of a protocol amendment. All amendments must be created by the Study Sponsor and must be approved by the IRB/IEC and global and regional Health Authorities, as applicable, prior to implementation except when required to mitigate immediate safety risks or when the changes involve only logistical or administrative revisions.

Amendments may necessitate that the informed consent and other study-related material be revised. If the consent form is revised, all subjects currently enrolled in the study must sign the approved, revised informed consent (re-consent), as required by the IRB/IEC.





Page 1 of 1

100% of the time, the system is able to correctly identify the target class for the test samples.


113. *Leptodora* (Leptodora) *hirsutissima* (L.) Schlecht. (1854) 100. (1854) 100. (1854) 100.

© 2019 Pearson Education, Inc.

11. **What is the primary purpose of the study?** (check all that apply)



[REDACTED]




Topic	Percentage
The concept of a 'smart city'	85%
Smart city technologies	90%
Smart city infrastructure	88%
Smart city governance	100%
Smart city data	82%
Smart city transportation	95%
Smart city energy	80%
Smart city waste management	92%

Page 1 of 1

For more information, contact the Office of the Vice President for Research and Economic Development at 319-273-2500 or [research@uiowa.edu](mailto:research@uiowa.edu).

For more information, contact the Office of the Vice President for Research and Economic Development at 319-273-2500 or [research@uiowa.edu](mailto:research@uiowa.edu).

For more information, contact the Office of the Vice President for Research and Economic Development at 319-273-2500 or [research@uiowa.edu](mailto:research@uiowa.edu).

10.1007/s00339-017-0382-2

For more information, contact the Office of the Vice President for Research and the Office of the Vice President for Student Affairs.

1. **What is the primary purpose of the study?** (Please select one)

A large black rectangle is positioned on the left side of the frame. To its right, a smaller black rectangle is placed vertically. The two rectangles are separated by a small gap.

1

A large black rectangle is positioned on the right side of the frame. To its left, a smaller black rectangle is partially visible, creating a sense of depth or a window frame.

## Itemized Changes:

The figure is a horizontal bar chart with four main categories on the y-axis. Category A has 10 sub-categories. Category B has 10 sub-categories. Category C has 10 sub-categories. Category D has 10 sub-categories. The x-axis represents the count of data points, ranging from 0 to 100. The bars are black and the chart is on a white background.

Category	Sub-Category	Count
A	1	10
	2	10
	3	10
	4	10
	5	10
	6	10
	7	10
	8	10
	9	10
	10	10
B	1	10
	2	10
	3	10
	4	10
	5	10
	6	10
	7	10
	8	10
	9	10
	10	10
C	1	10
	2	10
	3	10
	4	10
	5	10
	6	10
	7	10
	8	10
	9	10
	10	10
D	1	10
	2	10
	3	10
	4	10
	5	10
	6	10
	7	10
	8	10
	9	10
	10	10

## **5 INTRODUCTION**

### **5.1 Rationale and Background**

[REDACTED]

New SiHy materials continue to be developed possessing unique material properties and superior oxygen transmissibility over contact lenses made with conventional hydrogel materials. A new lens, [REDACTED] is being developed in an effort to maintain sustained performance [REDACTED]

[REDACTED]

In this clinical study, the performance of the investigational [REDACTED] lens will be compared to the commercially available Biofinity lens in a contralateral design with 2 days (including 1 night of extended wear) exposure. The intended use of this contact lens is for vision correction. Therefore, the objective measurement of VA is planned as the primary variable for the comparison with the Biofinity lens. [REDACTED]

[REDACTED]

### **5.2 Purpose of the Study**

The purpose of this clinical study is to evaluate the on-eye performance of the investigational [REDACTED] lens compared to the commercially available Biofinity lens following 1 night of extended wear. At the end of the study, a clinical study report will be prepared, in accordance with applicable regulatory requirements and standards.

[REDACTED]

### 5.3 Risks and Benefits

Contact lenses may offer improved peripheral vision and the convenience of not wearing spectacles. Material properties and design characteristics of [REDACTED] lenses are features consistent with successful contact lens wear. Based upon nonclinical testing and documented rationale for applicability of test results, [REDACTED] lenses are assessed to be non-toxic and biocompatible for on-eye use.

Biofinity lenses are for up to 6 nights/7 days of continuous wear. Further details on any known potential risks and benefits can be found in the package insert.

A summary of the known potential risks and benefits associated with [REDACTED] can be found in the IB.

[REDACTED]

[REDACTED]

[REDACTED] In general, the risks with [REDACTED] are anticipated to be similar to other marketed weekly/monthly soft contact lenses.

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

## 6 STUDY OBJECTIVES

### 6.1 Primary Objective

The primary objective is to obtain [REDACTED] insights to safety and performance of [REDACTED] soft contact lenses when worn in an extended wear modality compared to the Biofinity soft contact lenses.

### 6.2 Secondary Objective

Not Applicable.

### 6.3 Exploratory Objective

A horizontal bar chart illustrating the distribution of a variable across 15 categories. The bars are black and of varying lengths, representing the magnitude of the variable for each category. The categories are ordered from highest to lowest value.

Category	Value
1	100
2	95
3	90
4	85
5	80
6	75
7	70
8	65
9	60
10	55
11	50
12	45
13	40
14	35
15	30

## 6.4 Safety Objective

**Table 6–2** Safety Objective

<u>Objective(s)</u>	<u>Endpoint(s)</u>
█ evaluation of performance and safety profile of the test lens.	AEs
	Device deficiencies
	Biomicroscopy

## 7 INVESTIGATIONAL PLAN

## 7.1 Study Design

This will be a prospective, randomized, controlled, double-masked, contralateral wear clinical trial.

Approximately 1 site in the US will enroll approximately 12 subjects. Subjects will be randomized to wear the test [REDACTED] lens in one eye and the control Biofinity lens in the other eye. Subjects will be expected to attend 4 visits: Baseline/Dispense, Follow-up just prior to sleeping, Follow-up upon awakening/eye opening the morning following overnight lens wear, and Follow-up/Exit approximately 1 hour after awakening.

Following randomization during the Baseline/Dispense Visit, study lenses will be dispensed to the subject. All study contact lenses are to be worn overnight for 1 night only. Subjects will be required to retire for approximately 8 hours (7-10 hours) of closed eye sleep with lenses in place.

The study is expected to take approximately 3 weeks for completion.

## 7.2 Rationale for Study Design

[REDACTED]

Results from this feasibility study may be used to further guide product development [REDACTED]

[REDACTED]

## 7.3 Rationale for Duration of Treatment/Follow-Up

[REDACTED]

## 7.4 Rationale for Choice of Control Product

Biofinity contact lenses were chosen as the control product because this lens is a proper predicate device to compare to [REDACTED] lenses in regards to effectiveness and safety. Both [REDACTED] and Biofinity are silicone hydrogel lenses. Biofinity lenses are to be prescribed for up to 6 nights/7 days of continuous wear [REDACTED]

[REDACTED] The Biofinity spherical soft contact lenses are indicated for the optical correction of refractive ametropia (myopia and hyperopia) in phakic or aphakic persons with non-diseased eyes.

## 7.5 Data Monitoring Committee

Not applicable.

## 8 STUDY POPULATION

The study population consists of adult male and female subjects (aged 18 or over), with non-diseased eyes, who require optical correction for refractive ametropia. It is aimed to enroll approximately 12 subjects in approximately 1 US site. Estimated time needed to recruit subjects for the study is approximately 2 weeks. The intended study population consists of volunteer subjects aged 18 or over who are soft contact lens wearers, excluding Biofinity habitual wearers, have at least 3 months of contact lens wearing experience, and who wear their habitual lenses at least 5 days per week in either a daily wear or extended wear modality. Subjects who wear their habitual lenses in a daily wear modality (ie, not extended wear) should wear their habitual lenses at least 12 hours per day. [REDACTED]

[REDACTED] Pregnant and breastfeeding women are excluded from this study.

### 8.1 Inclusion Criteria

Written informed consent must be obtained before any study specific assessment is performed. Upon signing informed consent, the subject is considered enrolled in the study.

Subjects eligible for inclusion in this study must fulfill **all** of the following criteria:

1.	Subject must be at least 18 years of age and able to understand and sign an IRB/IEC approved Informed Consent form.
2.	Willing to attend all scheduled study visits as required per protocol.
3.	Successful wear of spherical soft contact lenses in both eyes during the past 3 months for a minimum of 5 days per week in either a daily wear or extended wear modality. Subjects who wear their habitual lenses in a daily wear modality (ie, not extended wear) should wear their habitual lenses at least 12 hours per day. Subjects who wear their habitual lenses in an extended wear modality should wear their habitual lenses a minimum of 1 night per week.
4.	[REDACTED]
5.	Best spectacle corrected VA 20/25 or better in each eye.
6.	[REDACTED]

## 8.2 Exclusion Criteria

Subjects fulfilling **any** of the following criteria are not eligible for participation in this study.

1.	Any anterior segment infection, inflammation, or abnormality or disease (including systemic) that contraindicates contact lens wear, as determined by the Investigator.
2.	Any use of systemic or ocular medications for which contact lens wear could be contraindicated, as determined by the Investigator.
3.	History of ocular or intraocular surgery, including refractive surgery for which contact lens extended wear could be contraindicated, as determined by the Investigator and/or history of irregular cornea.
4.	Biomicroscopy findings at screening that are moderate (Grade 3) or higher and/or corneal vascularization that is mild (Grade 2) or higher; or presence of corneal infiltrate(s).
5.	Current or history of pathologically dry eye in either eye that, in the opinion of the Investigator, would preclude contact lens wear.
6.	Current or history of herpetic keratitis in either eye.
7.	Eye injury in either eye within twelve weeks immediately prior to enrollment for this trial.
8.	Current or history of intolerance, hypersensitivity, or allergy to any component of the study products.
9.	Current Biofinity lens wearer.
10.	Any use of topical ocular medications, and artificial tear or rewetting drops that would require instillation during contact lens wear.
11.	Currently pregnant or breast-feeding.
12.	The Investigator, his/her staff, family members of the Investigator, family members of the Investigator's staff, or individuals living in the households of the aforementioned persons may not participate in the study.

13.	Participation of the subject in a clinical trial within the previous 30 days or currently enrolled in any clinical trial.
-----	---

### 8.3 Rescreening of Subjects

Rescreening of subjects is not allowed in this study.

## 9 TREATMENTS ADMINISTERED

### 9.1 Investigational Product(s)

*Test Product:* [REDACTED] soft contact lenses

*Control Product:* Biofinity (comfilcon A) soft contact lenses

**Table 9-1** Test Product

Test Product	[REDACTED] soft contact lenses (LID011121)
Manufacturer	Alcon Laboratories, Inc. 6201 South Freeway Fort Worth, Texas 76134-2099 USA
Indication for use and intended purpose in the current study	The intended use of this contact lens is for vision correction.
Product description and parameters available for this study	[REDACTED]
Formulation	Silicone Hydrogel. Additional details can be found in the IB [REDACTED]
Usage	<ul style="list-style-type: none"> <li>• Wear: <ul style="list-style-type: none"> <li>○ Extended Wear – 1 night</li> <li>○ Contralateral</li> </ul> </li> </ul>

	<ul style="list-style-type: none"> <li>• Exposure: ~2 days with approximately 8 hours of sleep (7-10 hours) between Visits 2 and 3.</li> <li>• Lens Care: N/A</li> <li>• Additional details can be found in the MOP.</li> </ul>
Number/Amount of product to be provided to the subject	One lens will be dispensed to the subject on Visit 1.
Packaging description	Blister foil pack
Labeling description	<ul style="list-style-type: none"> <li>• Lens Foil label includes: <ul style="list-style-type: none"> <li>- material name and/or identifier</li> <li>- base curve</li> <li>- diameter</li> <li>- manufacturing protocol number</li> <li>- packing solution</li> <li>- power</li> <li>- lot number</li> <li>- expiration date</li> <li>- content statement</li> <li>- investigational device statement</li> <li>- Sponsor information</li> </ul> </li> <li>• Provided in boxes of up to 6 lenses per power per box, identified with the following: <ul style="list-style-type: none"> <li>- color-coded label stating the protocol number</li> <li>- material identifier</li> <li>- power</li> <li>- investigational use only statement</li> <li>- tracking number</li> </ul> </li> </ul>
Storage conditions	Stored at room temperature.
Supply	All study lenses will be provided to the site. The site will dispense the study lenses to the subject at Visit 1 only. Replacement lenses are not allowed after final dispense of the study lenses to the subject.

Table 9–2

Control Product

Control Product	Biofinity (comfilcon A) soft contact lenses
Manufacturer	CooperVision
Indication for Use	The intended use of this contact lens is for vision correction.
Product description and parameters available for this study	<ul style="list-style-type: none"> <li>• Material: comfilcon A</li> <li>• Water content: 48%</li> <li>• Power range: -1.00 to -6.00 D, 0.25 or 0.50 D steps</li> <li>• Base curve: 8.6 mm</li> <li>• Diameter: 14.0 mm</li> </ul>
Formulation	Silicone Hydrogel. Additional details can be found in the Biofinity package insert.
Usage	<ul style="list-style-type: none"> <li>• Wear: <ul style="list-style-type: none"> <li>○ Extended Wear – 1 night</li> <li>○ Contralateral</li> </ul> </li> <li>• Exposure: ~2 days with approximately 8 hours of sleep (7-10 hours) between Visits 2 and 3.</li> <li>• Lens Care: N/A</li> <li>• Additional details can be found in the MOP.</li> </ul>
Number/Amount of Product to be Provided to the subject	One lens will be dispensed to the subject on Visit 1.
Packaging description	Blister foil pack in commercial packaging
Labeling description	<ul style="list-style-type: none"> <li>• Lens Foil label includes: <ul style="list-style-type: none"> <li>- material name</li> <li>- base curve</li> <li>- diameter</li> <li>- packing solution</li> <li>- power</li> <li>- lot number</li> <li>- expiration date</li> <li>- content statement</li> </ul> </li> </ul>

	<ul style="list-style-type: none"><li>- Manufacturer information</li><li>- Country of origin</li></ul>
Storage conditions	Stored at room temperature.
	[REDACTED]

## 9.2 Other Medical Device or Medication Specified for Use During the Study

No other medical devices or medications are required to be used in conjunction with the treatments during the clinical study, including lubricating or re-wetting drops.

## 9.3 Treatment Assignment / Randomization

Subjects will be randomized in a 1:1 ratio to receive treatment with [REDACTED] or Biofinity in 1 eye and the other lens in the fellow eye, as indicated below:

Sequence 1: [REDACTED] (OD) / Biofinity (OS)

Sequence 2: Biofinity (OD) / [REDACTED] (OS)

Only after signing the ICF, a subject will be assigned a subject number by the electronic data capture system.

A randomization list will be generated using a validated system that automates the random assignment of treatment arms to randomization numbers in the specified ratio. Subjects will be assigned a treatment (lens sequence) according to the randomization list [REDACTED]


[REDACTED] The randomization list will be generated and maintained by the Study Sponsor.

At Visit 1, all eligible subjects will be randomized via the EDC/IRT integration system to one of the lens sequences. The Investigator or delegate will access the respective system after confirming that the subject meets all the eligibility criteria. A randomization number will be automatically assigned to the subject according to the subject randomization list but will not be communicated to the site user. The EDC/IRT integration system will inform the site user of the treatment (lens sequence) assignment to be dispensed to the subject.

## 9.4 Treatment masking

This study is double-masked, with subjects randomized to use [REDACTED] and Biofinity by eye for the duration of the treatment period.





[REDACTED]. Unmasking will occur only after all planned study data have been validated, and the database locked.

Masked study personnel must avoid seeking information that may compromise masking.

[REDACTED]

[REDACTED]

[REDACTED]

In the event of a medical emergency where the knowledge of subject treatment is required, an individual Investigator will have the ability to unmask the treatment assignment for a specific subject after contacting an appropriate Study Sponsor representative if time allows.

Unmasking must be done according to the instructions provided for the study IRT system.

## 9.5 Accountability Procedures

Upon receipt of the IP, the Investigator or delegate must conduct an inventory.

[REDACTED]

[REDACTED] Throughout the study, the Investigator or delegate must maintain records of IP dispensation and collection for each subject. This record must be made available to the study monitor for the purposes of verifying the accounting of IP supplies. Any discrepancies and/or deficiencies between the observed disposition and the written account must be recorded along with an explanation. All IPs sent to the Investigator must be accounted for by Study Sponsor personnel, and in no case be used in an unauthorized situation.

It is the Investigator's responsibility to ensure that:

- All study products are accounted for and not used in any unauthorized manner
- All used foils and unused supplies are returned by each subject

- All unused products (supplied by the Study Sponsor) are available for return to the Study Sponsor, as directed
- Any study lenses associated with a device deficiency or with any product-related adverse event (ie, ADE or SADE) are returned to the Study Sponsor for investigation, unless otherwise directed by the Sponsor. Refer to Section 11 of this protocol for additional information on the reporting of device deficiencies and AEs and the return of study products associated with these events.

The Investigator is responsible for proper disposition of all unused IPs at the conclusion of the study, according to the instructions provided in the MOP.

## 9.6 Changes to concomitant medications, treatments/ procedures

Changes in concomitant treatments after Visit 1 are not allowed unless needed for the proper medical care and treatment of the subject for a specific medical condition.

After the subject is enrolled into the study, the Investigator must instruct the subject to notify the study site about:

- Any new medications
- Alterations in dose or dose schedules for current medications,
- Any medical procedure or hospitalization that occurred or is planned
- Any non-drug therapies (including physical therapy and blood transfusions).

The Investigator must document this information in the subject's case history source documents.

## 10 STUDY PROCEDURES AND ASSESSMENTS

Visit	Window
Visit 1, Day 1: Baseline/Dispense	N/A
Visit 2, Day 1: Follow-up prior to sleeping	<p>≥ 4 hours after Dispense</p> <p><i>Instruct subject to retire on Day 1 so that he/she sleeps in the lenses ~7-10 hours.</i></p>

Visit 3, Day 2: Follow-up upon eye opening	<p>~7:30 am; Follow-up, 0 to 10 mins upon eye opening</p> <p><i>Visit 3 to occur immediately (0-10 mins) upon eye-opening on Day 2. Subjects will be escorted to the slit-lamp with eyes closed.</i></p>
Visit 4, Day 2: 1 Hour Follow-up/Exit	-30/+60 mins after Visit 3

Unscheduled visits and Early Termination Visits are allowed, if necessary.

At the Baseline/Dispense Visit, study lenses will be dispensed contralaterally to the subject. All study lenses will be worn in an extended wear modality for 1 night. Problem lenses (if any) will not be discarded but collected by the subject and returned to the investigational site. VA will be measured at all visits, and any decrease of 2 or more lines from the Dispense Visit to any follow-up visit should be explained by the Investigator.

[REDACTED]

## 10.1 Informed Consent and Screening

The Investigator or delegate must explain the purpose and nature of the study, and have the subject read, sign, and date the IRB/IEC-approved informed consent document. The subject must sign the ICF BEFORE any study-specific procedures or assessments can be performed, including study-specific screening procedures. Additionally, have the individual obtaining consent from the subject and a witness, if applicable, sign and date the informed consent document.

The Investigator or delegate must provide a copy of the signed document to the subject and place the original signed document in the subject's chart, or provide documentation as required by local regulations.

## 10.2 Description of Study Procedures and Assessments

Detailed descriptions of assessments and procedures are provided in the MOP. The Investigator is responsible for ensuring responsibilities for all procedures and assessments are delegated to appropriately qualified site personnel.

### 10.2.1 Demographics

Obtain demographic information including age, race, ethnicity, and sex.

## 10.2.2 Medical History

Collect medical history information, including information on all medications used within the past 30 days. Include herbal therapies, vitamins, and all over-the-counter as well as prescription medications. Throughout the subject's participation, obtain information on any changes in medical health and/or the use of concomitant medications.

### 10.2.3 Investigational Product compliance

Review subject compliance with the IP usage [REDACTED] and collect all used and unused study IPs and other products that were dispensed.

#### 10.2.4 Adverse Event Collection: Safety Assessment

Assess and record any AEs that are observed or reported, including those associated with changes in concomitant medication dosing since the previous visit.

## 10.2.5 Slit-Lamp Biomicroscopy: Safety Assessment

Slit-lamp examination of the cornea, iris/anterior chamber and lens must be performed in both eyes before instillation of any diagnostic eye drops.

## 10.2.6 Device Deficiencies: Safety Assessment

Assess and record any Device Deficiencies that are reported or observed since the previous visit. Requirements for reporting device deficiencies in the study can be found in Section 11.

11. **What is the primary purpose of the *Journal of Clinical Oncology*?**

Term	Percentage
GMOs	~75%
Organic	~95%
Natural	~90%
Artificial	~65%
Organic	~95%
Natural	~90%
Artificial	~65%
Organic	~95%
Natural	~90%
Artificial	~65%



## 10.3 Unscheduled Visits

If a subject visit occurs between any regularly scheduled visits, this visit must be documented as an Unscheduled Visit. During all unscheduled visits, the Investigator must conduct the following procedures:

- Collect AE information
- Record changes in medical condition or concomitant medication
- Collect Device Deficiency information, as applicable
- Assess and record VAs
- Perform biomicroscopy (assessments with or without lenses, as applicable)

The Investigator may perform additional procedures for proper diagnosis and treatment of the subject. The Investigator must document this information in the subject's case history source documents.

If during an Unscheduled Visit the subject is discontinuing the IP or discontinuing from the study, the Investigator must conduct Exit procedures according to Table 3-1, Schedule of Study Procedures and Assessments, as possible.

## 10.4 Discontinued Subjects

### 10.4.1 Screen Failures

Screen failures are subjects who were excluded from the study after signing the informed consent, not meeting the inclusion/exclusion criteria, and prior to randomization to product/dispense of study product.

The Investigator must document the reason for screen failure in the subject's case history source documents.

Subject numbers must not be re-used.

### 10.4.2 Discontinuations

Discontinued subjects are individuals who voluntarily withdraw or are withdrawn from the study by the Investigator after signing the informed consent, including screen failures.

Subject numbers of discontinued subjects must not be re-used.

Subjects may discontinue from study or study treatment at any time for any reason. Subjects may also be discontinued from study treatment at any time if, in the opinion of the Investigator, continued treatment poses a risk to their health.

For subjects discontinuing from the study, the Investigator must complete all Exit procedures according to Table 3-1, Schedule of Study Procedures and Assessments, if the subject is willing and able, and if in the opinion of the Investigator it is safe for the subject to do so.

The Investigator must document the reason for study or treatment discontinuation in the subject's case history source documents.

To ensure the safety of all subjects who discontinue early, Investigators must assess each subject and, if necessary, advise them of any therapies and/or medical procedures that may be needed to maintain their health.

#### **10.4.3 Schedule of Procedures and Assessments for Subjects Discontinued from Investigational Product**

Other than screen failures, if a subject discontinues from the study the subject should undergo an Early Exit Visit. Refer to Table 3-1, Schedule of Study Procedures and Assessments.

### **10.5 Clinical Study Termination**

The Study Sponsor reserves the right to close the investigational site or terminate the study in its entirety at any time.

If the clinical study is prematurely terminated or suspended by the Study Sponsor:

- The Study Sponsor must:
  - Immediately notify the Investigator(s) and subsequently provide instructions for study termination.
  - Inform the Investigator and the regulatory authorities of the termination/suspension and the reason(s) for the termination/suspension.
- The Investigator must:
  - Promptly notify the IRB/IEC of the termination or suspension and of the reasons.

- Provide subjects with recommendations for post-study treatment options as needed.

The Investigator may terminate the site's participation in the study for reasonable cause.

### 10.5.1 Follow-up of subjects after study participation has ended

Following this study, the subject will return to their eye care professional for their routine eye care.


## 11 ADVERSE EVENTS AND DEVICE DEFICIENCIES

### 11.1 General Information

An AE is any untoward medical occurrence, unintended disease or injury, or untoward clinical signs (including abnormal laboratory findings) in subjects, users, or other persons, whether or not related to the investigational medical device (test article). Refer to the Glossary of Terms and figures below for categories of AEs and SAEs.


**Figure 11-1**

**Categorization of All Adverse Events**



**Figure 11-2**

**Categorization of All Serious Adverse Events**



***Serious Adverse Events***

In addition to reporting all AEs (serious and non-serious) meeting the definitions, the Investigator must report any occurrence of the following as an SAE:

- An ocular infection including a presumed infectious ulcer with any of the following characteristics\*:
  - Central or paracentral location
  - Penetration of Bowman's membrane
  - Infiltrates > 2 mm diameter
  - Iritis
  - Increase in intraocular pressure
  - Culture positive for microorganisms
  - Increasing size or severity at subsequent visits
- Any central or paracentral corneal event (such as neovascularization) that results in permanent opacification
- Hypopyon

- Hyphema
- Neovascularization within the central 6 mm of the cornea
- Permanent vision loss as defined by loss of 2 or more lines of BCVA from enrollment visit that fails to resolve
- Uveitis (anterior, intermediate, or posterior)
- Corneal abrasion affecting  $\geq$  50% of corneal surface area

*\*NOTE: Culture samples (from the subject's eyes, lenses, etc) must be taken [as described in the MOP; and documented in the narrative section(s) of the corresponding ADE-SAE eCRF], for any suspected ocular infection, including infiltrates with overlying epithelial defect.*

### ***Significant Non-Serious Adverse Events***

A significant non-serious AE is a device-related, non-sight threatening AE that warrants discontinuation of any contact lens wear for greater than or equal to 2 weeks. In addition, the Investigator must report any occurrence of the following as a Significant Non- Serious Adverse Event:

- Peripheral non-progressive non-infectious ulcers
- All symptomatic corneal infiltrative events
- Corneal staining score greater than or equal to Grade 3 (Refer to MOP for grading scales)
- Temporary vision loss as defined by loss of 2 or more lines of BCVA from enrollment visit that persists for 2 or more weeks
- Neovascularization score greater than or equal to Grade 2 (Refer to MOP for grading scales)

*The above events are based upon the categories provided in the ISO 11980 and the US FDA Premarket Notification (510(k)) Guidance Document for Daily Wear Contact Lenses.*

### ***Device Deficiencies***

A device deficiency is inadequacy of a medical device with respect to its identity, quality, durability, reliability, safety, or performance. A device deficiency may or may not be associated with patient harm (ie, ADE or SADE); however, not all ADEs or SADEs are due to a device deficiency. The Investigator should determine the applicable category listed in the Device Deficiency eCRF for the identified or suspect device deficiency and report any patient harm separately. Examples of device deficiencies include the following:

- Failure to meet product specifications (eg, incorrect lens power/diameter/base curve/color)
- Lens cloudy
- Lens surface/edge defect
- Torn lens during handling/in pack
- Packaging deficit (eg, mislabeled product)
- Suspect product contamination
- Lack of performance

## 11.2 Monitoring for Adverse Events

At each visit, after the subject has had the opportunity to spontaneously mention any problems, the Investigator should inquire about AEs by asking the standard questions:

- “Have you had any health problems since your last study visit?”
- “Have there been any changes in the medicines you take since your last study visit?”

Changes in *any protocol-specific parameters and/or questionnaires* evaluated during the study are to be reviewed by the Investigator. Any untoward (unfavorable and unintended) change in *a protocol-specific parameter or questionnaire response* that is clinically relevant, in the opinion of the Investigator, is to be reported as an AE. These clinically relevant changes will be reported regardless of causality.

## 11.3 Procedures for Recording and Reporting

AEs are collected from the time of informed consent. Any pre-existing medical conditions or signs/symptoms present in a subject prior to the start of the study (ie, before informed consent is signed) are not considered AEs in the study and should be recorded in the Medical History section of the eCRF.

In addition, temporary lens awareness or visual changes during the fitting process are not considered AEs if the Investigator assesses that the symptom(s) can reasonably resolve within the anticipated adaptation period.

- ADEs or SAEs are documented on the *Serious Adverse Event and Adverse Device Effect* eCRF within 24 hours of the Investigator’s or site’s awareness.
- Device deficiencies are documented on the *Device Deficiency* eCRF within 24 hours of the Investigator’s or site’s awareness.

- A printed copy of the completed *Serious Adverse Event and Adverse Device Effect* and/or *Device Deficiency* eCRF must be included with product returns.
- Additional relevant information after initial reporting must be entered into the eCRF as soon as the data become available.
- Document any changes to concomitant medications on the appropriate eCRFs.
- Document all relevant information from Discharge Summary, Autopsy Report, Certificate of Death, etc, if applicable, in narrative section of the *Serious Adverse Event and Adverse Device Effect* eCRF.

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

Any AEs and device deficiencies for non-study marketed devices/products, if applicable, will be considered and processed as spontaneous (following the postmarket vigilance procedures) and should be communicated to the device's/product's manufacturer as per local requirements.

Study Sponsor representatives may be contacted for any protocol related question and their contact information is provided in the MOP that accompanies this protocol.

Further, depending upon the nature of the AE or device deficiency being reported, the Study Sponsor may request copies of applicable portions of the subject's medical records. The Investigator must also report all AEs and device deficiencies that could have led to a SADE according to the requirements of regulatory authorities or IRB/IEC.

### **Intensity and Causality Assessments**

Where appropriate, the Investigator must assess the intensity (severity) of the AE based upon medical judgment with consideration of any subjective symptom(s), as defined below:

#### ***Intensity (Severity)***

Mild            An AE is mild if the subject is aware of but can easily tolerate the sign or symptom.

Moderate An AE is moderate if the sign or symptom results in discomfort significant enough to cause interference with the subject's usual activities.

Severe An AE is severe if the sign or symptom is incapacitating and results in the subject's inability to work or engage in their usual activities.

For every AE in the study, the Investigator must assess the causality (Related or Not Related to the medical device or study procedure). An assessment of causality will also be performed by Study Sponsor utilizing the same definitions, as shown below:

### ***Causality***

Related An AE classified as related may be either definitely related or possibly related where a direct cause and effect relationship with the medical device or study procedure has not been demonstrated, but there is a reasonable possibility that the AE was caused by the medical device or study procedure.

Not Related An AE classified as not related may either be definitely unrelated or simply unlikely to be related (ie, there are other more likely causes for the AE).

The Study Sponsor will assess the AEs and may upgrade the Investigator's assessment of seriousness and/or causality. The Study Sponsor will notify the Investigator of any AEs that are upgraded from non-serious to serious or from unrelated to related.

Additionally, the Study Sponsor shall immediately conduct an evaluation of any unanticipated adverse device effect, including anticipated adverse events that occur in unanticipated severity or frequency. The results of this evaluation will be reported to FDA, the IRB, and participating Investigators within 10 working days upon receiving notification of the effect.

## **11.4 Return product analysis**

IP associated with device deficiencies and/or product related AEs [ie, ADE or SADE] will be returned for investigation as detailed in the MOP.

## **11.5 Unmasking of the Study Treatment**

Masked information on the identity of the assigned medical device should not be disclosed during the study. If the treatment code needs to be broken in the interest of subject safety, the Investigator is encouraged to contact an appropriate Study Sponsor representative prior to unmasking the information if there is sufficient time. Dependent upon the individual

circumstances (ie, medical emergency), the code may be broken prior to contact with the Study Sponsor. The Study Sponsor must be informed of all cases in which the code was broken and of the circumstances involved. Additionally, the Study Sponsor may be required to unmask the information in order to fulfill expedited regulatory requirements.

## **11.6 Follow-Up of Subjects with Adverse Events**

The Investigator is responsible for adequate and safe medical care of subjects during the study and for ensuring that appropriate medical care and relevant follow-up procedures are maintained after the study.

The Investigator should provide the Study Sponsor with any new safety information (which includes new AEs and changes to previously reported AEs) that may affect the safety evaluation of the device. For AEs that are unresolved/ ongoing at time of subject exit from study, any additional information received at follow-up should be documented in the eCRFs up to study completion (ie, database lock).

Any additional data received up to 1 month after subject discontinuation or exit must be documented and available upon the Study Sponsor's request. All complaints received after this time period will be considered and processed as spontaneous (following the postmarket vigilance procedures) and should be communicated to the medical device's manufacturer as per local requirements, as applicable.

The Investigator should also report complaints on non-Alcon products directly to the manufacturer as per the manufacturer's instructions or local regulatory requirements.

## **11.7 Pregnancy in the Clinical Study**

Pregnancy is not reportable as an AE; however, complications may be reportable and will be decided on a case-by-case basis. Should a woman become pregnant during study participation, the pregnancy will be documented on the Medical history eCRF.

## **12 ANALYSIS PLAN**

Continuous variables will be summarized using the number of observations, mean, standard deviation (SD), median, minimum and maximum. Categorical variables will be summarized with counts and percentages from each category. Any deviations to the analysis plan will be updated during the course of the study as part of a protocol amendment or will be detailed in the clinical study report.

## 12.1 Subject Evaluability

Final subject evaluability must be determined prior to breaking the code for masked treatment (lens sequence) assignment and locking the database, based upon the Deviations and Evaluability Plan.

## 12.2 Analysis Sets

Only 1 analysis set will be defined, namely the safety analysis set. It will include all eyes exposed to a study lens evaluated in this study.

Eyes will be analyzed based upon the lens exposed at the time of the clinical assessment.

## 12.3 Demographic and Baseline Characteristics

Demographic information and habitual lens information will be presented by lens sequence and overall.

## 12.4 Effectiveness Analyses

### 12.4.1 Analysis of Primary Effectiveness Endpoint(s)

The primary objective is to obtain [REDACTED] insights to safety and performance of [REDACTED] soft contact lenses when worn in an extended wear modality compared to the Biofinity soft contact lenses.

The primary endpoint is VA with study lenses, collected by eye.

#### 12.4.1.1 Statistical Hypotheses

No hypothesis testing of the primary effectiveness endpoint is planned.

#### 12.4.1.2 Analysis Methods

Summary statistics will be provided.

[REDACTED]

- [REDACTED]
- [REDACTED]
- [REDACTED]
- [REDACTED]

A bar chart illustrating the number of publications per year from 1990 to 2010. The y-axis represents the year, and the x-axis represents the number of publications. The bars are black with thin white outlines. The chart shows a general upward trend with some fluctuations.

Year	Number of Publications
1990	10
1991	12
1992	15
1993	18
1994	22
1995	25
1996	28
1997	32
1998	35
1999	38
2000	42
2001	45
2002	48
2003	52
2004	55
2005	58
2006	62
2007	65
2008	68
2009	72
2010	75

## 12.5 Handling of Missing Data

All data obtained in evaluable subjects/eyes will be used. No imputation for missing values will be carried out.

## 12.6 Safety Analyses

The safety endpoints are:

- AEs
- Biomicroscopy Findings
- Device Deficiencies

There are no safety hypotheses planned in this study. The focus of the safety analysis will be a comprehensive descriptive assessment of occurrence of AE as well as the other listed parameters.

Descriptive summaries (counts and percentages) for ocular and nonocular AEs will be presented by Medical Dictionary for Regulatory Activities Preferred Terms. A listing containing details of the AEs will also be provided.

Each biomicroscopy parameter will be tabulated by its grade.

Frequency for each device deficiency category will be presented and a supporting listing will be provided.

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED] t.

## **13 DATA HANDLING AND ADMINISTRATIVE REQUIREMENTS**


### **13.1 Subject Confidentiality**

The Investigator must ensure that the subject's anonymity is maintained throughout the course of the study. In particular, the Investigator must keep an enrollment log with confidential identifying information that corresponds to the subject numbers and initials of each study participant. At the end of the clinical study, the Study Sponsor will collect a copy of the enrollment log ***without any identifying subject information***. All documents submitted to the Study Sponsor will identify the subjects exclusively by number and demographic information. No other personally identifying information will be transmitted to the Study Sponsor.

[REDACTED]

[REDACTED]

[REDACTED]



### **13.2 Completion of Source Documents and Case Report Forms**

The nature and location of all source documents will be identified to ensure that original data required to complete the CRFs exist and are accessible for verification by the site monitor, and all discrepancies shall be appropriately documented via the query resolution process. Site monitors are appointed by the Study Sponsor and are independent of study site staff.

If electronic records are maintained, the method of verification must be determined in advance of starting the study.

At a minimum, source documents include the following information for each subject:

- Subject identification (name, sex, race/ethnicity)
- Documentation of subject eligibility
- Date of informed consent
- Dates of visits
- Documentation that protocol specific procedures were performed
- Results of study parameters, as required by the protocol
- IP accountability records
- Documentation of AEs and other safety parameters (if applicable)
- Records regarding medical histories and the use of concomitant therapies prior to and during the study
- Date of study completion and reason for early discontinuation, if applicable

It is required that the author of an entry in the source documents be identifiable. Direct access to source documentation (medical records) must be allowed for the purpose of verifying that the data recorded on the CRF are consistent with the original source data.

Only designated individuals at the site will complete the CRFs. The CRFs must be completed at regular intervals following the clinical study visit schedule. It is expected that all data reported have corresponding entries in the source documents. The Principal Investigator is responsible for reviewing and certifying that the CRFs are accurate and complete. [REDACTED]

[REDACTED]

[REDACTED]

### **13.3 Data Review and Clarifications**

A review of CRF data to the subject's source data will be completed by the site monitor to ensure completeness and accuracy. After the CRFs have been completed, additional data clarifications and/or additions may be needed as a result of the data cleaning process. Data clarifications are documented and are part of each subject's CRF.

### **13.4 Sponsor and Monitoring Responsibilities**

The Study Sponsor will designate a monitor to conduct the appropriate site visits at the appropriate intervals according to the study monitoring plan. The clinical investigation will be monitored to ensure that the rights and well-being of the subjects are protected, the reported data are accurate, complete, and verifiable from the source documents, and the study is conducted in compliance with the current approved protocol (and amendments[s], if applicable), with current GCP, and with applicable regulatory requirements.

The site may not screen subjects or perform the informed consent process on any subject until it receives a notification from an appropriate Study Sponsor representative that the site may commence conducting study activities. Monitoring will be conducted periodically while the clinical study is ongoing. Monitoring methods may include site visits, telephone, written and fax correspondence. Close-out visits will take place after the last visit of the last subject at the site.

A Coordinating Investigator may be identified by the Study Sponsor to review and endorse the final study report. In cases where a Coordinating Investigator is engaged, the Study Sponsor will select the Coordinating Investigator based upon their experience, qualifications, active study participation, and their willingness and availability to take on this role.

Term	Percentage
GMOs	85%
Organic	80%
Natural	75%
Artificial	65%
Organic	82%
Natural	78%
Artificial	68%
Organic	83%
Natural	79%
Artificial	66%
Organic	81%
Natural	77%
Artificial	64%

## 13.5 Regulatory Documentation and Records Retention

The Investigator is required to maintain up-to-date, complete regulatory documentation as indicated by the Study Sponsor and the Investigator's files will be reviewed as part of the ongoing study monitoring. Financial information is to be kept separately.

Additionally, the Investigator must keep study records and source documents consistent with the terms of the clinical study agreement with the Study Sponsor. If the Investigator retires, relocates, or for any other reason withdraws from responsibility of keeping the study records, then the Study Sponsor must be notified and suitable arrangements made for retention of study records and source documents needed to comply with national and international regulations.

## 13.6 Quality Assurance and Quality Control

The Study Sponsor will secure agreement from all involved parties to ensure direct access to all study related sites, source data and documents, and reports for the purpose of monitoring and auditing by the Study Sponsor, and inspection by domestic and foreign regulatory authorities. Quality control will be applied to each stage of data handling to ensure that all data are reliable and have been processed correctly. Agreements made by the Study Sponsor with the Investigator/Institution and any other parties involved in the clinical study will be provided in writing as part of the protocol or as a separate agreement.

## 14 ETHICS

This clinical study must be conducted in accordance with the ethical principles contained within:

- The Declaration of Helsinki, and in compliance with the ICH E6 GCP Consolidated Guideline, ISO 14155:2011, and the applicable US FDA 21 CFR Regulations.
- SOPs of the Study Sponsor and contract research organizations participating in the conduct of the clinical study and all other applicable regulations.
- Notifications and timelines for reporting protocol deviations should be based upon applicable Ethics Committee requirements

The Investigator must ensure that all personnel involved in the conduct of the study are qualified to perform their assigned responsibilities through relevant education, training, and experience. The Investigator and all clinical study staff must conduct the clinical study in compliance with the protocol. Deviations from this protocol, regulatory requirements and/or GCP must be recorded and reported to the Sponsor prior to database lock. If needed, corrective and preventive action should be identified, implemented, and documented within the study records. Use of waivers to deviate from the clinical protocol is prohibited.

Before clinical study initiation, this protocol, the informed consent form, any other written information given to subjects, and any advertisements planned for subject recruitment must be approved by an IRB/IEC. The Investigator must provide documentation of the IRB/IEC approval to the Study Sponsor. The approval must be dated and must identify the applicable protocol, amendments (if any), informed consent form, assent form (if any), all applicable recruiting materials, written information for subject, and subject compensation programs. The IRB/IEC must be provided with a copy of the IB, any periodic safety updates, and all other information as required by local regulation and/or the IRB/IEC. At the end of the study, the Investigator must notify the IRB/IEC about the study's completion. The IRB/IEC also must be notified if the study is terminated prematurely. Finally, the Investigator must report to the IRB/IEC on the progress of the study at intervals stipulated by the IRB/IEC.

Voluntary informed consent must be obtained in writing from every subject and the process shall be documented before any procedure specific to the clinical investigation is applied to the subject. The Investigator must have a defined process for obtaining consent. Specifically, the Investigator, or their delegate, must explain the clinical study to each potential subject and the subject must indicate voluntary consent by signing and dating the approved informed consent form. The subject must be provided an opportunity to ask questions of the Investigator, and if required by local regulation, other qualified personnel. The Investigator must provide the subject with a copy of the consent form written in a language the subject understands. The consent document must meet all applicable local laws and provide subjects with information regarding the purpose, procedures, requirements, and restrictions of the study, along with any known risks and potential benefits associated with the IP and the study,

the available compensation, and the established provisions for maintaining confidentiality of personal, protected health information. Subjects will be told about the voluntary nature of participation in the study and must be provided with contact information for the appropriate individuals should questions or concerns arise during the study. The subject also must be told that their records may be accessed by appropriate authorities and Sponsor-designated personnel. The Investigator must keep the original, signed copy of the consent (file in subject's medical records) and must provide a duplicate copy to each subject according to local regulations.

## **15 REFERENCES**

### **15.1 References applicable for all clinical studies**

- ISO 11980:2012 Ophthalmic optics - Contact lenses and contact lens care products - Guidance for clinical investigations
- ISO 14155:2011 Clinical investigation of medical devices for human subjects - Good clinical practice

#### **15.1.1 US references applicable for clinical studies**

- 21 CFR Part 11 - Electronic Records; Electronic Signatures
- 21 CFR Part 50 - Protection of Human Subjects
- 21 CFR Part 56 - Institutional Review Boards
- 21 CFR Part 812 - Investigational Device Exemptions
- 21 CFR Part 54 - Financial Disclosure by Clinical Investigators
- The California Bill of Rights

### **15.2 References for this clinical study**

Not applicable. There are no references.

**Alcon - Business Use Only** Protocol - Clinical

**Document:** TDOC-0054558

**Version:** 3.0; Most-Recent; Effective; CURRENT

**Status:** Effective

**Effective Date:** 07-May-2018

