

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 Sponsor: FARAPULSE
		Device: FARAPULSE CARDIAC ABLATION SYSTEM PLUS
		Protocol Nr: CS1335 REVISION B
		ICON Project Nr: 5326/0019

**Feasibility Study of the FARAPULSE Cardiac Ablation System Plus in the Treatment
of Persistent Atrial Fibrillation**
PersAFOne II
Statistical Analysis Plan

Versions	Date	Documents used	Statistical Lead	Senior Reviewer
V1.0	24/03/2023	<ul style="list-style-type: none"> - CS1335_Rev A to Rev B PersAFOne II Clinical Protocol_19May2022.pdf - PersAFOnell_Annnotated CRF_revC_21Sep2022.pdf - PersAFOnell_Annnotated CRF_CEC.pdf - Jan 2022_Endpoints for Persafone CZ II.xlsx 	Najeh DAABEK	Marie Christine REYMOND
V0.3	17/03/2023	<ul style="list-style-type: none"> - CS1335_Rev A to Rev B PersAFOne II Clinical Protocol_19May2022.pdf - PersAFOnell_Annnotated CRF_revC_21Sep2022.pdf - PersAFOnell_Annnotated CRF_CEC.pdf - Jan 2022_Endpoints for Persafone CZ II.xlsx 	Najeh DAABEK	Marie Christine REYMOND

Sponsor:
BOSTON SCIENTIFIC, INC.
Way Marlborough
MA 01752
USA
France

CRO:
ICON Clinical Research SARL
Immeuble Le Capitole
55 Avenue des Champs Pierreux
92000 Nanterre

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 Sponsor: FARAPULSE
		Device: FARAPULSE CARDIAC ABLATION SYSTEM PLUS
		Protocol Nr: CS1335 REVISION B
		ICON Project Nr: 5326/0019

V0.2	09/03/2023	<ul style="list-style-type: none"> - CS1335_Rev A to Rev B PersAFOne II Clinical Protocol_19May2022.pdf - PersAFOnell_Annnotated CRF_revC_21Sep2022.pdf - PersAFOnell_Annnotated CRF_CEC.pdf - Jan 2022_Endpoints for Persafone CZ II.xlsx 	Najeh DAABEK	Marie REYMOND	Christine REYMOND
V0.1	06/03/2023	<ul style="list-style-type: none"> - CS1335_Rev A to Rev B PersAFOne II Clinical Protocol_19May2022.pdf - PersAFOnell_Annnotated CRF_revC_21Sep2022.pdf - PersAFOnell_Annnotated CRF_CEC.pdf - Jan 2022_Endpoints for Persafone CZ II.xlsx 	Najeh DAABEK	Marie REYMOND	Christine REYMOND

Sponsor:
BOSTON SCIENTIFIC, INC.
 Way Marlborough
 MA 01752
 USA
 France

CRO:
ICON Clinical Research SARL
 Immeuble Le Capitole
 55 Avenue des Champs Pierreux
 92000 Nanterre

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 Sponsor: FARAPULSE
		Device: FARAPULSE CARDIAC ABLATION SYSTEM PLUS
		Protocol Nr: CS1335 REVISION B
		ICON Project Nr: 5326/0019

SIGNATURE PAGE

PERSAFONE II **V1.0**

FARAPULSE, INC

Name	Position	Date	Signature
Anitha Achyutha	Director, Clinical	<i>Anitha Achyutha</i> 24 Mar 2023 15:50:02 UTC (Z) REASON: I approve this document	Anitha Achyutha 24 Mar 2023 15:50:02 UTC (Z) REASON: I approve this document
Annelies Vanderper	EU Senior Manager Clinical Trials	<i>Annelies Vanderper</i> 24 Mar 2023 10:08:19 UTC (Z) REASON: I approve this document	Annelies Vanderper 24 Mar 2023 10:08:19 UTC (Z) REASON: I approve this document

ICON CLINICAL RESEARCH SARL

Name	Position	Date	Signature
Marie-Christine REYMOND	Senior Manager, Biostatistics	<i>Marie-Christine Reymond</i> 27 Mar 2023 07:26:30 UTC (Z) REASON: I approve this document	Marie-Christine Reymond 27 Mar 2023 07:26:30 UTC (Z) REASON: I approve this document
Najeh DAABEK	Biostatistician II	<i>Najeh Daabek</i> 24 Mar 2023 09:25:38 UTC (Z) REASON: I approve this document	Najeh Daabek 24 Mar 2023 09:25:38 UTC (Z) REASON: I approve this document

REASON: I approve this document

d0fe0951-4e46-41f5-9c58-01c89f6dec07

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

TABLE OF CONTENTS

List of abbreviations and definition of terms	5
1. Overview.....	6
1.1. Study Objective.....	6
1.2. Study Design.....	6
1.3. Study Plan.....	6
1.3.1. Sample Size	6
1.3.2. Patient's Follow-up	6
1.3.3. Study Device use	6
1.3.4. Study Assessments.....	7
2. Statistical methods	7
2.1. General Statistical Considerations	7
2.1.1. Time Points Definition	7
2.1.2. Handling Missing Data.....	7
2.1.3. Descriptive Statistics in Summary Tables	8
2.1.4. Inferential Analysis.....	8
2.1.5. Data Listings.....	8
2.2. Sample size calculation.....	8
2.3. Analysis Sets.....	8
2.3.1. Definition of patient populations.....	8
2.3.2. Protocol Deviations	9
2.3.1. Databases.....	9
2.4. Statistical Analyses.....	10
2.4.1. Patient Disposition and Follow-up	10
2.4.2. Baseline Patient Characteristics.....	10
2.4.3. Index Procedure.....	10
2.4.4. Follow-up at remapping procedure.....	11
2.4.5. 6 and 12 Months follow up.....	11
2.4.6. Feasibility Analysis	11
2.4.7. Additional endpoints.....	12
2.4.8. Safety Endpoint.....	12
2.4.9. Concomitant Medications.....	13
2.5. Derived Criteria Calculation.....	13
3. Statistical Software	25
4. Statistical tables, Listings and Graphs (Table of contents).....	25
4.1. Statistical Tables.....	25
4.1.1. Table 1. Subject enrollment per Site	25
4.1.2. Table 2. Demographics.....	25
4.1.3. Table 3. Baseline Medical History.....	26
4.1.4. Table 4. Procedure.....	29
4.1.5. Table 5. Remap at 60 Days.....	30
4.1.6. Table 6. Per Remapped Veins Analysis	31
4.1.7. Table 7. Arrhythmia Recurrence at 12 Months.....	31
4.1.8. Table 8. Primary Safety Endpoint –Composite Safety Endpoint (CSE)	32
4.1.9. Table 9. Primary Feasibility Endpoint – Acute Procedural Success (ITT population)	33
4.1.10. Table 10. Primary Feasibility Endpoint – Acute Procedural Success (PP population)	33
4.1.11. Table 11. Additional Endpoints.....	33
4.1.12. Table 12. Secondary Safety Endpoints.....	34

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

4.1.13. Table 13. Secondary Feasibility Endpoints	35
4.1.14. Table 14. Adverse Events (AEs) (not including SAEs)	36
4.1.15. Table 15. Serious Adverse Events (SAEs).....	37
4.1.16. Table 16. All Deaths.....	37
4.1.17. Table 17. Protocol Deviations according investigator without COVID-19	37
4.1.18. Table 18. Protocol Deviations according investigator with COVID-19	38
4.1.19. Table 19. Device Deficiency Summary Table.....	38
4.2. Listings	39
4.2.1. Listing 1: Patients populations, reasons of non-inclusion in the different population, withdrawals and follow-up duration.....	39
4.2.2. Listing 2: Protocol deviations according the investigator	39
4.2.3. Listing 3: Demographics at baseline	39
4.2.4. Listing 4: AF History.....	39
4.2.5. Listing 5: Medical History (1/3).....	39
4.2.6. Listing 6: Medical History (2/3).....	39
4.2.7. Listing 7: Medical History (3/3).....	39
4.2.8. Listing 8: Procedure data (1/4).....	39
4.2.9. Listing 9: Procedure data (2/4).....	39
4.2.10. Listing 10: Procedure data (3/4).....	39
4.2.11. Listing 11: Procedure data (4/4).....	39
4.2.12. Listing 12: Additional Ablation.....	39
4.2.13. Listing 13: Chronic Procedural Success at remap.....	39
4.2.14. Listing 14: Primary Safety endpoints	39
4.2.15. Listing 15: Device Deficiencies.....	39
4.2.16. Listing 16: Adverse events according CEC review.....	39
4.2.17. Listing 17: Serious adverse events according CEC review.....	39
4.2.18. Listing 18: Concomitant Medications	39
4.3. Graphs.....	40
4.3.1. Graph 1: Patient disposition.....	40
4.3.2. Graph 2: Freedom from AF/AFL/AT on ADDs.....	40
4.3.3. Graph 3: Freedom from AF/AFL/AT off ADDs.....	41
4.3.4. Graph 4: Freedom from AF on AADs.....	41
4.3.5. Graph 5: Freedom from AF off AADs	41
4.3.6. Graph 6: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days.....	41
4.3.7. Graph 7: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days on AADS	41
4.3.8. Graph 8: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days off AADS	41
4.3.9. Graph 9: Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days on AADs.....	41
4.3.10. Graph 10: Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days off AADs.....	41
4.3.11. Graph 11: Death	41



STATISTICAL ANALYSIS PLAN

FARAPULSE_PersAFOne II SAP
V1.0

PERSAFONE II
V1.0

Page 6 /42

LIST OF ABBREVIATIONS AND DEFINITION OF TERMS

AAD	Antiarrhythmic drug
ABL	Ablation
AE	Adverse event
AF	Atrial fibrillation
AFL	Atrial Flutter
BMI	Body mass index
CPS	Chronic Procedural Success
CT	Computed tomography
CSE	Composite Safety Endpoint
CTI	Cavotricuspid isthmus
DCCV	Direct current cardioversion
ITT	Intention to treat
LA	Left atrium or left atrial
MRI	Magnetic resonance imaging
SAP	Statistical analysis plan
PEF	Pulsed electric field
PersAF	Persistent atrial fibrillation
PP	Per protocol
PT	Preferred term
PV	Pulmonary vein
SAE	Serious adverse event
SOC	System Organ system
TIA	Transient ischemic attack

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 7 /42
---	----------------------------------	---

1. OVERVIEW

This statistical analysis plan (SAP) describes the planned statistical analyses of the data collected during the clinical study PersAFOne II.

This SAP provides additional details concerning the statistical analyses outlined in the protocol (CS1335_Rev A to Rev B PersAFOne II Clinical Protocol_19May2022).

1.1. Study Objective

The objective of this safety and feasibility study is to assess whether the endocardial creation of electrically nonconductive lesions via PEF catheter ablation applied using the FARAPULSE Cardiac Ablation System Plus is a feasible and safe treatment for Persistent atrial fibrillation (PersAF) and associated Atrial flutter (AFL)

1.2. Study Design

This is a prospective, multi-center, unblinded single arm safety and feasibility study. Subjects will undergo percutaneous PEF ablation for pulmonary vein isolation and at the clinical discretion of the investigator receive PEF (Pulsed electric field) ablation of additional arrhythmogenic locations. Subjects will be followed at 30 days, 60 days, 6 months and 12 months for adverse events, recurrence of arrhythmia after a 90-day Blanking Period and other relevant outcome measures.

1.3. Study Plan

1.3.1. Sample Size

Up to 60 subjects in total will be enrolled at up to four sites in this clinical safety and feasibility study. Approved sites will enroll at least 5 and not more than 30 subjects.

1.3.2. Patient's Follow-up

7 visits are scheduled:

- Visit 1: Baseline
- Visit 2: Index Procedure
- Visit 3: Pre-Discharge
- Visit 4: 30-Day Visit
- Visit 5: 60-Day Visit and Remapping Procedure
- Visit 6: 6-Month Visit
- Visit 7: 12-Month Visit

1.3.3. Study Device use

The FARAPULSE Cardiac Ablation System Plus is indicated for the treatment of drug refractory, recurrent, symptomatic persistent atrial fibrillation.

The FARAPULSE Pulsed Field Ablation Catheter is used as an adjunctive device in the endocardial treatment of persistent atrial fibrillation with the following intended uses:

- Gap ablation to complete electrical isolation of the pulmonary veins,

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 8 /42
---	----------------------------------	--

- Focal ablation of cardiac arrhythmias, and
- Creation of ablation line between the inferior vena cava and the tricuspid valve.

1.3.4. Study Assessments

The following flowchart applies to the study:

Assessment	Baseline	Procedure	Pre-Discharge	30-Days Post-Procedure (± 7 days)	60-Days (± 15 days)	6-Month (180 ± 30 days)	12-Month (365 ± 30 days)	Unscheduled
Medical History, CHA ₂ DS ₂ -VASc	X							
AFD and Anticoagulant Medications	X		X	X	X	X	X	X
Symptoms of recurrent arrhythmia				X	X	X	X	X
History of cardioversions, ablations, hospital admissions since last visit			X	X	X	X	X	X
Pregnancy test (for females of childbearing potential)	X	X			X	X ⁴	X ⁴	
12-lead ECG	X		X	X	X	X	X	X
24-Hour Continuous ECG Monitor (e.g., Holter)						X	X	
Cardiac CT/MRI for LA and PV dimensions	X				X			
Mediastinal MRI			X ¹					
TEE or other imaging modality (to exclude left atrial thrombus)	X							
Electroanatomical Mapping		X			X			
Event Monitor readiness/compliance					X	X	X	X
NIHSS	X		X					
Neurologic exam			X ²					
NYHA Classification	X		X	X	X	X	X	X
Fluoroscopic Examination of Diaphragm		X			X if remapped, X ³ if not	X ³	X ³	
Adverse Events		X	X	X	X	X	X	X

2. STATISTICAL METHODS

2.1. General Statistical Considerations

2.1.1. Time Points Definition

Baseline data is defined as the last available observation recorded before the first study device exposition for the patient.

Visit(n) data is defined as the last available observation on or before the Visit(n) time point following the first study device exposition for the patient.

2.1.2. Handling Missing Data

In order to provide unbiased and informative findings, no replacement of missing values is planned for any parameters. Analyses is performed with all available data only.

Data collected by the Investigators and reviewed by the Clinical Event Committee (CEC):

When the data are collected by the investigators and reviewed by the CEC, the analysis will be as follows:

- The data reviewed by the CEC will prevail and will be analyzed with no replacement of missing data from the CEC with the data collected by the investigator

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 9 /42
---	----------------------------------	---

2.1.3. Descriptive Statistics in Summary Tables

- *Continuous variables* will be summarized using standard quantitative statistics: number of non-missing observations, mean, standard deviation, median, quartiles and range (minimum and maximum observed values). The number of missing observations will also be specified.
- *Categorical variables* will be summarized using classical frequency statistics: number of non-missing observations and percentages by categories. Percentages will be calculated on the number of non-missing observations and will be displayed using one decimal. The number of missing observations will also be specified

2.1.4. Inferential Analysis

- Confidence intervals:

When applicable, bilateral asymptotic or exact confidence intervals for binomial distributions will be calculated at the 95% level.

For categorical variables, if pertinent, the 95% asymptotic confidence interval will be calculated if theoretical assumptions are verified. If this is not the case, and the corresponding proportion is 0% or 100%, then the Agresti-Coull confidence interval will be calculated instead. In all other cases, the exact confidence interval will be calculated.

- Survival analysis:

Time-to-event variables will be described over time by survival curves as per the Kaplan-Meier method together with the associated estimators, considering the upper window of the corresponding visits. The survival rate will be presented at each time point (30 days, 90 days, 180 days, 365 days) and overall.

2.1.5. Data Listings

Patient data listings will be selected data supportive of summary statistical tables, including derived/calculated data from statistical process. These data listings are performed on selected analysis sets according to the focus of the analysis and are sorted by subject number and visit. Data entry into a free field in the CRF will be described in the listings.

2.2. Sample size calculation

Up to 60 subjects in total will be enrolled at up to four sites in this clinical safety and feasibility study. Approved sites will enroll at least 5 and not more than 30 subjects.

2.3. Analysis Sets

2.3.1. Definition of patient populations

3 populations will be defined:

- The enrolled population will include all subjects who provided their informed consent.
- Safety Subjects/Intent to Treat Subjects (ITT) population will include all enrolled subjects except those who terminate their participation prior to the beginning of the Index Procedure.
- The Per Protocol (PP) population will include Intent-to-treat subjects for whom the Index Procedure is finished without interfering investigational device deficiency or malfunction.

The primary feasibility endpoint will be analyzed on ITT and on PP population.

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 10 /42
---	----------------------------------	--

2.3.2. Protocol Deviations

A description of the deviation from the protocol and justification must be recorded on the Protocol Deviation Form. The deviations are defined as follows:

- Consentement inappropriately obtained
- Procedure not performed according to protocol/IFU
- Inclusion/Exclusion Criteria deviation
- Protocol required test not completed
- Missed Follow-up visit
- Follow-up visit performed outside of window
- Other

2.3.1. Databases

The Remapping at 60 Days, recurrence and Additional endpoints analysis will be based on the MDT Excel file sent by sponsor, which will contain:

- Received ablation at remap"
- Received RF/PFA ablation at remap
- Patient with Early recurrence of atrial fibrillation (AF)
- AF status at remapping procedure
- AF free survival days
- Atrial fibrillation/ Atrial flutter / Atrial tachycardia (AF/AFL/AT) status at remapping procedure
- AF/AFL/AT free survival days
- 6-month holter
- 12-month holter
- Transtelephonic monitor (TTM) weeks expected
- TTM weeks received
- Number TTM transmissions
- Radiofrequency (RF) touchup
- Ablation (ABL) >90 days Blanking period
- Class I/III antiarrhythmics drug (AAD) >90 days
- Amiodarone
- Direct current cardioversion (DCCV) >90 days
- Arrhythmia-related (AF, AFL, AT) hospitalization

The Excel file sent by the Sponsor will have the same structure (in terms of variables names, labels, formats) for each analysis to ensure the continuity of the programming.

Corelab data will be received, but no analysis will be done. All the data will be received in the MDT file from the sponsor

All other analysis will be done on the e-CRF clinical database.

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 11 /42
---	----------------------------------	--

2.4. Statistical Analyses

2.4.1. Patient Disposition and Follow-up

2.4.1.1. Patient Populations Sample Size

The number and percentage of patients included in each population and the reasons of non-inclusion in each subsequent population will be presented on the enrolled population.

The number of patients included, as well as the distribution by center will be presented globally. The number and percentage of patients present at each visit will be presented with the study exit reasons.

A listing will be provided to present patient disposition. The following variables will be listed:

- Date of consent
- Procedure date
- Remap date
- Most recent follow-up
- Date of study exit
- Follow up time (days)

2.4.1.2. Protocol Deviations

Frequency of patients with at least one protocol deviation, as well as the total number of events, will be summarized on the ITT population, by Type of deviation.

The detail will be also presented in a listing.

2.4.2. Baseline Patient Characteristics

Descriptive statistical data will be used to draw up a recapitulation of the characteristics of the patients at the time of enrolment. It will be summarized on the ITT population.

- Demographics at baseline: Age, Gender, height, weight, BMI
- Baseline Medical History: CHA2DS2-VASc Score, heart rate, blood pressure, Months between Date of first documented AF diagnosis and Enrollment date, number of AF episodes past year, Months between Date of most recent AF episode and Enrollment date, failed drug, Hypertension, Cardiac arrhythmias, CT/MRI results, Echocardiography, LVEF, Left Atrial AP Diameter, Hypertension

Previous parameters will also be presented in a listing.

2.4.3. Index Procedure

Descriptive statistical data will be used to draw up a recapitulation of the intervention details at the time of the procedure.

The following data will be presented on the ITT population:

- Procedure Data: Procedure time (hours), Farawave dwell total time (hours), PVI ablation time (hours), FARAPPOINT CTI dwell TOTAL (hours), CTI Experimental ablation catheter time, Total fluoroscopy time (minutes), Total volume contrast (cc), Total Volume IV Fluids (cc); Total number of PVI ablations, Total number of ablations for each vein, Total number of PVI ablations by patient, PVI status at end of procedure by each attempted vein, Num PV treated, Num PV isolated, Proportion of veins isolated, CTI acute isolation, Total number of ablations of CTI applications, LA floor attempted and LA floor Isolated/Blocked, LA roof attempted and LA roof

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 12 /42
---	----------------------------------	---

Isolated/Blocked, LA MI attempted and LA MI Isolated/Blocked, LAPW attempted and LAPW Isolated/Blocked, Total number of ablations for LAPW, Total number of ablations for LAPW with FARAWAVE, LAPW Experimental ablation catheter time,

A listing will be provided to present previous parameters.

2.4.4. Follow-up at remapping procedure

Descriptive statistical data will be used to draw up a recapitulation of the follow-up details at the remap procedure.

The following data will be presented on the ITT population:

Remapping details: Days follow up at remap, Durable PVI, LSPV chronic isolation, LIPV chronic isolation, LCPV chronic isolation, RSPV chronic isolation, RIPV chronic isolation, RMPV chronic isolation, CTI chronic isolation, LAPW chronic isolation, Number of veins durably isolated, Number of veins with gaps.

The details will also be presented in a listing.

2.4.5. 6 and 12 Months follow up

Descriptive statistical data will be used to draw up a recapitulation of the follow-up details at 6 and 12 Months follow up.

The following data will be presented on the ITT population:

TTM compliance and Holter compliance at 6 months and 12 months. Freedom from AF/AFL/AT (for patient On AADs and on patient Off AADs), Freedom from AF (for patient On AADs and on patient Off AADs), Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days (for patient On AADs and on patient Off AADs), Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days (for patient On AADs and on patient Off AADs).

2.4.6. Feasibility Analysis

The feasibility analysis will be performed on the ITT population.

2.4.6.1. Primary Feasibility Analysis

The proportion of subjects that achieve Acute Procedural Success (APS) defined as the percutaneous endocardial creation of a complete, electrically isolating set of lesions around the ostia of the pulmonary veins (PVI) using the FARAPULSE Cardiac Ablation System Plus during the Index Procedure, as clinically assessed by entrance and/or exit block performed ≥ 20 minutes after the last PVI lesion is made.

2.4.6.2. Secondary Feasibility Analysis

The secondary feasibility analyses are:

- The proportion of subjects who undergo the protocol-specified 60-Day Remapping Procedure and achieve Chronic Procedural Success (CPS) defined as persistent electrical isolation of all initially ablated pulmonary veins. Chronic Procedural Success will be subdivided by Index Procedure only CPS and Reablated CPS subjects.
- The proportion of attempted subjects that achieve Acute CTI Success, defined as the creation of bi-directional electrical block across the CTI using the investigational devices.
- Durability of the CTI and/or other extra-PV lesion set(s), when applicable. Durability is defined as unaltered integrity of each such lesion assessed at the Remapping Procedure.
- The proportion of subjects that achieve Therapeutic Success, defined as freedom from:
 - Post-Blanking Period through assessment: occurrence of AF, AFL or AT, or ablation for AF/AFL/AT using the study device
 - At any time: ablation for AF/AFL/AT with a non-study device

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 13 /42
---	----------------------------------	---

Therapeutic Success will be assessed from the end of the 90 Day Blanking Period through Months 6 and 12 and will be subdivided by on / off AFDs (Atrial Fibrillation Drugs) post-Blanking Period.

2.4.7. Additional endpoints

The additional observations will be performed on the ITT population and will include:

- Proportion of subjects with early recurrence of atrial fibrillation (ERAF) by 90 days after the initial study ablation.
- Time to Therapeutic Failure
- Proportion of all ablated pulmonary veins that are isolated at the Index Procedure using the study device.
- Proportion of all ablated pulmonary veins isolated using the study device during the Index Procedure that remain isolated at the 60-day Remapping Procedure.
- The proportion of attempted subjects that achieve Chronic CTI Success, defined as persistent bi-directional electrical block across the CTI as assessed at the 60-day remapping procedure.
- The proportion of attempted subjects that achieve Chronic Focal Success, defined as persistent electrical nonconductivity of extra-PV tissue targeted for ablation in the Index Procedure, excluding the CTI as assessed at the 60 day remapping procedure.

2.4.8. Safety Endpoint

Descriptive statistical data is used to analyze the Primary Safety Endpoint (PSE), the secondary safety endpoints and others analysis. The analysis is performed on the ITT populations.

The Safety endpoint will be performed on the ITT population and will include

2.4.8.1. Primary Safety Analysis

The primary safety endpoint for this study is the Composite Safety Endpoint (CSE) defined as the incidence of the following early-onset and late-onset serious adverse events (SAEs) which are device- or procedure-related, as adjudicated by the CEDMC.

Composite Safety Endpoint Definitions.

- Early onset (within 30 days of an Index or Remap Procedure)
 - Death
 - Myocardial infarction (MI)
 - Persistent diaphragmatic paralysis
 - Stroke or transient ischemic attack (TIA)
 - Peripheral or organ thromboembolism
 - Pericarditis
 - Cardiac tamponade / perforation
 - Vascular access complications requiring intervention
 - Heart block
- Late onset (any time during follow-up through 12 months)
 - Pulmonary vein (PV) stenosis (> 70% diameter reduction from baseline)
 - Atrio-esophageal fistula

2.4.8.2. Secondary Safety Analysis

- The Primary Safety Endpoint assessed at 7 rather than 30 days

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 14 /42
---	----------------------------------	---

- The proportion of subjects with a device- or procedure-related SAE
- The proportion of subjects with stroke or TIA
- The proportion of subjects requiring cardioversion
- The proportion of subjects requiring an arrhythmia-related hospitalization

2.4.9. Concomitant Medications

Concomitant medications will be listed on the ITT population.

2.5. Derived Criteria Calculation

If at least one of the items needed to calculate a derived criteria is missing, then the corresponding derived criteria will be considered as missing.

Patient Disposition:

- Derived criteria “Enrolled population” will be derived as follows:
IF the date of patient consent is filled in **Patient eligibility form (IE)** **THEN** “Enrolled population” = ‘Yes’
ELSE “Enrolled population” = ‘No’.
- Derived criteria “ITT population” will be defined as follows:
IF “Enrolled population” = ‘Yes’ and the date of the Procedure is filled in **Procedural data form (PR)** **THEN** “ITT population” = ‘Yes’
ELSE “ITT population” = ‘No’.
- Derived criteria “Per Protocol population PP” will be defined as follows:
IF “ITT population” = ‘Yes’ **AND** Device Deficiency Number is not filled **AND** Serial # is not filled in **Device Deficiency DD Form** **THEN** Per Protocol population = ‘Yes’
ELSE Per Protocol population = ‘No’.
- Derived criteria “Time of patient follow-up” will be calculated as:
Date of study exit/withdrawal (*) in **Patient eligibility form (IE)** – date of procedure in **Procedural data form (PR)**
*if the patient is still in the study, the last visit date will be used instead of the exit/ withdrawal one.
- Derived criteria “Study exit by visit” will be calculated as:
 - **IF** STUDY EXIT DATE <PROCEDURE DATE **THEN** patient exited study Before Procedure.
 - **ELSE IF** STUDY EXIT DATE≥PROCEDURE DATE **AND** (DISCHARGE DATE is missing **AND** 30 DAYS VISIT DATE is missing **AND** REMAP DATE is missing **AND** 6 MONTHS VISIT DATE is missing **AND** 12 MONTHS VISIT DATE is missing) **THEN** patient exited study after Procedure.
 - **ELSE IF** STUDY EXIT DATE≥ DISCHARGE DATE **AND** (30 DAYS VISIT DATE is missing **AND** REMAP DATE is missing **AND** 6 MONTHS VISIT DATE is missing **AND** 12 MONTHS VISIT DATE is missing) **THEN** patient exited study after Discharge.
 - **ELSE IF** STUDY EXIT DATE≥ 30 DAYS VISIT DATE **AND** (REMAP DATE is missing **AND** 6 MONTHS VISIT DATE is missing **AND** 12 MONTHS VISIT DATE is missing) **THEN** patient exited study after 30 days visit.

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 15 /42
---	----------------------------------	---

- **ELSE IF STUDY EXIT DATE \geq REMAP DATE AND** (6 MONTHS VISIT DATE is missing **AND** 12 MONTHS VISIT DATE is missing) **THEN** patient exited study after Remap procedure.
- **ELSE IF STUDY EXIT DATE \geq 6 MONTHS VISIT DATE AND** (12 MONTHS VISIT DATE is missing) **THEN** patient exited study after 6 months.
- **ELSE IF STUDY EXIT DATE \geq 12 MONTHS VISIT DATE THEN** patient exited study after 12 months.

Baseline Patient Characteristics:

- Derived criteria “Time between the first documented AF diagnosis and the enrollment (months)” will be defined as follows:
(Informed consent date – date of first documented AF diagnosis) /30.4167 in **AF history form (AH)**.
- Derived criteria “Time between the most recent AF episode and the enrollment (months)” will be defined as follows:
(Informed consent date – date of most recent AF episode) /30.4167 in **AF history form (AH)**.

Index Procedure:

- Derived criteria “Procedure time (hours)” will be defined as follows:
Procedure end- Procedure start in **Procedural data form (PR)**.
- Derived criteria “PVI ablation time (hours)” will be defined as follows:
Last PVI ablation* time - First PVI ablation* time in **Procedural data form (AB)**
- *Where ablation location ablation location : ablation location = (“Left superior”, “Left inferior”, “Right superior”, “Right inferior”, “Left common ostium”, “Right middle”) and (Ablation location=’Other” and Ablation location, Other specify=’LCPV’)) .
- Derived criteria “FARAWAVE dwell time (hours)” will be defined as follows:
SUM of [(FARAWAVE ablation catheter remove 1 – FARAWAVE ablation catheter introduction 1), (FARAWAVE ablation catheter remove n – FARAWAVE ablation catheter introduction n)] in **Procedural data form (PR)**
* FARAWAVE ablation catheter remove/introduction = Ablation catheter remove/introduction
WHERE FARAWAVE: Type of endocardial catheter used for ablation is Ticked **OR** Type of endocardial catheter used for ablation =’FARAWAVE’
- Derived criteria “CTI (FARAPoint) dwell time (hours)” will be defined as follows:
SUM of [(FARAPoint ablation catheter remove 1 – FARAPoint ablation catheter introduction 1),
(FARAPoint ablation catheter remove n – FARAPoint ablation catheter introduction n)] in **Procedural data form (PR)**
* FARAPoint ablation catheter remove/introduction = Ablation catheter remove/introduction
WHERE FARAPoint: Type of endocardial catheter used for ablation is Ticked **OR** Type of endocardial catheter used for ablation = FARAPoint
- Derived criteria “LAPW ablation time (hours)” will be defined as follows:
Last LAPW ablation time - First LAPW ablation time in **Procedural data form (AB)**

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	<h1>STATISTICAL ANALYSIS PLAN</h1>	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 16 /42
---	------------------------------------	--

- Derived criteria “Total number of PVI applications” will be defined as follows:
 $\text{SUM of } [(\text{Count (Ablation location IN ("Left superior", "Left inferior", "Right superior", "Right inferior", "Right middle", "Left common ostium") OR (Ablation location="Other" AND Ablation location, Other specify='LCPV'))}]$ in **Procedural data form (AB)**
- Derived criteria “Number of LAPW applications” will be defined as follows:
 $\text{SUM of } [(\text{Count (Ablation location='Other" AND Ablation location, Other specify IN ("LA PW", "LAPW", "LEFT ATRIAL POSTERIOR WALL"))}]$ in **Procedural data form (AB)**
- Derived criteria “Total number of CTI applications” will be defined as follows:
 $\text{SUM of } [(\text{Count (Ablation location=" CTI"))}]$ in **Procedural data form (AB)**
- Derived criteria “Total number of veins treated” will be defined as follows:
 $\text{COUNT of } [\text{left superior", "left inferior", "left common ostium", "right middle", "right superior", "right inferior"}]$ **WHERE** isolated/Blocked in ('Yes', 'No') in **Procedural data form (PR)**
- Derived criteria “Total number of veins isolated” will be defined as follows:
 $\text{SUM of } [\text{left superior, left inferior, left common ostium, right middle, right superior, right inferior}]$ **WHERE** isolated/Blocked = 'Yes' in **Procedural data form (PR)**
- Derived criteria “Of the total number of veins treated, the number of veins isolated (Proportion of veins isolated)” will be defined as follows:
 $\text{Derived criteria "Total number of veins isolated"/ Derived criteria "Total number of veins treated"}$
- Derived criteria Proportion “CTI acute isolation” will be defined as follows:
 $\text{COUNT of } [\text{CTI isolated/Blocked='Yes' /CTI Attempted='Yes'}]$ in **Procedural data form (PR)**

Remapping at 60 Days

- Derived criteria “Days of follow-up at remapping procedure” will be defined as follows:
Assessment date (**WHERE** Visit='60-Day Follow-up') in **Follow-up form (FU)** – Procedure date in **Procedural data form (AB)**
- Derived criteria “Durable PVI” will be defined as follows:
IF Any PV reconnection noted? = 'No' in **Follow-up 3D electroanatomical mapping form (RM)**
THEN PVI durable = 'Yes'
ELSE IF Any PV reconnection noted? = 'Yes' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** PVI Durable=No
- Derived criteria “LAPW chronic isolation” will be defined as follows:
IF Anatomical Structure/Area other specification 1 **IN** ("POSTERIOR WALL" "LA POSTERIOR WALL" "LAPW" "PA WALL" "BOX POSTERIOR WALL") **AND** Other 1 Lesion durable = 'yes' **OR** Anatomical Structure/Area other specification 2 **IN** ("POSTERIOR WALL" "LA POSTERIOR WALL" "LAPW" "PA WALL" "BOX POSTERIOR WALL") **AND** Other 2 Lesion durable = 'Yes' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** LAPW chronic isolation= 'Yes'
ELSE LAPW chronic isolation = 'No'
- Derived criteria “Num veins with isolation” will be defined as follows:

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 17 /42
---	----------------------------------	--

SUM OF [(Anatomical structure/Area Isolated = ('Left superior, left inferior, Left common ostium, right superior, right inferior, right middle) **AND** Isolated (Yes/No/'N/A') ='Yes'] in **Follow-up 3D electro anatomical mapping form (RM)**

- Derived criteria “Num veins with gaps” will be defined as follows:

SUM OF [(Anatomical structure/Area Isolated = ('Left superior, left inferior, Left common ostium, right superior, right inferior) **AND** Isolated (Yes/No/'N/A') ='No'] in **Follow-up 3D electro anatomical mapping form (RM)**

- Derived criteria “Percentage of PVs Isolated per-vein at remap” will be defined as follows:
(Num veins with isolation /(Num veins with isolation + Num veins with gaps))*100

- Derived criteria “Num veins remain isolated” will be defined as follows:
Derived criteria “Num PV isolated” – Derived criteria “Num veins with gaps”

- Derived criteria “Proportion PVs Remaining isolated” will be defined as follows:
Derived criteria “Num veins remain isolated” /Derived criteria “Num PV isolated”

6 and 12 Months follow up

- Derived criteria “TTM compliance (%) (calculated)” will be defined as follows:
(TTM weeks received /TTM weeks expected) *100 in **MDT excel file**
- Derived criteria “Holter compliance (%) at 6 months” will be defined as follows:
IF 6mo Holter in ('AF', 'none', 'SVT') in **MDT excel file** **THEN** Holter compliance (%) at 6 months = 'Yes'
ELSE IF 6mo Holter in ('MISSING') in **MDT excel file** **THEN** Holter compliance (%) at 6 months='No'
- Derived criteria “Holter compliance (%) at 12 months” will be defined as follows:
IF 12mo Holter in ('AF', 'none', 'SVT') in **MDT excel file** **THEN** Holter compliance (%) at 12 months = 'Yes'
ELSE IF 12mo Holter in ('MISSING') in **MDT excel file** **THEN** Holter compliance (%) at 12 months='No'
- Derived criteria “Freedom from AF/AFL/AT” will be defined as follows:
IF AF/AFL/AT status =0 in **MDT excel file** **THEN** Freedom from AF/AFL/AT ='Yes'
ELSE Freedom from AF/AFL/AT ='No'
- Derived criteria “Freedom from AF/AFL/AT with Class I/III AAD >90 days” will be defined as follows:
IF AF/AFL/AT status =0 **AND** Class I/III AAD >90 days =1 in **MDT excel file** **THEN** Freedom from AF/AFL/AT with Class I/III AAD >90 days ='Yes'
ELSE Freedom from AF/AFL/AT with Class I/III AAD >90 days ='No'
- Derived criteria “Freedom from AF/AFL/AT without Class I/III AAD >90 days” will be defined as follows:

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	<h1>STATISTICAL ANALYSIS PLAN</h1>	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 18 /42
---	------------------------------------	--

IF AF/AFL/AT status =0 **AND** Class I/III AAD >90 days =0 in **MDT excel file THEN** Freedom from AF/AFL/AT without Class I/III AAD >90 days = 'Yes'

ELSE Freedom from AF/AFL/AT without Class I/III AAD >90 days ='No'

- Derived criteria "Freedom from AF" will be defined as follows:

IF AF status =0 in **MDT excel file THEN** Freedom from AF ='Yes'

ELSE Freedom from AF ='No'

- Derived criteria "Freedom from AF with Class I/III AAD >90 days" will be defined as follows:

IF AF status =0 **AND** Class I/III AAD >90 days =1 in **MDT excel file THEN** Freedom from AF with Class I/III AAD >90 days ='Yes'

ELSE Freedom from AF with Class I/III AAD >90 days ='No'

- Derived criteria "Freedom from AF without Class I/III AAD >90 days" will be defined as follows:

IF AF status =0 **AND** Class I/III AAD >90 days =0 in **MDT excel file THEN** Freedom from AF without Class I/III AAD >90 days ='Yes'

ELSE Freedom from AF without Class I/III AAD >90 days ='No'

- Derived criteria "Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days" will be defined as follows:

IF AF/AFL/AT status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 in **MDT excel file THEN** Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days ='Yes'

ELSE Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days ='No'

- Derived criteria "Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days" will be defined as follows:

IF AF/AFL/AT status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 **AND** Class I/III AAD >90 days =1 in **MDT excel file THEN** Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days='Yes'

ELSE Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days='No'

- Derived criteria "Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days" will be defined as follows:

IF AF/AFL/AT status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 **AND** Class I/III AAD >90 days =0 in **MDT excel file THEN** Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days= 'Yes'

ELSE Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days='No'

- Derived criteria "Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days" will be defined as follows:

IF AF status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 in **MDT excel file THEN** Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days ='Yes'

ELSE Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days ='No'

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 19 /42
---	----------------------------------	--

- Derived criteria “Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days” will be defined as follows:
IF AF status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 **AND** Class I/III AAD >90 days =1 in **MDT excel file** **THEN** Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days ='Yes'
ELSE Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days with Class I/III AAD >90 days ='No'
- Derived criteria “Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days” will be defined as follows:
IF AF status =0 **AND** DCCV90D=0 **AND** RF touch up=0 **AND** ABL>90 day=0 **AND** Class I/III AAD >90 days =0 in **MDT excel file** **THEN** Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days='Yes'
ELSE Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days without Class I/III AAD >90 days= 'No'

Primary Feasibility endpoint

- Derived criteria “Acute Procedural Success (APS)” will be defined as follows
IF Ablation with study device was successful = ‘Yes’ in **procedural form (PR)** **THEN** APS='Yes'
ELSE IF Ablation with study device was successful = ‘No’ in **procedural form (PR)** **THEN** APS='No'

Secondary Feasibility endpoint

- Derived criteria “Chronic Procedural Success (CPS)” will be defined as follows:
IF Has all ablation with the study device(s) been successful? ='Yes' in **Procedural data form (PR)** **AND** Has all ablation with the study device remain successful? ='Yes' **AND** Additional ablation(s) performed? ='No' in **FOLLOW UP 3D ELECTRO ANATOMICAL MAPPING form (RM)** **THEN** Chronic Procedural Success (CPS) ='Yes, index procedure only'
ELSE IF Has all ablation with the study device(s) been successful? ='Yes' in **Procedural data form (PR)** **AND** Has all ablation with the study device remain successful? ='No' **AND** Any PV reconnections noted? in = ‘Yes’ **AND** Additional ablation(s) performed? ='Yes' in **FOLLOW UP 3D ELECTRO ANATOMICAL MAPPING form (RM)** **THEN** Chronic Procedural Success (CPS) ='Yes, reblated'
ELSE IF Has all ablation with the study device(s) been successful? ='No' in **Procedural data form (PR)** **THEN** Chronic Procedural Success (CPS) ='No'

- Derived criteria “Acute CTI Success” will be defined as follows:
COUNT of [CTI isolated/Blocked='Yes'/ CTI Attempted='Yes'] in **Procedural data form (PR)**

- Derived criteria “Durability of CTI at remap,” will be defined as follows:

IF CTI lesion durable ='Yes' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of CTI at remap = 'Yes'
ELSE IF CTI lesion durable ='No' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of CTI at remap = 'No'

- Derived criteria “Durability of Mitral isthmus at remap” will be defined as follows:

IF Mitral isthmus lesion durable ='Yes' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Mitral isthmus at remap = 'Yes'
ELSE IF Mitral isthmus lesion durable ='No' in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Mitral isthmus at remap = 'No'

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 20 /42
---	----------------------------------	--

- Derived criteria “Durability of Left atrial roof line at remap” will be defined as follows:
IF Left atrial roof line lesion durable =’Yes’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Left atrial roof line at remap = ’Yes’
ELSE IF Left atrial roof line lesion durable =’No’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Left atrial roof line at remap = ’No’
- Derived criteria “Durability of Left atrial Posterior line” will be defined as follows:
IF Left atrial Posterior line lesion durable =’Yes’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Left atrial Posterior line at remap = ’Yes’
ELSE IF Left atrial Posterior line lesion durable =’No’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** Durability of Left atrial Posterior line at remap = ’No’
 - Derived criteria “Durability of Left atrial Posterior wall” will be defined as follows:
IF Anatomical Structure/Area other specification 1 **IN** (“POSTERIOR WALL” “LA POSTERIOR WALL” “LAPW” “PA WALL” “BOX POSTERIOR WALL”) **AND** Other 1 Lesion durable = ’yes’ **OR** Anatomical Structure/Area other specification 2 **IN** (“POSTERIOR WALL” “LA POSTERIOR WALL” “LAPW” “PA WALL” “BOX POSTERIOR WALL”) **AND** Other 2 Lesion durable =’Yes’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN** LAPW chronic isolation= ’Yes’
- ELSE LAPW chronic isolation = ’No’
- Derived criteria “Therapeutic Success ON/OFF AAD” will be defined as follows:
IF “AF/AFL/AT status”= 0 **AND** “RF touchup”= 0 **AND** “ABL >90 days” = 0 in **MDT excel file** **THEN** therapeutic success=’Yes’
ELSE therapeutic success=’No’
- Derived criteria “Therapeutic Success Off ADD” will be defined as follows:
IF Therapeutic success =’Yes’ **AND** Class I/III AAD>90days=0 in **MDT excel file** **THEN** therapeutic success Off ADD =’YES’
ELSE IF Therapeutic success =’No’ **AND** Class I/III AAD>90days=0 in **MDT excel file** **THEN** therapeutic success Off ADD =’NO’
- Derived criteria “Therapeutic Success On ADD” will be defined as follows:
IF Therapeutic success =’Yes’ **AND** Class I/III AAD>90days=1 in **MDT excel file** **THEN** therapeutic success On ADD =’No’
Else IF Therapeutic success =’No’ **AND** Class I/III AAD>90days=1 in **MDT excel file** **THEN** therapeutic success On ADD =’No’

Additional endpoints

- Derived criteria “Proportion of all ablated pulmonary veins that are isolated at the Index Procedure using the study device” will be defined as follows:
COUNT of [(Left Superior Isolated/Blocked =’Yes’, Left Inferior Isolated/Blocked=’Yes’, Left Common Ostium Isolated/Blocked=’Yes’, Right Middle Isolated/Blocked=’Yes’, Right Superior Isolated/Blocked=’Yes’, Right Inferior Isolated/Blocked=’Yes’)] / COUNT of [(Left Superior Isolated/Blocked in (Yes, No) , Left Inferior Isolated/Blocked in (Yes, No), Left Common Ostium Isolated/Blocked in (Yes, No), Right Middle Isolated/Blocked in (Yes, No), Right Superior

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 21 /42
---	----------------------------------	--

Isolated/Blocked in (Yes, No), Right Inferior Isolated/Blocked in (Yes, No))] in **Procedural data form (PR)**

- Derived criteria “Proportion of all ablated pulmonary veins isolated using the study device during the Index Procedure that remain isolated at the 60-day remapping procedure” will be defined as follows:
(Derived criteria “Total number of veins isolated” - Derived criteria “Number of veins with gaps”)/
Derived criteria “Total number of veins isolated”
- Derived criteria Proportion “Chronic CTI at 60d remap” will be defined as follows:
IF CTI lesion durable =’Yes’ in **Follow-up 3D electroanatomical mapping form (RM)** **THEN**
Chronic CTI at 60d remap =’Yes’
ELSE IF CTI lesion durable =’No’ in **Follow-up 3D electroanatomical mapping form (RM)**
THEN Chronic CTI at 60d remap =’No’
- Derived criteria “Chronic Focal Success” will be defined as follows:
IF Anatomical Structure/Area other specification 1 **IN** (“POSTERIOR WALL” “LA POSTERIOR WALL” “LAPW” “PA WALL” “BOX POSTERIOR WALL”) **AND** Other 1 Lesion durable = ‘yes’ **OR**
Anatomical Structure/Area other specification 2 **IN** (“POSTERIOR WALL” “LA POSTERIOR WALL” “LAPW” “PA WALL” “BOX POSTERIOR WALL”) **AND** Other 2 Lesion durable = =’Yes’ in
Follow-up 3D electroanatomical mapping form (RM) **THEN** Chronic focal succes= ‘Yes’
ELSE Chronic focal succes = ‘No’

Primary Safety endpoint

Unless otherwise indicated, all the following variables will be derived from the Adverse event adjudication form (CE)

- Derived criteria “Death” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-≤REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) AND PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Death’ **THEN**
DEATH=’Yes’
ELSE DEATH=’No’
- Derived criteria “Myocardial infarction” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Myocardial infarction’
THEN Myocardial infarction=’Yes’.
ELSE Myocardial infarction =’No’.
- Derived criteria “PV Stenosis” will be defined as follows:
IF (0≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) ≤360 **AND**
PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ PV Stenosis’ **THEN** PV Stenosis =’Yes’.
ELSE PV Stenosis =’No’.

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 22 /42
---	----------------------------------	--

- Derived criteria “Diaphragmatic Paralysis” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Diaphragmatic Paralysis’ **THEN** Diaphragmatic Paralysis =’Yes’.
ELSE Diaphragmatic Paralysis =’No’.
- Derived criteria “AE fistula” will be defined as follows:
IF (0≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) ≤360 **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ AE fistula’ **THEN** AE fistula =’Yes’.
ELSE AE fistula =’No’.
- Derived criteria “TIA” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ TIA’ **THEN** TIA=’Yes’.
ELSE TIA=’No’.
- Derived criteria “CVA” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ CVA’ **THEN** CVA=’Yes’.
ELSE CVA=’No’.
- Derived criteria “Major pericarditis” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Major pericarditis’ **THEN** Major pericarditis =’Yes’.
ELSE Major pericarditis =’No’.
- Derived criteria “Tamponade/Perforation” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Tamponade/Perforation’ **THEN** Tamponade/Perforation =’Yes’
ELSE Tamponade/Perforation =’No’.
- Derived criteria “Heart Block” will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=’ Heart Block’ **THEN** Heart Block =’Yes’
ELSE Heart Block =’No’

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

- Derived criteria "Vascular Access Complication" will be defined as follows:
IF (30≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (30≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=' Vascular Access Complication' **THEN** Vascular Access Complication ='Yes'
ELSE Vascular Access Complication ='No'
- Derived criteria "Total Early and Late Onset Composite Safety Endpoint" will be defined as follows:
IF Death=Yes **OR** Myocardial infarction =Yes **OR** PV Stenosis =Yes **OR** Diaphragmatic Paralysis =Yes **OR** AE fistula =Yes **OR** TIA =Yes **OR** CVA=Yes **OR** Major pericarditis =Yes **OR** Tamponade/Perforation =Yes **OR** Heart Block =Yes **OR** Vascular Access Complication =Yes **THEN** Total Early and Late Onset Composite Safety Endpoint ='Yes'
ELSE Total Early and Late Onset Composite Safety Endpoint ='No'
- Derived criteria "Total Early Onset Composite Safety Endpoint" will be defined as follows:
IF Death=Yes **OR** Myocardial infarction =Yes **OR** Diaphragmatic Paralysis =Yes **OR** TIA =Yes **OR** CVA=Yes **OR** Major pericarditis =Yes **OR** Tamponade/Perforation =Yes **OR** Heart Block =Yes **OR** Vascular Access Complication =Yes **THEN** Total Early Onset Composite Safety Endpoint='Yes'
ELSE Total Early Onset Composite Safety Endpoint='No'
- Derived criteria "Total Late (any time during follow-up through 12 months)" will be defined as follows:
IF PV Stenosis =Yes **OR** AE fistula =Yes **THEN** Total Late (any time during follow-up through 12 months) ='Yes'
ELSE Total Late (any time during follow-up through 12 months) ='No.

Secondary Safety endpoint

Unless otherwise indicated, all the following variables will be derived from the Adverse event adjudication form (CE)

- Derived criteria "Death 7 days" will be defined as follows:
IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=' Death' **THEN** DEATH='Yes'
ELSE DEATH='No'
- Derived criteria "Myocardial infarction 7 days" will be defined as follows:
IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA=' Myocardial infarction' **THEN** Myocardial infarction='Yes'.
ELSE Myocardial infarction = 'No'.
- Derived criteria "Diaphragmatic Paralysis 7 days" will be defined as follows:
IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING**

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 24 /42
---	----------------------------------	--

FORM (RM)) AND PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' Diaphragmatic Paralysis' **THEN** Diaphragmatic Paralysis = 'Yes'.
ELSE Diaphragmatic Paralysis = 'No'.

- Derived criteria "TIA 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' TIA' **THEN** TIA='Yes'.
ELSE TIA='No'.

- Derived criteria "CVA 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' CVA' **THEN** CVA='Yes'.
ELSE CVA='No'.

- Derived criteria "Major pericarditis 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' Major pericarditis' **THEN** Major pericarditis = 'Yes'.
ELSE Major pericarditis = 'No'.

- Derived criteria "Tamponade/Perforation 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' Tamponade/Perforation' **THEN** Tamponade/Perforation = 'Yes'.
ELSE Tamponade/Perforation = 'No'.

- Derived criteria "Heart Block 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' Heart Block' **THEN** Heart Block = 'Yes'.
ELSE Heart Block = 'No'.

- Derived criteria "Vascular Access Complication 7 days" will be defined as follows:

IF (7≤ (EVENT DATE-PROCEDURE DATE IN **PROCEDURAL DATA FORM (PR)**) **OR** (7≤ (EVENT DATE-REMAPPING DATE IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)**) **AND** PRIMARY SAFETY ENDPOINT EVENT CRITERIA= ' Vascular Access Complication' **THEN** Vascular Access Complication = 'Yes'.
ELSE Vascular Access Complication = 'No'.

- Derived criteria "Total Early Onset Composite Safety Endpoint assessed at 7 days" will be defined as follows:

IF Death 7 days='Yes' **OR** Myocardial infarction 7 days='Yes' **OR** Diaphragmatic Paralysis 7 days='Yes' **OR** TIA 7 days='Yes' **OR** CVA 7 days='Yes' **OR** Major pericarditis 7 days='Yes' **OR**

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

Tamponade/Perforation 7 days='Yes' **OR** Heart Block 7 days='Yes' **OR** Vascular Access Complication 7 days='Yes' **THEN** Total Early Onset Composite Safety Endpoint assessed at 7 days = 'Yes'.

ELSE Total Early Onset Composite Safety Endpoint assessed at 7 days = 'No'.

- Derived criteria "Device-Related Serious Adverse Events (SAEs)" will be defined as follows:
IF CEC seriousness = "Serious" **AND** CEC relationship to study Device in (Causal relationship, Probably related, Possibly related, Unlikely to be related) in adverse event adjudication form **THEN** Device-Related Serious Adverse Events (SAEs)='Yes'; **ELSE** Device-Related Serious Adverse Events (SAEs)='No'
- Derived criteria "Procedure-Related Serious Adverse Events (SAEs)" will be defined as follows:
IF CEC relationship to study procedure in (Causal relationship, Probably related, Possibly related, Unlikely to be related) **AND** CEC seriousness = 'Serious' **THEN** Procedure-Related Serious Adverse Events (SAEs)='Yes'
ELSE IF CEC relationship to study procedure ='Not related' **AND** CEC seriousness = " Serious" **THEN** Procedure-Related Serious Adverse Events (SAEs)='No'
- Derived criteria "Cardioversion" will be defined as follows:
IF Arrhythmia-Treatment DC Cardioversion is ticked in **Arrhythmia episodes form** **THEN** Cardioversion ='Yes';
ELSE Cardioversion ='No'
- Derived criteria "Stroke after index procedure" will be defined as follows:
IF Event name='STROKE' in **Adverse event form** **AND** Event start date ≥ Procedure date in **Procedural data Form (PR)** **THEN** STROKE='Yes'
ELSE STROKE='No'
- Derived criteria "Transient Ischemic Attack (TIA) after index procedure" will be defined as follows:
IF Event name='Transient Ischemic Attack' in **Adverse event form** **AND** Event start date ≥ Procedure date in **Procedural data Form (PR)** **THEN** TIA='Yes'
ELSE TIA='No'
- Derived criteria "Stroke after Remap procedure" will be defined as follows:
IF Event name='STROKE' in **Adverse event form** **AND** Event start date ≥ Remap date IN **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)** **THEN** STROKE='Yes'
ELSE STROKE='No'
- Derived criteria "Transient Ischemic Attack (TIA) after Remap procedure" will be defined as follows:
IF Event name='Transient Ischemic Attack' in **Adverse event form** **AND** Event start date ≥ Remap date in **FOLLOW-UP 3D ELECTROANATOMICAL MAPPING FORM (RM)** **THEN** TIA='Yes'
ELSE TIA='No'
- Derived criteria "Arrhythmia-related (AF, AFL, AT) hospitalization" will be defined as follows:
IF Arrhythmia-related (AF, AFL, AT) hospitalization ='Yes' in **MDT excel file** **THEN** Arrhythmia-related (AF, AFL, AT) hospitalization ='Yes'
ELSE IF Arrhythmia-related (AF, AFL, AT) hospitalization ='No' in **MDT excel file** **THEN** Arrhythmia-related (AF, AFL, AT) hospitalization ='No'

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 26 /42
---	----------------------------------	---

- Derived criteria "Death" will be defined as follows:

IF SAE: Death='Yes' in **Adverse event form** **OR** Reason ='Patient death' in **study exit form**
THEN Death='Yes'
ELSE Death='No'

3. STATISTICAL SOFTWARE

All statistical outputs (summary tables and data listings) will be generated using SAS® version 9.4 or further.

4. STATISTICAL TABLES, LISTINGS AND GRAPHS (TABLE OF CONTENTS)

4.1. Statistical Tables

4.1.1. Table 1. Subject enrollment per Site

		Enrolled patients	ITT patients	PP patients
		(N=XX)	(N=XX)	(N=XX)
Site #	Site Name			
XX		XX (XX%)	XX (XX%)	XX (XX%)
XX		XX (XX%)	XX (XX%)	XX (XX%)

4.1.2. Table 2. Demographics

	ITT Population N = XX	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
Age	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Height	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Weight	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
BMI	XX±X	[XX - XX]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	XX (XX, XX) (XX, XX)	
	n/N (%)	[95% CI]
Male Gender	XX/XX (XX.X%)	[XX% - XX%]
Female Gender	XX/XX (XX.X%)	[XX - XX%]

4.1.3. Table 3. Baseline Medical History

	ITT Population N = XX	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
Left atrial diameter (cm)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
AF history (number of months between Date of first documented AF diagnosis and Enrollment date)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Number of AF episodes past year	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
AF history (number of between date of most recent AF episode and Enrollment date)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Heart rate	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Blood pressure	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
LVEF (%)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
CHA2DS2VASC	XX±X XX (XX, XX)	[XX - XX]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.



STATISTICAL ANALYSIS PLAN

FARAPULSE_PersAFOne II SAP
V1.0

PERSAFONE II
V1.0

Page 28 /42

	(XX, XX)	
	n/N (%)	[95% CI]
Patient that failed any AAD	XX/XX (XX.X%)	[XX% - XX%]
If Yes: Class I	XX/XX (XX.X%)	[XX% - XX%]
If Yes: Beta Blocker/Class II	XX/XX (XX.X%)	[XX% - XX%]
If Yes: Class III	XX/XX (XX.X%)	[XX% - XX%]
If Yes: Calcium Channel Blocker/Class IV	XX/XX (XX.X%)	[XX% - XX%]
Hypertension	XX/XX (XX.X%)	[XX% - XX%]
CT/MRI results		
Left Superior Pulmonary Vein: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Left Inferior Pulmonary Vein: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Right Superior Pulmonary Vein: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Right Inferior Pulmonary Vein: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Mitral Valve: Result		

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.



STATISTICAL ANALYSIS PLAN

FARAPULSE_PersAFOne II SAP
V1.0

PERSAFONE II
V1.0

Page 29 /42

Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Left Ventricule: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Right Atrium: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Other finding 1: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Other finding 2: Result		
Normal	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Not Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Abnormal Clinically Significant	XX/XX (XX.X%)	[XX% - XX%]
Echocardiography		
Protocol-required imaging for exclusion of LA thrombus performed		
Yes	XX/XX (XX.X%)	[XX% - XX%]
No	XX/XX (XX.X%)	[XX% - XX%]
Thrombus present		
Yes	XX/XX (XX.X%)	[XX% - XX%]
No	XX/XX	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 30 /42
-------------	----------------------------------	--

	(XX.X%)	
--	---------	--

4.1.4. Table 4. Procedure

	ITT Population N = XX	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
Procedure time (hours)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
FARAWAVE dwell time (hours)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
PVI ablation time (hours)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
FARAPoint CTI dwell time (hours)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
LAPW ablation time (hours)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Fluoroscopy time (min)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Contrast volume (cc)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
IV fluid volume (cc)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
Total number of PVI applications	XX	NA
Number of applications/patient	XX±X XX	[XX - XX]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	(XX, XX) (XX, XX)	
Number of applications/vein	XX	NA
Number of LAPW applications	XX	NA
	N, n, n/N (%)	95% CI
Total number of veins treated	XX	NA
Total number of veins isolated	XX	NA
Of the total number of veins treated, the number of veins isolated	XX/XX (XX.X%)	[XX% - XX%]
CTI acute isolation	XX/XX (XX.X%)	[XX% - XX%]
Total number of CTI applications	XX	NA

4.1.5. Table 5. Remap at 60 Days

	ITT Population N = 25	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
Days follow up at remap	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
	n/N (%)	[95% CI]
Durable PVI	XX/XX (XX.X%)	[XX% - XX%]
LSPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
LIPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
LCPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
RSPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
RIPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
RMPV chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
CTI chronic isolation	XX/XX (XX.X%)	[XX% - XX%]
LAPW chronic isolation	XX/XX	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 32 /42
-------------	----------------------------------	--

	(XX.X%)	
--	---------	--

4.1.6. Table 6. Per Remapped Veins Analysis

	Veins Remapped N = XX	
	n/N (%)	[95% CI]
Number of veins durably isolated	XX/XX (XX.X%)	[XX% - XX%]
Number of veins with gaps	XX/XX (XX.X%)	[XX% - XX%]
Missing Data	XX	NA

4.1.7. Table 7. Arrhythmia Recurrence at 12 Months

	ITT Population N = XX	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
TTM compliance	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
	n/N (%)	[95% CI]
Holter compliance (6 month)	XX/XX (XX.X%)	[XX% - XX%]
Holter compliance (12 month)		
Freedom from AF/AFL/AT	XX/XX (XX.X%)	[XX% - XX%]
On AADs	XX/XX (XX.X%)	[XX% - XX%]
Off AADs	XX/XX (XX.X%)	[XX% - XX%]
Freedom from AF	XX/XX (XX.X%)	[XX% - XX%]
On AADs	XX/XX (XX.X%)	[XX% - XX%]
Off AADs	XX/XX (XX.X%)	[XX% - XX%]
Freedom from AF/AFL/AT, no DCCV>90 days, no RF touch up, no ablation >90 days	XX/XX (XX.X%)	[XX% - XX%]
On AADs	XX/XX	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	(XX.X%)	
Off AADs	XX/XX (XX.X%)	[XX% - XX%]
Freedom from AF, no DCCV>90 days, no RF touch up, no ablation >90 days	XX/XX (XX.X%)	[XX% - XX%]
On AADs	XX/XX (XX.X%)	[XX% - XX%]
Off AADs	XX/XX (XX.X%)	[XX% - XX%]

4.1.8. Table 8. Primary Safety Endpoint –Composite Safety Endpoint (CSE)

	ITT Population N = XX	
	n/N (%)	[95% CI]
Early and Late Onset Composite Safety Endpoint	XX/XX (XX.X%)	[XX% - XX%]
Total Early (within 30 days of Index or Remap Procedure)	XX/XX (XX.X%)	[XX% - XX%]
Death	XX/XX (XX.X%)	[XX% - XX%]
Myocardial infarction (MI)	XX/XX (XX.X%)	[XX% - XX%]
Persistent diaphragmatic paralysis	XX/XX (XX.X%)	[XX% - XX%]
Stroke	XX/XX (XX.X%)	[XX% - XX%]
Transient ischemic attack (TIA)	XX/XX (XX.X%)	[XX% - XX%]
Peripheral or organ thromboembolism	XX/XX (XX.X%)	[XX% - XX%]
Pericarditis	XX/XX (XX.X%)	[XX% - XX%]
Cardiac tamponade / perforation	XX/XX (XX.X%)	[XX% - XX%]
Vascular access complications requiring intervention	XX/XX (XX.X%)	[XX% - XX%]
Heart block	XX/XX (XX.X%)	[XX% - XX%]
Total Late (any time during follow-up through 12 months)	XX/XX (XX.X%)	[XX% - XX%]
Pulmonary vein (PV) stenosis (> 70% diameter reduction from baseline)	XX/XX (XX.X%)	[XX% - XX%]
Atrio-esophageal fistula	XX/XX (XX.X%)	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 34 /42
---	----------------------------------	--

4.1.9. Table 9. Primary Feasibility Endpoint – Acute Procedural Success (ITT population)

	ITT Population N = XX	
	n/N (%)	[95% CI]
Acute Procedural Success*	XX/XX (XX.X%)	[XX% - XX%]

4.1.10. Table 10. Primary Feasibility Endpoint – Acute Procedural Success (PP population)

	PP Population N = XX	
	n/N (%)	[95% CI]
Acute Procedural Success*	XX/XX (XX.X%)	[XX% - XX%]

4.1.11. Table 11. Additional Endpoints

	ITT Population N = XX	
	Mean±S.D. Median (min, max) (Q1, Q3)	[95% CI]
Time to Therapeutic Failure (days)	XX±X XX (XX, XX) (XX, XX)	[XX - XX]
	n/N (%)	[95% CI]
Proportion of subjects with Early recurrence of atrial fibrillation (ERAF) by 90 days after the index procedure	XX/XX (XX.X%)	[XX% - XX%]
Proportion of all ablated pulmonary veins that are isolated at the Index Procedure using the study device	XX/XX (XX.X%)	[XX% - XX%]
Proportion of all ablated pulmonary veins isolated using the study device during the Index Procedure that remain isolated at the 60-day remapping procedure	XX/XX (XX.X%)	[XX% - XX%]
The proportion of attempted subjects that achieve Chronic CTI Success, defined as persistent bi-directional electrical block	XX/XX (XX.X%)	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 35 /42
-------------	----------------------------------	--

across the CTI assessed at the 60-day Remapping Procedure		
The proportion of attempted subjects that achieve Chronic Focal Success, defined as persistent electrical nonconductivity of all extra-PV tissue targeted for ablation in the Index Procedure, excluding the CTI, assessed at 60-day Remapping Procedure	XX/XX (XX.X%)	[XX% - XX%]

4.1.12. Table 12. Secondary Safety Endpoints

	ITT Population N = XX	
	n/N (%)	[95% CI]
Secondary Safety Endpoint (Early and Late Onset Composite Safety Endpoint) assessed at 7 days	XX/XX (XX.X%)	[XX% - XX%]
Total Early (within 7 days of Index or Remap Procedure)	XX/XX (XX.X%)	[XX% - XX%]
Death	XX/XX (XX.X%)	[XX% - XX%]
Myocardial infarction (MI)	XX/XX (XX.X%)	[XX% - XX%]
Persistent diaphragmatic paralysis	XX/XX (XX.X%)	[XX% - XX%]
Stroke	XX/XX (XX.X%)	[XX% - XX%]
Transient ischemic attack (TIA)	XX/XX (XX.X%)	[XX% - XX%]
Peripheral or organ thromboembolism	XX/XX (XX.X%)	[XX% - XX%]
Pericarditis	XX/XX (XX.X%)	[XX% - XX%]
Cardiac tamponade / perforation	XX/XX (XX.X%)	[XX% - XX%]
Vascular access complications requiring intervention	XX/XX (XX.X%)	[XX% - XX%]
Heart block	XX/XX (XX.X%)	[XX% - XX%]
Total Late (any time during follow-up through 12 months)	XX/XX (XX.X%)	[XX% - XX%]
Pulmonary vein (PV) stenosis (> 70% diameter reduction from baseline)	XX/XX (XX.X%)	[XX% - XX%]
Atrio-esophageal fistula	XX/XX (XX.X%)	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 36 /42
---	----------------------------------	---

Device-Related Serious Adverse Events (SAEs)	XX/XX (XX.X%)	[XX% - XX%]
Procedure-Related Serious Adverse Events (SAEs)	XX/XX (XX.X%)	[XX% - XX%]
Stroke	XX/XX (XX.X%)	[XX% - XX%]
Transient Ischemic Attack (TIA)	XX/XX (XX.X%)	[XX% - XX%]
Cardioversion	XX/XX (XX.X%)	[XX% - XX%]
Arrhythmia-related (AF, AFL, AT) hospitalization	XX/XX (XX.X%)	[XX% - XX%]

4.1.13. Table 13. Secondary Feasibility Endpoints

	ITT Population N = XX	
	n/N (%)	[95% CI]
Chronic Procedural Success (CPS)* for subjects who undergo the protocol-specified 60-Day Remapping Procedure	XX/XX (XX.X%)	[XX% - XX%]
Acute Cavotricuspid isthmus (CTI) Success**	XX/XX (XX.X%)	[XX% - XX%]
Therapeutic Success***	XX/XX (XX.X%)	[XX% - XX%]
On AADs	XX/XX (XX.X%)	[XX% - XX%]
Off AADs	XX/XX (XX.X%)	[XX% - XX%]
Durability of the CTI and/or other extra PV lesion set(s)		
Durability of CTI at remap	XX/XX (XX.X%)	[XX% - XX%]
Durability of Mitral isthmus at remap	XX/XX (XX.X%)	[XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 37 /42
-------------	----------------------------------	--

Durability of LAPW at remap	XX/XX (XX.X%)	[XX% - XX%]
Durability of Left atrial roof line at remap		
Durability of Left atrial Posterior line		

4.1.14. Table 14. Adverse Events (AEs) (not including SAEs)

	ITT Population N = XX		
SOC Term/PT Term	Total Number of Events	Number of patients n	n/N (%) [95% CI]
ALL	XX	XX	XX/XX (XX.X%) [XX% - XX%]
SOC1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT2	XX	XX	XX/XX (XX.X%) [XX% - XX%]
SOC2	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT3	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT4	XX	XX	XX/XX (XX.X%) [XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 38 /42
-------------	----------------------------------	--

4.1.15. Table 15. Serious Adverse Events (SAEs)

	ITT Population N = XX		
SOC Term/PT Term	Total Number of Events	Number of patients n	n/N (%) [95% CI]
ALL	XX	XX	XX/XX (XX.X%) [XX% - XX%]
SOC1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT2	XX	XX	XX/XX (XX.X%) [XX% - XX%]
SOC2	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT3	XX	XX	XX/XX (XX.X%) [XX% - XX%]
PT4	XX	XX	XX/XX (XX.X%) [XX% - XX%]

4.1.16. Table 16. All Deaths

	ITT Population N = XX	
	n/N (%)	[95% CI]
All-cause Mortality	XX/XX (XX.X%)	[XX% - XX%]

4.1.17. Table 17. Protocol Deviations according investigator without COVID-19

	ITT Population N = XX		
Type of Deviation	Total Number of Events	Number of patients n	n/N (%) [95% CI]
Protocol Deviation Type 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description n	XX	XX	XX/XX (XX.X%)

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

ICON	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 39 /42
-------------	----------------------------------	--

			[XX% - XX%]
Protocol Deviation Type n	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description n	XX	XX	XX/XX (XX.X%) [XX% - XX%]

4.1.18. Table 18. Protocol Deviations according investigator with COVID-19

Type of Deviation	ITT Population N = XX		
	Total Number of Events	Number of patients n	n/N (%) [95% CI]
Protocol Deviation Type 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description n	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Type n	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description 1	XX	XX	XX/XX (XX.X%) [XX% - XX%]
Protocol Deviation Description n	XX	XX	XX/XX (XX.X%) [XX% - XX%]

4.1.19. Table 19. Device Deficiency Summary Table

	ITT Population N = 25
Device deficiency rate	XX/XX (XX%) [XX% - XX%]

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

	STATISTICAL ANALYSIS PLAN	FARAPULSE_PersAFOne II SAP V1.0 PERSAFONE II V1.0 Page 40 /42
---	----------------------------------	--

4.2. Listings

- 4.2.1. **Listing 1: Patients populations, reasons of non-inclusion in the different population, withdrawals and follow-up duration**
- 4.2.2. **Listing 2: Protocol deviations according the investigator**
- 4.2.3. **Listing 3: Demographics at baseline**
- 4.2.4. **Listing 4: AF History**
- 4.2.5. **Listing 5: Medical History (1/3)**
- 4.2.6. **Listing 6: Medical History (2/3)**
- 4.2.7. **Listing 7: Medical History (3/3)**
- 4.2.8. **Listing 8: Procedure data (1/4)**
- 4.2.9. **Listing 9: Procedure data (2/4)**
- 4.2.10. **Listing 10: Procedure data (3/4)**
- 4.2.11. **Listing 11: Procedure data (4/4)**
- 4.2.12. **Listing 12: Additional Ablation**
- 4.2.13. **Listing 13: Chronic Procedural Success at remap**
- 4.2.14. **Listing 14: Primary Safety endpoints**
- 4.2.15. **Listing 15: Device Deficiencies**
- 4.2.16. **Listing 16: Adverse events according CEC review**
- 4.2.17. **Listing 17: Serious adverse events according CEC review**
- 4.2.18. **Listing 18: Concomitant Medications**


CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

4.3. Graphs

4.3.1. Graph 1: Patient disposition

4.3.2. Graph 2: Freedom from AF/AFL/AT on ADDs



- Summary table

	Group 1 (N=XXX)
Nb of patients	XXX
Nb of patients with an event	XX (XX.X%)
Nb of patients without an event	XX (XX.X%)
Survival rate	XX %
Time to event (days)	
Median	XX
(95% CI)	(XX, XX)
25th-75th percentile	XXX – XXX
Range	X.XX - X.XX

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.

- 4.3.3. Graph 3: Freedom from AF/AFL/AT off ADDs**
- 4.3.4. Graph 4: Freedom from AF on AADs**
- 4.3.5. Graph 5: Freedom from AF off AADs**
- 4.3.6. Graph 6: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days**
- 4.3.7. Graph 7: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days on AADS**
- 4.3.8. Graph 8: Freedom from AF/AFL/AT, no DCCV>90 days, no RF, no ablation >90 days off AADS**
- 4.3.9. Graph 9: Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days on AADs**
- 4.3.10. Graph 10: Freedom from AF, no DCCV>90 days, no RF, no ablation >90 days off AADs**
- 4.3.11. Graph 11: Death**

CONFIDENTIEL / CONFIDENTIAL

Toute reproduction ou distribution, totale ou partielle, de ce document, sans autorisation préalable de ICON est interdite.
Reproduction and distribution of this document, in whole or in part, is strictly forbidden without prior authorization from ICON.