



**CLINICAL STUDY PROTOCOL
CP-0008 Rev. 12.7**

REVISION DATE: MARCH 30, 2020

**Prospective, Multicenter, Single Arm Safety and Effectiveness
Confirmatory Study of Endovascular Abdominal Aortic
Aneurysm Repair using the Nellix® System**

EVAS II Confirmatory Study

Table of Contents

<u>Section</u>	<u>Page</u>
1. INVESTIGATOR SIGNATURE PAGE.....	22
2. STUDY CONTACT PERSONNEL.....	23
2.1. SPONSOR	23
2.2. CONTACT NAMES AND TELEPHONE NUMBERS.....	23
3. PROTOCOL SYNOPSIS	24
4. TRIAL OVERVIEW	28
4.1. OBJECTIVE	28
4.2. BACKGROUND	28
4.3. STUDY DESIGN	30
5. STUDY POPULATION	31
5.1. ENROLLMENT	31
5.2. INCLUSION CRITERIA	31
5.3. EXCLUSION CRITERIA	32
6. RESPONSE MEASURES	32
6.1. PRIMARY SAFETY	32
6.2. PRIMARY EFFECTIVENESS.....	33
6.3. ADDITIONAL EVALUATIONS	34
7. STUDY MATERIALS.....	35
7.1. DEVICE DESCRIPTION	35
7.2. INVESTIGATOR TRAINING AND EXPERIENCE	38
7.3. DEVICE ACCOUNTABILITY	38
7.4. PATIENT AND DEVICE PREPARATION	39
8. STUDY METHODS.....	39
8.1. GENERAL ENTRY PROCEDURES.....	39
8.2. INFORMED CONSENT	39
8.3. SUBJECT IDENTIFICATION	39
8.4. SCHEDULE OF MEASUREMENTS	40
8.5. STUDY PROCEDURES AND EVALUATIONS	40
8.5.1. Informed Consent.....	40
8.5.2. CT Scan Protocol	40
8.5.3. Pre-Operative Activities.....	41
8.5.4. Subject Enrollment.....	41
8.5.5. Post-Operative Follow-up Visits	41
8.5.6. Enhanced Post-Operative Follow-up Visits	42
8.6. RECOMMENDED SECONDARY INTERVENTION TREATMENT OPTIONS	42
8.6.1. Treatment Options for Type 1 Endoleaks.....	43
8.6.1.1. Type 1A Endoleaks.....	43
8.6.1.2. Type 1B Endoleaks	43
8.6.2. Treatment Options for Migration	44
8.6.3. Treatment Options for Aneurysm Expansion.....	44

8.7. PROTOCOL DEVIATIONS	45
8.7.1. Deviations with Expedited Reporting Requirements	45
8.7.2. Deviations Requiring Prior Approval.....	45
8.7.3. Non-Urgent Deviations	45
8.8. ADVERSE EVENT REPORTING	46
8.8.1. General Definitions	46
8.8.2. Reporting of Adverse Events	47
8.8.3. Adverse Events: required reporting timeframes.....	48
8.8.4. Secondary procedures	48
8.8.5. Anticipated Adverse Events	48
8.8.6. Unanticipated Adverse Device Effect	49
8.8.7. Adverse Event Definitions	49
9. RISK ANALYSIS	51
10. STATISTICAL PROCEDURES	52
10.1 ANALYSIS POPULATIONS	52
10.2 PATIENT ACCOUNTABILITY AND MISSING DATA	52
10.3 GENERAL NOTES	52
10.4 BASELINE COMPARABILITY ANALYSIS.....	52
10.5 POOLING	53
10.6 PRIMARY SAFETY ANALYSIS.....	53
10.6.1 PRIMARY SAFETY HYPOTHESIS	53
10.7 SAMPLE SIZE.....	53
10.8 ADDITIONAL EVALUATIONS	53
11. ADMINISTRATIVE PROCEDURES	55
11.1 STATEMENT OF COMPLIANCE	55
11.2 RESPONSIBILITIES	55
11.3 PATIENT PROTECTION.....	56
11.4 PROTOCOL CHANGES	57
11.5 DOCUMENTATION	57
11.6 MONITORING PLAN	58
12. REFERENCES.....	59
13. APPENDIX A: LIST OF COILS INVESTIGATORS MAY USE ADJUNCTIVELY WITH ONYX®...	60

List of Attachments

- Attachment 1: Instructions for Use
Attachment 2: Informed Consent Form Template
Attachment 3: Electronic Case Report Form

Revision 02 Change Summary

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 01 to version 02, Date 03 July 2012 to 9 Dec 2012. Reason: Reflected revision date change in the header.
Table of Contents	Updated to reflect revision changes
1 (4)	Updated Investigator Signature page Reason: Modified first paragraph to reflect accordance with all applicable global laws.
2.1 (5)	Updated General Sponsor Contact Reason: to reflect change in Sponsor personnel.
3 (6) 5.1 (13)	Updated subject population Reason: to clarify the number of Roll-in subjects and up to 150 patients in the Intent to Treat Pivotal Cohort.
3 (6) 7.2 (18)	Updated Physician Training Reason: The training requirement has been changed to 3 required cases rather than 5 to be cleared for study enrollment.
4.1 (10)	Updated Objective: Added EndoVascular Aneurysm Sealing System Reason: To reflect updated product name
5.2 (13)	Inclusion Criteria has been updated: changed d. proximal non-aneurysmal aortic neck; length \geq 10mm; lumen diameter 16 to 32mm; angle \leq 60° to the aneurysm sac added f. Common Iliac artery lumen diameter between 8 and \leq 35mm with blood lumen diameter with blood lumen diameter \leq 35mm g. Ability to preserve at least one hypogastric artery Reason: to reflect Anatomic eligibility for the Nellix System per the instructions for use
6.2 (14)	Primary Effectiveness has been updated. Removed defined as migration resulting in serious adverse event or requiring secondary intervention through 12 months. Revised Secondary endovascular procedure for resolution for endoleak added (type I or type III) Reason: To clarify safety and effectiveness
6.3 (15)	Updated Additional Evaluations. Revised Secondary endovascular procedure within 30 days, at 6 months, and annually to 5 years for resolution of endoleak Type I and II. Reason: to identify the procedure and resolution of endoleak Type I and III
4.2 (11)	Updated Background to include ePTFE covered stent. Reason: To reflect the Instructions for Use
8.6 (23)	Updated Protocol Deviation Reason: to clearly define a protocol deviation
8.7.1 (24)	Updated Serious Adverse Event Definition Reason: For alignment with ISO 14155 definition.
8.7.3 (26)	Added reference to Social Security Death Index (SSD) and CDC National Death Index (NDI) to obtain patient death information. Reason: Added additional information for sites to acquire patient death information
(43)	Updated monitoring will be performed under the direction of Avi Sharma, Director, Clinical Affairs Reason: A new CRO will be contracted.
Attachment 2 (2)	Revised Informed Consent section 3. Up to 180 patients will be enrolled in the study overall. Reason: To reflect the change in the protocol.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

Section (Page)	Change and Reason for the Change
Attachment 2 (3)	Revised Informed Consent section 4. The US sites will be conducted under an IDE, or Canada under an investigational test authorization.
Attachment 3	Formatting changed to DD/MMM/YY throughout all Case Report Forms Reason: To correct formatting
Attachment 3 (1)	Revised Screening and Baseline case report form inclusion criteria section 3. Patient aneurysm increase of 1.0cm. Reason: to reflect the IFU
Attachment 3 (1)	Revised Screening and Baseline case report form inclusion criteria to include in section 4. Most caudal renal artery to aortoiliac bifurcation length \geq 70 mm Reason: to reflect the current IFU

Revision 03 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 02 to version 03, Date 09 Dec 2012 to 12 June 2013. Reason: Reflected revision date change in the header.
Table of Contents	Updated to reflect revision changes
1 (13)	‘Global’ changed to ‘regional’ Reason: PIs follow regional regulations. Protocol requirements will specify any non-regional requirements.
2 (23)	CEC and DSMB Name and Address added (Syntactx) Reason: CRO is now under contract with Endologix.
3 (24)	Regions of study populations added Reason: Clarification of enrollment sites by region.
3 (24)	Site enrollment maximum set to 10% Reason: To ensure a more equitable enrollment profile across sites
3(25) 7.2 (38)	Training section – changed from 3 roll-in per site to 1 roll-in. Grammatical changes. Reason: Reflect correct roll-in population as specified in Subject Population (page 10) and other sections.
3 (25) 5.3 (32)	Exclusion Criteria #3: Removed words ‘enrollment or 30 day follow-up phase of’ Reason: Subjects should not be involved in a competing study at any time frame, as the study population should be examining only the Nellix system as a treatment.
3 (26) 6.2 (33)	Primary Effectiveness Endpoint: “Endoleak type I > 30 days, and Endoleak III” changed to “Endoleak type I or Endoleak type III at 12 months” and additionally “Secondary endovascular procedure up to 12 months for resolution of Endoleak (Type I or Type III)” was added. Reason: The existence of an endoleak is, by itself, not necessarily an indication of a clinically significant leakage. Clinically, a failure of effectiveness related to endoleaks would be indicated by a re-intervention to re-exclude the aneurysm, or the existence of an endoleak sufficient to cause the aneurysm to expand. Both of those occurrences are considered failures by this endpoint. Type I or III Endoleaks at 12 months (regardless of intervention or effect) is included to conservatively allow for possible future corrective actions.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

Section (Page)	Change and Reason for the Change
3 (26) 3(18) 10.8 (53)	Primary Effectiveness Endpoint: Added in 'device occlusion', and 'may be due to thrombus or other causes' to list of secondary procedure causes. Reason: Clarity, as requested by FDA (question 44a)
3 (26)	Primary Effectiveness Endpoint: Re-arrangement of metrics. Reason: Clarity (most of the metrics refer to secondary interventions for various device issues; these were compiled into a list).
3 (26)	Additional Evaluations: MAE Individual Components: Changed 'years 1 through 5' to 'annually to 5 years' Reason: Clarity (to match other metric descriptions in this section).
3 (26)	Additional Evaluations: Composite Major Adverse Events: added '30 days' to time frames of analysis. Reason: Error - should have been included (as in the other metrics in this section).
3 (26) 4.3 (30) 6.3 (34) 8.4 (40) 10.8 (53)	Additional Evaluations: Distal Blood Flow (ABI): Removed Reason: Per the recommendation of the study P.I., ABI (ankle brachial index) is a poor indicator of device performance. Due to the co-morbidities associated with this study population (peripheral artery disease), ABI can be permanently elevated in patients with calcified arteries, rendering ABI a poor tool for identifying future changes in blood flow. Peripheral blood flow, should of course, be examined as per physician standard of care. J Am Coll Cardiol. 2008 Apr 1;51(13):1292-8. doi: 10.1016/j.jacc.2007.11.064. doi: 10.2337/diacare.26.12.3333 <i>Diabetes Care December 2003 vol. 26 no. 12 3333-3341</i>
3 (26) 6.3 (34) 8.7.7 (46) 10.8 (53)	Additional Evaluations: Luminal Thrombus Requiring Intervention at each study timepoint: Metric and definition added Reason: Response to FDA's request concerning thrombus formation (question 46), thrombus has also been added as a separate metric for evaluation (in addition to being part of the effectiveness endpoint).
3	Statistical Considerations: Study Population: 1 st paragraph changes Reason: Clarity to Intent to treat population definition, and to enrolled status.
3	Statistical Considerations: Study Population: Per Protocol population paragraph added. Reason: Clarify the Per Protocol population definition.
3	Statistical Considerations: Study Population: Completed Cases paragraph added. Reason: Clarify the Completed Cases population definition.
7 (35)	Study Materials: Single use or multiple use added to each component. Reason: Clarity (and Dispenser should have been specified multiple use).
7.2 (38)	Investigator Training item #2: added word 'approximately' to case number requirement Reason: Sub-investigator qualifications for implantation with regard to prior EVAR experience will be done on a case-by-case basis; however, 25 cases will be treated as a rough baseline.
7.3 (38)	Device accountability: added in 'prior to implantation' Reason: Clarity. Return of all device components can only occur if the device was not implanted.
8.1 (39) 11.2 (49)	Addition of 'independent physician review. Reason: Addition of more detail regarding patient approval process.
8.2 (39)	'global standards' changed to 'Good Clinical Practices' Reason: Clarity

Section (Page)	Change and Reason for the Change
8.3 (27)	Changed title of Section 8.3 to properly reflect subsequent content Reason: Error, and as noted in FDA question #49.
8.3 (27)	Removal of methodology for patient naming. Reason: Patient IDs will be automatically generated within the eDC system.
8.5.4 (40)	Addition of sentence “If, during the index procedure, the device is not implanted and/or there is conversion to open surgery, the patient will be followed until the 30 day follow-up” Reason: Clarification of follow-up process for patients that either do not receive the study device, or are converted to open surgery.
8.7.1 (46)	‘...administered a product...’ changed to ‘...enrolled in the trial...’ Reason: Correction. AE reporting also applies to subjects who did not receive the product, but were enrolled.
8.7.2 (46) 8.7.3 (46)	Sections rewrite Reason: Consolidation. Many paragraphs in these two sections contained repetitious instructions. The adverse event reporting process is roughly the same regardless of the type of adverse event. AE reporting time requirements (from the site to Endologix), while mentioned, were scattered throughout text and difficult to access quickly. A table of “Adverse Events: required reporting timeframes” was generated to address this. Requirements for reporting and investigating a subject death were consolidated into one section. There were no substantive changes.
8.7.4 (46)	Removal of word ‘endovascular’ from section title Reason: Correction. Procedures discussed also refer to open repair.
8.7.4 (46)	Removal of fax information. Reason: reporting in this trial will be electronic via EDC.
8.7.6 (46)	Addition of Endologix reporting timeframes of UADEs. Removal of fax information Reason: Reporting will be done by the EDC system in this trial. Clarify Endologix response to such an event.
8.7.7 (46) 10.8 (53)	Revised definition for Clinically Significant Migration: Reason: Updated for consistency within the document.
8.7.7 (46)	Removal of definitions of various adverse events Reason: Consolidation. Definitions in the protocol will be limited to endpoints and device specific issues. All other clinical events will be defined by the independent CEC (Syntactx)
8.7.7 (46)	Revised Distal Ischemia Definition: Removed”...A reduction in ABI of 0.15 or more attributable to the index procedure and not related to natural progression of atherosclerotic disease is included in this definition. Reason: See change #17
8.7.7 (46) 10.8 (53)	Revised Migration definition: Reason: Defined per migration from the renal artery and removed intervention criteria to include all device migrations >5 mm.
8.7.7 (46)	Addition of device occlusion to secondary procedures. Reason: Treatment can also be open surgery. Obstruction and occlusion will cover any thrombus related events (FDA Question 44).

Section (Page)	Change and Reason for the Change
8.7.7 (46)	Removal of word ‘endovascular’ from Secondary Procedure definition Reason: See change#47
10.1 (52)	Removal of ‘with imputation for missing data...’ Reason: Removed for redundancy as this is explained in the next section.
10.1 (52)	Addition of ‘or have expired prior to 12 months’ Reason: Per FDA request (Question 47), provide further clarification on the Completed Cases analysis population.
10.1 (52)	Completed case (CC) changes; removal of one sentence, addition of another Reason: Clarity.
10.2 (52)	Addition ‘prior to procedure’ Reason: Clarity (this listing is for patients who never officially enrolled)
10.2 (52)	Change of ‘15 subjects’ to ‘10% of study enrollment’ (2 places in paragraph) Reason: Quantify the amount of total enrollment a given clinical site can contribute to the study.
10.2 (52)	Removal of sentence “In the control group, no subjects were...” Reason: Per FDA request (Question 51), there will be no control group comparator for this study.
10.2 (52)	Addition of paragraph beginning with “In addition to the above...” Reason: Per FDA request (Question 52), Tipping Point Analyses has been included.
10.5 (53)	Addition of sentence beginning “The estimates obtained...” Reason: Per FDA request (Question 53), provided additional clarification pertaining to heterogeneity of study sites.
10.6 (53)	For entire section: removal of word ‘control’. Reason: Clarity. SVS open surgery group provides a historical OPC for this trial. The SVS group is not a direct control (i.e., patient level data) in this trial.
10.6 (41)	For entire section: change ‘test group’ to ‘study population’ Reason: Clarity. ‘Test group’ normally refers to one arm of a two arm trial. This trial is not designed as a two arm trial.
10.6.2	Changes to sentence beginning with ‘The sample size of this trial...’ Reason: Clarity. ‘Test group’ normally refers to one arm of a two arm trial. This trial is not designed as a two arm trial.
10.7.1 (51)	Addition of sentence beginning with “The hypothesis test will be...” Reason: Clarification of the significance level and for consistency within the document.
10.7.2 (53)	Revision of endpoint description Reason: Updated to reflect endpoint changes.
11.2.5 (48)	Added “An independent clinical safety review will examine all AEs and SAEs. MAEs will undergo adjudication by the CEC panel.” Reason: Events that the CEC will review were clarified per FDA IDE response question #11.
11.5 (57)	Revised Study log maintenance description Reason: Updated to reflect current practices for study log maintenance.
Attachment 2 (2)	Added prior clinical studies information to informed consent Reason: This change was implemented per a request in FDA IDE Response question #13.

Section (Page)	Change and Reason for the Change
Attachment 2 (4)	<p>Added information pertaining to follow-up of enrolled patients where the study device is not successfully implanted during the index procedure or patients where the device is explanted.</p> <p>Reason: Protocol CP-0008 Rev.03 was updated with a 30-day follow-up period for patients where the device is not successfully implanted, if the patient is converted to open surgery, or if the device is explanted with open surgical repair.</p>
Attachment 2 (9)	<p>Added US Clinical Trials.gov publishing requirement</p> <p>Reason: To inform patients that information on this study will be available on the internet.</p>
Attachment 3 (All)	Paper CRFs were converted to electronic CRFs
Attachment 3 (5)	<p>Revised from: "Most caudal renal artery to aorto-iliac bifurcation length \geq 70 mm". To: "Most caudal renal artery to each hypogastric artery length \geq 100mm"</p> <p>Reason: Updated to reflect current inclusion exclusion criteria per discrepancy noted in FDA IDE response question #14.</p>
Attachment 3 (5)	<p>Revised from: "Is there clinically significant infrarenal mural thrombus (>5mm thick over >60% circumference)" To:...(>5 mm thick over >50% circumference)</p> <p>Reason: Updated to reflect current inclusion exclusion criteria per discrepancy noted in FDA IDE response question #14.</p>
Attachment 3 (17)	<p>Removed: Nellix Extenders were removed from the CRFs</p> <p>Reason: Extenders are not planned to be used in this clinical study. This addresses FDA IDE response question #12</p>

Revision 04 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 03 to version 04, Date 12 June 2013 to May 2014. Reason: Reflected revision date change in the header.
2 (13)	Added "...with conditions of the approval imposed by the reviewing Ethics Committee or Institutional Review Board....I agree to maintain adequate source documentation records throughout the clinical investigation and make them available as requested during monitoring visits. I agree to maintain the device accountability records and ensure that the investigational device is used solely by authorized users as specified by this protocol. I agree to ensure that the requirements for obtaining informed consent are met. Additionally, I agree to disclose financial interests in accordance with 21 CFR 54, and certify that such financial interests, if any, will not interfere with my responsibilities as an investigator or influence study outcomes under my supervision. Reason: These statements were added to the investigator signature page to centralize requirements from 21 CFR 812.43 (C) and ISO 14155 that were reflected across several study agreements.
3 (15)	Removal of Canada from the list of countries Reason: Canada no longer has active sites in this trial. No patients were enrolled at either site.
3 (15)	Change the site maximum enrollment from 10% to 15%. Reason: For consistency with prior EVAR trials.
3 (16)	Revised Inclusion criteria #4 for the maximum abdominal aortic aneurysm (AAA) sac diameter size from: " $\geq 5.5\text{cm}$ or $\geq 4.5\text{ cm}$ which has increased by $>1.0\text{cm}$ within the last year" to: " $\geq 5.0\text{cm}$, or $\geq 4.5\text{ cm}$ which has increased by $\geq 0.5\text{cm}$ within the last 6 months, or which exceeds 1.5 times the transverse dimension of an adjacent non-aneurysmal aortic segment". Reason: As recommended by the trial Physician Steering Committee, for alignment with SVS practice guideline recommendations and aneurysm dimensional inclusion criteria among most recently approved products/labeling, including Medtronic Endurant, Trivascular Ovation, Lombard Anaconda, and Endologix (PEVAR).
3 (16)	Revised Exclusion criteria#10 of section #3 of clinically significant mural thrombus from "Clinically significant infrarenal mural thrombus ($>5\text{mm}$ thickness over $>50\%$ circumference)" to "Clinically significant mural thrombus within the proximal landing zone (minimum 10mm) of the infrarenal non-aneurysmal neck ($>5\text{mm}$ thickness over $>50\%$ circumference)" Reason: Clarification of exclusion criterion. This revision clarifies that the mural thrombus exclusion applies to the infrarenal non-aneurysmal neck.
4.2 (20)	Added text beginning with "There are a number of qualifiers..." Reason: For clarification only.
5.1 (22)	Change the site maximum enrollment from 10% to 15% Reason: For consistency with prior EVAR trials.

Section (Page)	Change and Reason for the Change
5.2(22)	<p>Revised Inclusion criteria #4 for the maximum abdominal aortic aneurysm (AAA) sac diameter size from: “$\geq 5.5\text{cm}$ or $\geq 4.5\text{ cm}$ which has increased by $>1.0\text{cm}$ within the last year” to: “$\geq 5.0\text{cm}$, or $\geq 4.5\text{ cm}$ which has increased by $\geq 0.5\text{cm}$ within the last 6 months, or which exceeds 1.5 times the transverse dimension of an adjacent non-aneurysmal aortic segment”.</p> <p>Reason: As recommended by the trial Physician Steering Committee, for alignment with SVS practice guideline recommendations and aneurysm dimensional inclusion criteria among most recently approved products/labeling, including Medtronic Endurant, Trivascular Ovation, Lombard Anaconda, and Endologix (PEVAR).</p>
5.3 (23)	<p>Revised Exclusion criteria #10 of section 5.3 of clinically significant mural thrombus from “Clinically significant infrarenal mural thrombus ($>5\text{mm}$ thickness over $>50\%$ circumference)” to “Clinically significant mural thrombus within the proximal landing zone (minimum 10mm) of the infrarenal non-aneurysmal neck ($>5\text{mm}$ thickness over $>50\%$ circumference)”</p> <p>Reason: Clarification of exclusion criterion.</p>
7.2 (28)	<p>Revised review requirement for investigator training from “review of Investigator Brochure” to “review of prior clinical safety and effectiveness information”.</p> <p>Reason: To reflect actual practice.</p>
7.2 (28)	<p>Added criteria for sub-Investigators who lead the implant procedures “For procedures led by a sub-investigator, the PI must be present to provide supervision of the case, unless the sub-Investigator has completed investigational device training and has received documented approval by Endologix.”</p> <p>Reason: Clarification of possible sub-investigator qualifications. Some sub-investigators will receive the full PI/proctor training.</p>
8.3 (30)	<p>Added subject number allocation system “Subject numbers will be automatically assigned by the EDC system”</p> <p>Reason: Subject number allocation was inadvertently missed in the previous revision.</p>
8.3 (30)	<p>Clarified screening log maintenance “either on paper or electronically in the EDC”</p> <p>Reason: Screening log can be maintained by either method (paper or electronically).</p>
8.5.2(30)	<p>Clarified slice thickness criteria from “$<3\text{mm}$” to “$\leq 3\text{mm}$”</p> <p>Reason: Typographical error: CT Scan slice thickness equal to 3mm are acceptable.</p>
8.5.4 (30)	<p>Revised subject enrollment section to clarify screen failures. “If the device does not contact the patient, the patient will not be considered enrolled and the assigned study number will not be given to another patient and the patient is considered as a screen failure”</p> <p>Reason: Further clarified screen failures and patient numbers assignment in screen failures.</p>
8.5.4 (31)	<p>Revised conditions, when device is not implanted during index procedure. “If, during the index procedure, the device is introduced but is not implanted and/or there is conversion to open surgery, the patient will be followed until the 30 day follow-up.”</p> <p>Reason: Further clarified situation when devices are not implanted during index procedure.</p>
8.5.5 (32)	<p>Removed site reference manual from instruction to complete, from “site reference manual” to “site initiation visit”.</p> <p>Reason: The site personnel are trained during site initiation visit on how to enter data into eCRFs.</p>
8.7.2 (35)	<p>Added “If approval is obtained from the patient or a patient’s authorized family member, the site should notify Endologix of the explant”.</p> <p>Reason: Added requirement to include patient’s authorization and to notify the sponsor once obtained.</p>

Section (Page)	Change and Reason for the Change
8.7.7 (36)	<p>Pulmonary related death was moved from page 35</p> <p>Reason: Pulmonary related death was located in the wrong section.</p>
8.7.7(36)	<p>Removed definition for Migration from “death” section to “other definitions” section</p> <p>Reason: Editorial correction only.</p>
8.7.7 (37)	<p>Added “...and is still present at the 30 day CT as read by the Core Lab” to Procedural Technical Failure definition.</p> <p>Reason: Clarify definition of unresolved endoleak.</p>
10.2 (41)	<p>Removed “...and imputing the mean from all subjects with outcomes.”</p> <p>Reason: Including this additional method is not relevant since the endpoint is dichotomous and not a continuous variable from which a mean can be calculated across patients.</p>
10.5 (42)	<p>Replaced Logistic Regression using a univariate model of study site with ... “extension of the Fisher’s exact test (Fisher-Freeman-Halton).”</p> <p>Reason: This replacement is better suited for testing homogeneity across study sites.</p>
10.5 (42)	<p>Added “Both the safety and effectiveness measures will employ weighted proportions as the basis for the statistical test.”</p> <p>Reason: Changed for clarity, as the approach to weighing estimates of the proportions as described by the Method of Fleiss (1993) is the same for all endpoints.</p>
10.6 (43)	<p>Changed: “...greater...” to “...less...”</p> <p>Reason: This was a typographical error. The sentence was updated to reflect actual statistical test that shows significant heterogeneity when the p-value is less than 0.15.</p>
10.6 (43)	<p>Removed “Secondary Procedures” from the primary safety hypothesis</p> <p>Reason: This was inadvertently added to the primary safety hypothesis as secondary procedures are represented in the primary effectiveness analysis.</p>
10.6.3 (44)	<p>Updated equation:</p> $W_k = \left(\frac{1}{n_k} p_k (1 - p_k) \right)^{-1}$ <p>Reason: There was a typographical error in the previous version of this equation. This equation was intended to show that weight is equal to the inverse of the variance.</p>
ICF (3)	<p>Study Questions, #3: Changed to ‘enroll up to 31 patients’</p> <p>Reason: to match up to maximum site enrollment</p>
ICF (4)	<p>Blood Pressure : removed descriptive text ‘... measured at your arms and ankles’</p> <p>Reason: protocol only specifies a normal blood pressure exam.</p>
ICF (10)	<p>Change “Endologix fenestrated device” to “Nellix system”</p> <p>Reason : Incorrect system title.</p>

Revision 05 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 04 to version 05, Date Jun 2014 to Nov 2014. Reason: Reflected revision date change in the header.
Table of Contents	Updated to reflect revision changes
3 (16)	Updated protocol title, both short and long, to reflect the addition of Continued Access (as a separate cohort)
3 (16)	Clarified 'Study Devices' to include full name of device
3 (18)	Enrollment section changed to clarify the two trial phases: the Primary Investigation, and the Extended Investigation (Continued Access). Added the expected Extended Investigation enrollment and closure timeframe.
5.1 (23)	Changed title of section to 'Primary Investigation' for clarity.
5.2 (23)	New section added to explain details of the Extended Investigation. The section elucidates 1) the expected enrollment, 2) the expected time of enrollment, and 3) the sites that invited to participate.
8.7.2 (36)	Updated company clinical fax number
10.9 (49)	New section added to explain statistical treatment of the Continued Access cohort.

Revision 06 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 05 to version 06, Date Nov 2014 to Feb 2015. Reason: Reflected revision date change in the header.
Table of Contents	Updated to reflect revision changes
3 (17)	Updated Continued Access population size to 100 Reason: Reflected total population size for Continued Access
5 (24)	Updated Continued Access population size to 100 Reason: Reflected total population size for Continued Access

Revision 07 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
Title Page	Updated company address Reason: 2 Musick, Irvine, CA 92618 is location of new headquarters
All Pages	Changed all pages to protocol version from 06 to version 07, Date Feb 2015 to Apr 2015. Reason: Reflected revision date change in the header.
Table of Contents	Updated to reflect revision changes

Section (Page)	Change and Reason for the Change
2 (16)	Updated company address Reason: 2 Musick, Irvine, CA 92618 is location of new headquarters
3 (18)	Updated minimum proximal non-anuerysmal aortic neck lumen diameter from 16mm to 18mm and the minimum common iliac artery lumen diameter from 8mm to 9mm Reason: Corresponds with updated IFU. The original anatomical parameters in both the IFU and the earlier protocol versions allowed for the possible use of 8mm Nellix stent diameters. The Core Lab has already been using these more restrictive parameters to determine eligibility throughout the trial, as Nellix 8mm stents were never utilized or implanted during the study.
5 (25)	Updated minimum proximal non-anuerysmal aortic neck lumen diameter from 16mm to 18mm and the minimum common iliac artery lumen diameter from 8mm to 9mm Reason: Corresponds with updated IFU. The original anatomical parameters in both the IFU and the earlier protocol versions allowed for the possible use of 8mm Nellix stent diameters. The Core Lab has already been using these more restrictive parameters to determine eligibility throughout the trial, as Nellix 8mm stents were never utilized or implanted during the study.
11 (54)	Updated company address Reason: 2 Musick, Irvine, CA 92618 is location of new headquarters

Revision 08 changes (VERSION WAS NOT RELEASED)

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 07 to version 08, Date May 2015 to Aug 2015. Reason: Reflected revision date change in the header.
3 (17)	Updated Continued Access population size to 270 Reason: Reflected total population size for Continued Access
5 (24)	Updated Continued Access population size to 270 Reason: Reflected total population size for Continued Access

Revision 08.1 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 08 to version 08.1, Date May 2015 to October 2015 Reason: Reflected revision date change in the header.
3 (18)	Updated Continued Access population size to 150 Reason: Reflected total population size for Continued Access
5 (25)	Updated Continued Access population size to 150 Reason: Reflected total population size for Continued Access

Revision 09 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 08.1 to version 09, Date October 2015 to April 2016 Reason: Reflected revision date change in the header.
3 (18)	Updated Continued Access population size to 250 Reason: Reflected total population size for Continued Access
5 (25)	Updated Continued Access population size to 250 Reason: Reflected total population size for Continued Access

Revision 10 changes (VERSION WAS NOT RELEASED)

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 09 to version 10, Date April 2016 to June 2016 Reason: Reflected revision date change in the header.
2.1 (18)	Changed sponsor contact Reason: Updated based on current organizational structure.
3 (20)	Changed anatomical inclusion criteria of blood lumen diameter from \leq 60 mm to \leq 70 mm Reason: Expanded treatable blood lumen diameters in updated Instructions for Use (IFU).
5 (27)	Changed anatomical inclusion criteria of blood lumen diameter from \leq 60 mm to \leq 70 mm Reason: Expanded treatable blood lumen diameters in updated Instructions for Use (IFU).
7 (30)	Changed dispenser from multiple use to single use Reason: Disposable dispensers are now available to use.
7.1 (30)	Replaced "polyurethane bag (EndoBag) having a polyester reinforcement sleeve" with polyurethane EndoBag with welded seams and a lumen having an inner polyester (PET) sleeve is attached to the stent proximally and distally using polyethylene sutures Reason: Describes the changes to the the EndoBag, with the distal end now also attached to the stent.
7.1 (31)	Updated the name of the pressure monitor from Mirador Compass™ to Centurion™ Biomedical Compass™ Reason: Mirador Biomedical is now a subsidiary of Centurion Medical Products, Inc.
7.1 (32)	Replaced the picture of the console Reason: Console was updated from 4 ports to 3.
8.7 (40)	Added sentence: "These events must be reported if clinically significant." Reason: To emphasize the importance of reporting any clinically significant AE.
11.6 (57)	Changed sponsor contact Reason: Updated based on current organizational structure.

Revision 10.1 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 09 to version 10.1, Date April 2016 to September 2016 Reason: Reflected revision date change in the header.
2.1 (18)	Changed sponsor contact Reason: Updated based on current organizational structure.
7 (30)	Changed dispenser from multiple use to single use Reason: Disposable dispensers are now available to use.
7.1 (30)	Replaced "polyurethane bag (EndoBag) having a polyester reinforcement sleeve" with polyurethane EndoBag with welded seams and a lumen having an inner polyester (PET) sleeve is attached to the stent proximally and distally using polyethylene sutures Reason: Describes the changes to the the EndoBag, with the distal end now also attached to the stent.
7.1 (31)	Updated the name of the pressure monitor from Mirador Compass™ to Centurion™ Biomedical Compass™ Reason: Mirador Biomedical is now a subsidiary of Centurion Medical Products, Inc.
7.1 (32)	Replaced the picture of the console Reason: Console was updated from 4 ports to 3.
8.7 (40)	Added sentence: "These events must be reported if clinically significant." Reason: To emphasize the importance of reporting any clinically significant AE.
11.6 (57)	Changed sponsor contact Reason: Updated based on current organizational structure.

Revision 11.1 changes (VERSION WAS NOT RELEASED)

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 10.1 to version 11.0 Date September 2016 to May 2017 Reason: Reflected revision date change in the header.
3 (21)	Updated proximal non-anuerysmal aortic neck lumen diameter from 18mm-32mm to a neck diameter of 18mm-28mm and added: Distal iliac artery seal zone with length of ≥ 10 mm, diameter range of 9 to 25mm and ratio of maximum aortic aneurysm diameter to maximum aortic blood lumen diameter < 1.40 . Reason: Corresponds with updated IFU.
5 (28)	Updated proximal non-anuerysmal aortic neck lumen diameter from 18mm-32mm to a neck diameter of 18mm-28mm and added: Distal iliac artery seal zone with length of ≥ 10 mm, diameter range of 9 to 25mm and ratio of maximum aortic aneurysm diameter to maximum aortic blood lumen diameter < 1.40 . Reason: Corresponds with updated IFU.
8.4 (37)	Updated table to reflect enhanced follow-up visits Reason: Schedule of Measurements summarize assessments at each timepoint in the study, including enhanced follow-up. Added footnote to table to clarify population affected by the new follow-up recommendations.

Section (Page)	Change and Reason for the Change
8.5.6 (38)	Added section for enhanced post-operative follow-up visit. Reason: Enhanced follow-ups are recommended to monitor at-risk for migration, Type 1A endoleaks and/or AAA sac enlargement.
8.6 (38)	Added section of recommended secondary intervention treatment options. Reason: Provides options to Nellix Investigators to treat subjects who experience migration, Type 1A endoleaks and/or AAA sac enlargement.

Revision 11.2 changes (VERSION WAS NOT RELEASED)

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 11.0 to version 11.2 Date June 2017 to July Reason: Reflected revision date change in the header.
8.6 (38)	Added section of recommended secondary intervention treatment options. Reason: Provides options to Nellix Investigators to treat subjects who experience migration, Type 1A endoleaks and/or AAA sac enlargement.

Revision 11.3 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 11.2 to version 11.3 Date July 2017 to August Reason: Reflected revision date change in the header.
8.6.1 (40)	Updated table to reflected recommended secondary intervention treatment options for Type 1A Endoleaks Reason: When positioned correctly, with adequate seal, Onyx® can be used to treat Type 1A endoleaks.
Appendix	Added appendix of coils to be used with the Onyx Liquid Embolics System Reason: Reflects list of peripheral vascular coils available in the US, that allows physician to use preferred standard coil.

Revision 12.0 changes (VERSION WAS NOT RELEASED)

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
Cover Page	Added Confirmatory indication Reason: Updated based on protocol being a confirmatory evaluation.
All Pages	Changed all pages to protocol version from 11.2 to version 12.0 Date July 2017 Reason: Reflected revision date change in the header.
3 (21)	Updated Protocol Synopsis Reason: Updates align with details found in Trial Overview, Study Population and Response Measures sections.
5.1 (29)	Revised to reflect study design for confirmatory study Reason: Updates align with conformational design
6.3 (32)	Updated Additional Evaluations to reflect evaluation time points used in Pivotal study Reason: Adverse events are analyzed through the duration of the 1-year endpoint.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

Section (Page)	Change and Reason for the Change
7.3 (36)	Revised Device Accountability Reason: Provided clarification of specific components of the Nellix System tracked on the device accountability log.
8.1 (37) 11.2 (60)	Updated Physician Reviewer to Case Review Board Reason: Reflects updates needed to ensure IFU is followed
8.8.7 (50)	Removed duplicate definition of <i>Clinically Significant Migration</i> . Reason: Duplication not needed.
10.6 (54)	Revised Primary Safety Reason: Details of the analysis plan are in SAP.
10.7 (55)	Revised Primary Effectiveness Reason: Details of the analysis plan are in SAP.

Revision 12.1 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
Cover Page	Added Confirmatory indication Reason: Updated based on protocol being a confirmatory evaluation.
All Pages	Changed all pages to protocol version from 12.0 to version 12.1 Date August 2017 Reason: Reflected revision date change in the header.
3 (22)	Updated total number of sites from up to 35 to up to 28 Reason: To help control enrollment distribution across sites.
3 (25)	Removed Statistical Considerations referenced in the synopsis Reason: The SAP is the reference for statistical methodology.
5 (59)	Updated total number of sites from up to 35 to up to 28 Reason: To help control enrollment distribution across sites.
5.1 (30)	Added “In the Primary Investigation, a single site may not enroll more than 15% of the total enrollment.” Reason: To clarify that there is a site enrollment maximum.
8.8.6 (48)	Clarified that type 1 endoleaks, device migration and aneurysm sac enlargement are not considered UADEs Reason: Aligns with refining IFU to mitigate those events previously noted to be at higher than expected rates.
7 (30) 7.1 (31) 7.2 (34)	Included Ovation iX Iliac Stent Graft as part of the Nellix System. Included device description and training requirements.

Revision 12.2 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.1 to version 12.2 Date September 2017 Reason: Reflected revision date change in the header.
3 (24-25) 6.3 (31)	Added Additional Evaluations from SAP Reason: Updated to align with SAP per FDA's request.

Revision 12.3 changes

Note: minor punctuation or formatting changes are not included in this table.

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.2 to version 12.3 Date March 15, 2018 Reason: Reflected revision date change in the header.
2.1 (21)	Updated General Sponsor Contact Reason: Updated based on current organizational structure
4.3 (28)	Revised language Reason: To reflect current terminology
5.1 (29)	Revised language Reason: To provide additional clarification
6.3 (32)	Relocation of existing footnotes Reason: Formatting change to reflect the accurate location of the footnote entries
7.1 (33)	Deleted existing footnotes Reason: Formatting change to reflect the accurate location of the footnote entries
7.2 (36)	Revised language Reason: To provide additional clarification
7.3 (36)	Revised language Reason: To provide additional clarification
8.2 (37)	Added language Reason: To provide additional clarification
8.4 (38)	Revised language Reason: To provide additional clarification and correct editorial error
8.5.1 (38)	Revised language Reason: To provide additional clarification
8.5.3 (39)	Revised language Reason: To provide additional clarification
8.5.6 (40)	Added language Reason: To define inadequate procedures in the revised definition of "at risk" to ensure enhanced follow-up is recommended.
8.6 (41-42)	Added secondary intervention terminology Reason: To reflect current terminology
11.6 (57)	Revised language Reason: To provide additional clarification
11.6 (58)	Updated Monitoring contact Reason: Updated based on current organizational structure

Revision 12.4 changes

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.3 to version 12.4 Date March 28, 2018. Reason: Reflected revision date change in the header.
8.5.6 (40)	Added punctuation to an inadequate procedure definition. Reason: Updated per FDA request

Revision 12.5 changes

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.4 to version 12.5 Date April, 2018. Reason: Reflected revision date change in the header.
8.5.6 (41)	Added clarification. Reason: Updated per FDA request

Revision 12.6 changes

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.5 to DRAFT September 21, 2018. Reason: Reflected revision date change in the header.
2.1 (22)	Updated General Sponsor Contact Reason: Updated based on current organizational structure
3.0 (23)	Changed 90 subjects to 105 subjects Reason: To align with the revised SAP
3.0 (24)	Clarified Total Enrollment, the Pivotal Cohort, and the Roll-In Cohort Reason: To align with the revised SAP
3.0 (24)	Changed maximum of 28 US sites to 38 US sites Reason: Additional sites required due to enrollment challenges
3.0 (25)	Changed Primary Effectiveness Endpoint to 2 year Reason: To align with the revised SAP
5.1 (30,31)	Changed 90 subjects to 105 subjects. Clarified Total Enrollment, the Pivotal Cohort, and the Roll-In Cohort. Revised Section Title. Reason: To align with the revised SAP
5.1 (30)	Changed maximum of 28 US sites to 38 US sites Reason: Additional sites required due to enrollment challenges
5.1 (31)	Changed maximum of 28 US sites to 38 US sites Reason: Additional sites required due to enrollment challenges
5.2 (31)	Revised Inclusion numbering for consistency with the Synopsis Reason: Administrative
5.2 (31)	Revised Inclusion numbering for consistency with the Synopsis Reason: Administrative
6.1 (33)	Updated to reference the SAP Reason: Administrative
6.2 (33)	Changed Primary Effectiveness Endpoint to 2 year Reason: To align with the revised SAP
6.2 (33)	Updated to reference the SAP Reason: Administrative
7.2 (38)	Deleted Roll-In Subjects section Reason: Language included in Section
8.2 (38)	Deleted reference to the Informed Consent Form Reason: Is not included in the protocol as an attachment
8.8.1 (46)	Replaced current definition of Adverse Event Reason: Updated language for consistency with ISO 14155
8.8.1 (46)	Added reference to 21 CFR 812(s) Reason: To provide additional clarification
8.8.1 (46, 47)	Added definition of device-related and procedure-related Reason: To provide additional clarification

Section (Page)	Change and Reason for the Change
8.8.2 (47)	Revised language Reason: To align with the protocol
8.8.2 (47)	Revised language Reason: To provide additional clarification
8.8.5 (49)	Added Fistula as an anticipated Adverse Event Reason: To align with the protocol
8.8.7 (50)	Revised language Reason: To provide additional clarification
10.7.1 (53)	Deleted section 10.7.1 Reason: Section is repeated in the protocol
10.7.2 (53)	Moved Sample Size from 10.7.1 to 10.7 Reason: Formatting change
10.7.2 (53)	Changed from Bayesian approach to Frequentist approach Reason: To align with the revised SAP
11.2 (56)	Revised language Reason: To align with the protocol
11.2 (56)	Added language Reason: To provide additional clarification
11.3 (57)	Deleted language Reason: Section is repeated in the protocol
11.5 (57)	Deleted language Reason: To align with HIPAA
11.6 (58)	Deleted language Reason: Administrative change

Revision 12.7 changes

Section (Page)	Change and Reason for the Change
All Pages	Changed all pages to protocol version from 12.6 to version 12.7, dated March 30, 2020. Reason: Reflected revision date change in the header.
8.5.6 (41)	Updated sample size from 105 subjects to 95 subjects and added footnotes detailing the impact of existing sites that already reached the 15 subject enrollment cap. Reason: Updated due to COVID-19 considerations

1. INVESTIGATOR SIGNATURE PAGE

I agree to conduct the study as detailed in the protocol and in accordance with conditions of the approval imposed by the reviewing Ethics Committee or Institutional Review Board, all applicable regional laws and regulations. I agree to maintain adequate source documentation records throughout the clinical investigation and make them available as requested during monitoring visits. I agree to maintain the device accountability records and ensure that the investigational device is used solely by authorized users as specified by this protocol. In addition, I agree to provide all the information requested in the electronic Case Report Forms presented to me by the Sponsor in a manner to assure completeness and accuracy.

I agree to ensure that the requirements for obtaining informed consent are met. I agree to actively enroll consecutive subjects into this study and confirm that I am not currently participating in any clinical investigations for similar types of medical devices that would inhibit my ability to participate fully in this study. Additionally, I agree to disclose financial interests in accordance with 21 CFR 54, and certify that such financial interests, if any, will not interfere with my responsibilities as an investigator or influence study outcomes under my supervision.

I also agree that all information provided to me by the Sponsor, including pre-clinical data, protocols, Case Report Forms, and any verbal and written information, will be kept strictly confidential and confined to the clinical personnel involved in conducting the study. It is recognized that this information may be relayed in confidence to the Ethics Committee or Institutional Review Board, or to regulatory authorities.

In addition, no reports or information about the study or its progress will be provided to anyone not involved in the study other than the Sponsor, or its representatives, the Ethics Committee(s), the Institutional Review Board(s), or the core lab. Any such submission will indicate that the material is confidential.

Investigator Signature

Date


Investigator Printed Name

2. STUDY CONTACT PERSONNEL

2.1. SPONSOR



2.2.



3. PROTOCOL SYNOPSIS

Title:	Prospective, Multicenter, Single Arm Safety and Effectiveness Confirmatory Study of <u>EndoVascular</u> Abdominal Aortic <u>Aneurysm</u> Repair Using the Nellix® <u>System</u>
Short Title:	EVAS II Confirmatory Study
Study Devices:	Nellix EndoVascular Aneurysm Sealing System (Nellix System)
Study Sponsor:	Endologix, Inc.
Objectives:	To study the safety and effectiveness of the Nellix System for Endovascular Abdominal Aortic Aneurysm (AAA) repair. Procedures will be performed per the instructions for use, and per institutional protocols and standard of care for endovascular aneurysm repair. As such, this study will evaluate the safety and effectiveness of the device system among a wide range of physicians and in consecutively enrolled subjects to assess outcomes generalizability.
Study Design:	Prospective, multicenter, single arm study with consecutive, eligible subject enrollment at each site. All subjects will undergo the Endovascular Aneurysm repair procedure with the Nellix System. Sites have been chosen with a suitable research infrastructure and physician experience in endovascular aneurysm repair to ensure adequate enrollment. Subjects will be followed procedurally to discharge, at 30 days (primary safety endpoint), six months, one year (primary effectiveness endpoint) and annually thereafter to five years (total follow-up commitment).
Investigators:	Physicians (vascular surgeons or interventional cardiologists/radiologists as part of a multi-specialty team with vascular surgeons) with well-established experience in open repair and endovascular aneurysm repair techniques (i.e., ≥ 25 cases in the prior year) may participate. Details of the procedural and device usage techniques are incorporated into the study protocol on the basis of prior experience with an earlier generation device system, and more recent device refinements that have been developed. In those cases where the investigator is an interventionalist, a vascular surgeon must be immediately available during the procedure to perform any necessary surgical intervention.
Subject Population:	Patients diagnosed with an abdominal aortic or aortoiliac aneurysm who are considered candidates for endovascular repair and who meet the study eligibility criteria may be screened for enrollment in the study.
Enrollment:	A total of 95 subjects will be enrolled at a maximum of 38 US sites. There will be two (2) cohorts in the trial. There will be a Pivotal Cohort and a Roll-In Cohort. The Roll-In Cohort consists of the first enrolled subject at a site who has never used the Nellix System. A single site may not enroll more than 15% of the total enrollment*.

* Per the protocol revision 12.6, which required a total of 105 subjects, an enrollment cap of 15% limited a site to enroll no more than 15 subjects in order to guard against overrepresentation of higher enrollers. With the enrollment revision from 105 subjects to 95 subjects, this can lead to a slightly higher percentage than 15% (15 subjects is 15.8%) for any site that already reached the enrollment cap. As described in the corresponding Statistical Analysis Plan (SAP) maintained by the Endologix, this is not a substantial deviation from the original intent.

Training:

Nellix operators will be directly supervised in each Nellix case by an Endologix trainer or Proctor. This training period will continue until a minimum of one case has been successfully completed and each investigator is cleared to proceed with Study enrollment without proctor supervision. An Endologix Clinical Specialist will attend every case. This training requirement may be modified at the discretion of Endologix.

Inclusion Criteria:

1. Male or female at least 18 years old;
2. Informed consent understood and signed;
3. Patient agrees to all follow-up visits;
4. Have an infrarenal abdominal aortic aneurysm (AAA) with maximum sac diameter ≥ 5.0 cm, or ≥ 4.5 cm which has increased by ≥ 0.5 cm within the last 6 months, or which exceeds 1.5 times the transverse dimension of an adjacent non-aneurysmal aortic segment. No AAA < 4 cm will be included.
5. Anatomic eligibility for the Nellix System *per the instructions for use*:
 - a. Adequate iliac/femoral access compatible with the required delivery systems (diameter ≥ 6 mm);
 - b. Aneurysm blood lumen diameter ≤ 60 mm;
 - c. Proximal non-aneurysmal aortic neck: length ≥ 10 mm; diameter 18 to 28 mm; angle $\leq 60^\circ$ to the aneurysm sac;
 - d. Most caudal renal artery to each hypogastric artery length ≥ 100 mm;
 - e. Common iliac artery lumen diameter between 9 and 35 mm;
 - f. Distal iliac artery seal zone with length of ≥ 10 mm and diameter range of 9 to 25 mm;
 - g. Ability to preserve at least one hypogastric artery.
 - h. Ratio of maximum aortic aneurysm diameter to maximum aortic blood lumen diameter < 1.40

Exclusion Criteria:

1. Life expectancy < 2 years;
2. Psychiatric or other condition that may interfere with the study;
3. Participating in another clinical study
4. Known allergy to any device component;
5. Coagulopathy or uncontrolled bleeding disorder;
6. Ruptured, leaking or mycotic aneurysm;
7. Serum creatinine level > 2.0 mg/dL;
8. CVA or MI within three months of enrollment/treatment;
9. Aneurysmal disease of the descending thoracic aorta;

10. Clinically significant mural thrombus within the proximal landing zone (minimum 10mm) of the infrarenal non-aneurysmal neck (>5mm thickness over >50% circumference);
11. Connective tissue diseases (e.g., Marfan Syndrome)
12. Unsuitable vascular anatomy;
13. Pregnant (females of childbearing potential only).

Primary Endpoints:^{*}

Safety: Major adverse events at 30 days.

Effectiveness: Treatment Success at 2 year.

This is defined as procedural technical success and the absence of:

- Abdominal aortic aneurysm rupture;
- Conversion to open surgical repair;
- Endoleak Type I or Type III at Year 2
- Clinically significant migration;
- > 5mm aneurysm sac enlargement of the maximal diameter as measured by the difference from the one month CT time point; or
- Secondary endovascular procedure up to Year 2 for resolution of:
 - Endoleak (Type I or Type III)
 - Device obstruction or occlusion
 - Device migration
 - Abdominal aneurysm sac expansion
 - Device defect.

Additional Evaluations:

A summary analysis of the Effectiveness endpoint will also be presented at 1 year.

A descriptive analysis and a Kaplan-Meier analysis will be performed on All-Cause Mortality, Cardiovascular Mortality, and Adverse Cardiovascular Events (ACE).

Additional evaluations include:

- Procedural and in-hospital evaluations:
 - Anesthesia type;
 - Vascular access type;
 - Contrast media volume used;
 - Estimated blood loss;
 - Fluoroscopy time;
 - Catheter time;
 - Polymer fill volume;
 - Polymer fill pressure

^{*} Detailed definitions are provided in subsequent sections in the protocol.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

- Adjunctive device placement;
- Subjects requiring blood transfusion;
- Total procedure time;
- Anesthesia time;
- Concomitant procedures;
- Time in ICU;
- Time to hospital discharge.
- Mortality (all-cause and aneurysm-related) within 30 days, 31 to 365 days, and annually at years 1 to 5;
- MAE Individual Components within 30 days, 31 to 365 days, and at 1 year;
- Aneurysm Rupture within 30 days, 31 to 365 days, and annually at years 1 to 5;
- Conversion to open repair within 30 days, 31 to 365 days, and annually at years 1 to 5;
- Adverse Events: All serious and non-serious events within 30 days, 31 to 365 days, and annually at years 1 to 5;
- Renal function as assessed by estimated glomerular filtration rate (eGFR) pre-discharge and at 30 days, 6 months, and annually at years 1 to 5;
- Endoleak as reported by Core Lab at 30 days, 6 months, and annually at years 1 to 5;
- Device Patency and Integrity at 30 days, 6 months, and annually at years 1 to 5;
- Aneurysm Sac Diameter as reported by Core Lab at 30 days, 6 months, and annually at years 1 to 5;
- Migration (>5mm, >10mm) and clinically significant migration as reported by Core Lab at 30 days, 6 months, and annually at years 1 to 5;
- Secondary endovascular procedures (including those performed for resolution of endoleaks, device thrombosis/occlusion, rupture, migration, fracture, kinking, infection, aneurysm sac expansion, renal artery stenosis, and/or a device defect) within 30 days, 31 to 365 days, and annually at years 1 to 5.
- Gender Analysis for the primary safety and effectiveness endpoints.
- Site Comparisons Analysis for the primary safety and effectiveness endpoints.

Schedule of Tests: Pre-procedural high resolution, contrast-enhanced CT scan evaluation to determine anatomical eligibility for enrollment will be performed within three months of the study procedure. Following EC/IRB approval of the study and the written informed consent form, patients will be screened for eligibility. Following informed consent, a physical exam and laboratory testing will be performed prior to the procedure. Subjects will be followed procedurally and to hospital discharge, and will then be followed at intervals: 30 days; 6 months; 1 year; and annually to 5 years.

4. TRIAL OVERVIEW

4.1. OBJECTIVE

The objective of this study is to assess the safety and effectiveness of the Endologix Nellix® EndoVascular Aneurysm Sealing System for the endovascular repair of infrarenal abdominal aortic aneurysms (AAA). Procedures will be performed per the Nellix Instructions For Use (IFU) and per institutional protocols and standard of care for endovascular aneurysm repair. As such, this study will evaluate the safety and effectiveness of this device system among a wide range of physicians and in consecutively enrolled subjects to assess outcomes generalizability.

4.2. BACKGROUND

Mortality due to AAA rupture ranks 13th in the United States and 10th in Canada among men older than 65 years.* In addition, several authors have suggested an increase in the mortality rate due to rupture of AAA in the past decades in both the United States and England, with more than 14,000 and 7,259 deaths/year in those respective countries.†‡ Despite being significant, such figures are probably underestimated, because many deaths resulting from ruptured aneurysms are not verified by autopsy, and are consequently not documented.§ Some researchers have reported that the overall lethality associated with rupture of AAA approaches 80%, including dead on arrival subjects or those who die before the diagnosis is made.**

AAA is defined clinically as a focal dilatation of the aorta causing a diameter increase of >50% of the expected normal diameter.†† Although any artery may develop an aneurysm, they are most commonly observed in the infrarenal abdominal aorta, thoracic aorta, popliteal artery and common iliac artery.

The principal risks related to aneurysms are rupture and thrombus migration. Aneurysms slowly and continually increase in size leading to the aneurysm rupture. The larger an aneurysm becomes, the likelihood of eventual rupture increases. The natural history of aortic aneurysms is to enlarge and rupture. Other potential complications of the aneurysm include compression of adjacent organs which may result in aortoenteric fistula, or aortocaval fistula. If the thrombus embolizes and flows down the blood stream, this can induce acute or chronic arterial obliteration of the lower limbs.

The risk of rupture is weighed against the risk of perioperative morbidity. The United Kingdom Small Aneurysm trial (UKSAT) reported 103 aneurysm ruptures in 2,257 subjects over a period

*Semenciw R, Morrison H, Wigle D, et al. Recent trends in morbidity and mortality rates for abdominal aortic aneurysms. Rev Canadien Santé Publique 1992;83:274-6.

†Lelienfeld DE, Gurdenson PD, Sprafka JM, et al. Epidemiology of aortic aneurysms: Mortality trends in the United States 1951-1981. Atherosclerosis 1987;7:637-43.

‡Fowkes FRG, Macintyre CCA, Rucjerley CV. Increasing incidence of aortic aneurysms in England and Wales. Brit Med J 1989;298:33-5.

§Quil DS, Colgan MP, Summer DS. Ultrasonic screening for the detection of abdominal aortic aneurysms. Surg Clin North Am 1989;69:713-20.

**Bengtsson H, Bergqvist D. Ruptured abdominal aortic aneurysm: a population-based study. J Vasc Surg 1993; 18:74-80.

††Hallett JW Jr. Diseases of the aorta and its branches – aneurysms. Merck Manual, Online Medical Manual (2008). Retrieved on 19 March 2010 from <http://www.merck.com/mmpe/sec07/ch079/ch079b.html>.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

of seven years, with an annual rupture rate of 2.2%.^{*} The decision to treat a patient that presents with an asymptomatic aneurysm is primarily dependent upon the size of the aneurysm. Current Society for Vascular Surgery (SVS) practice guidelines recommend surveillance for most subjects with a fusiform AAA in the range of 4.0 to 5.4cm in maximum diameter; therefore, surgical repair of abdominal aneurysms of 5.5 cm or greater in diameter is recommended in healthy subjects, as is repair of saccular aneurysms. There are a number of qualifiers within the SVS guidelines that recommend AAA repair at 5.0cm, such as rapidly expanding aneurysms, and women.

It is estimated that approximately 25% to 40% of infrarenal AAA are not suitable for endovascular aneurysm repair (EVAR) due to unfavorable proximal neck anatomy (e.g., highly angulated, dilated, short, or encroaching on or involving the renal arteries).^{†‡} In most studies of endovascular AAA repair, the infrarenal non-aneurysmal neck length and angulation to the aneurysm sac requirements are ≥ 15 mm and $\leq 60^\circ$, respectively; shorter lengths or greater angulation have been reported to increase the risk of migration and type 1A endoleak and associated need for intervention.^{§,**}

When selecting the specific stent graft to be used for EVAR, the characteristics of the graft must be considered in light of the patient's anatomic and physiologic characteristics. Endovascular devices vary in the type of stent design. For example, most of the currently available devices seal aneurysm by proximal and distal fixation of the stent graft by either active fixation (anchoring pins) or oversizing the stent diameter for increased radial force thereby achieving seal and excluding the aneurysm sac lumen. Greater than 25% of subjects develop an endoleak (mostly type II endoleaks due to back-bleeding of lumbar or visceral arteries within the aneurysm sac) within the first two years following endoluminal stent-graft repair,^{††} and approximately 15-20% of the endoleaks persist at five years.[§] Endoleaks may lead to sac enlargement and migration and may require reintervention, using either catheter-based techniques or conversion to open repair. These secondary procedures increase the risk to the patient and increase the cost of treatment.^{‡‡} Accordingly, EVAR subjects are routinely monitored annually by contrast-enhanced CT scan after treatment contributing further to treatment costs and increasing patient exposure to nephrotoxic contrast agents and radiation.

^{*}Mortality results for randomized controlled trial of early elective surgery or ultrasonographic surveillance for small abdominal aortic aneurysms Lancet 1998; 352:1649-55.

[†]Carpenter JP, Baum RA, Barker CF, et al. Impact of exclusion criteria on patient selection for endovascular abdominal aortic aneurysm repair. J Vasc Surg 2001;34:1050-4.

[‡]Arko FA, Filis KA, Seidel SA, et al. How many patients with infrarenal aneurysms are candidates for endovascular repair? J Endovasc Ther 2004;11:33-40.

[§]Leurs LJ, Kievit J, Dagnelie PC, et al. Influence of infrarenal neck length on outcome of endovascular abdominal aortic aneurysm repair. J Endovasc Ther 2006;13:640-8.

^{**}AbuRahma A, Campbell J, Stone PA, et al. The correlation of aortic neck length to early and late outcomes in endovascular repair patients. J Vasc Surg 2009;50:738-48.

^{††}Stavropoulos SW and Charagundla SR. Imaging techniques for detection and management of endoleaks after endovascular aortic aneurysm repair. Radiology 2007; 243(3):641-55.

^{‡‡}Vogel TR, Symons RG, Flum DR. Longitudinal outcomes after endovascular repair of abdominal aortic aneurysms. J Vasc Endovasc Surg 2008;47(2):264-9.

Recent reports suggest contrast nephrotoxicity affects 7 to 12% of subjects after CT angiography.*

The Nellix System was designed to withstand the migration and lateral displacement forces acting on endografts, while at the same time excluding the aneurysm sac lumen and minimizing endoleaks of any kind. The system is comprised of two independent flow channels, one to each iliac artery. Each flow channel consists of a balloon-expanded ePTFE covered stent surrounded by a Polymer-filled EndoBag which fills the blood lumen within the aorta, thus providing positional stability of the endograft and sealing the aneurysm completely from side-branch flow.† A clinical trial has been performed to assess the safety and effectiveness of the first generation Nellix system for endovascular AAA repair and results serve as the basis for CE Mark approval (October 2012); initial results to one year have been reported.‡ Results of this trial demonstrate the versatility of the Nellix System in treating a variety of AAA anatomies, including those that have common iliac artery involvement bilaterally.

Continued clinical evaluation of the Nellix System in a broader group of institutions and physicians is therefore appropriate to assess the safety and effectiveness of the device for AAA repair and the generalizability of the approach.

4.3. STUDY DESIGN

This is a US multicenter, prospective, single arm clinical study. Subjects with infrarenal AAA who are suitable candidates for endovascular repair using the Nellix System, based on protocol Inclusion/Exclusion criteria, will be considered for enrollment.

After this protocol and the patient informed consent form are reviewed and approved by the local Ethics Committee/Institutional Review Board (EC/IRB), patients having infrarenal AAA will be offered participation in the study. This will be accomplished through the patient's reading of the informed consent form in the patient's native language and discussion of the study with the patient by the Principal Investigator (PI) and site personnel. Agreement to participate and to attend all follow-up visits will be documented with the patient's signature on the informed consent form, with appropriate signatures of the site PI and an impartial witness.

After providing written informed consent, screening and eligibility determinations will be performed by the site, independent Core Lab, Case Review Board and Endologix. Subjects will undergo a high resolution, contrast-enhanced computed tomography angiography (CT) scan of the relevant aortic and aortoiliac vasculature within the prior three months of the scheduled procedure. Evaluation of the aortic and vascular anatomy suitability per this protocol, as depicted on the CT scan, will be performed by the site PI, reviewed by an independent core laboratory, and will undergo Case Review Board. Other tests include a physical examination, review of medical history for exclusionary conditions, and selected blood laboratory analyses.

*Chaer RA, Gushchin A, Rhee R, et al. Duplex ultrasound as the sole long-term surveillance method post-endovascular aneurysm repair: a safe alternative for stable aneurysms. *J Vasc Surg* 2009; 49:845-9.

†Donayre CE, Zarins CK, Krievins DK, et al. Initial clinical experience with a sac anchoring endoprostheses for abdominal aortic aneurysm repair. *J Vasc Surg* 2011;53:574-82.

‡Krievins DK, Holden A, Savlovskis J, et al. EVAR using the Nellix sac anchoring endoprostheses: treatment of favorable and adverse anatomy. *Eur J Vasc Endovasc Surg* 2011;42:38-46.

Endologix will notify, in writing, each potential patients final eligibility status before a case may be scheduled.

Following discharge from the hospital, the first follow-up visit will be made at 30 days (± 2 weeks). A CT scan will be performed to assess aneurysm morphology and device integrity and patency, as well as the status of the implanted devices. Subsequent follow-up visits will be made at six months, one year, and annually to five years per institutional standard of care for subjects with endovascular stent grafts. Continued subject follow-up beyond five years is outside of the scope of this study. Nonetheless, all subjects should be monitored and evaluated per the institutional standards of care for patients with an implanted endovascular stent graft.

5. STUDY POPULATION

5.1. ENROLLMENT

A total of 95 subjects will be enrolled at a maximum of 38 US sites. There will be two (2) cohorts in the trial. There will be a Pivotal Cohort and a Roll-In Cohort. The Roll-In Cohort consists of the first enrolled subject at a site who has never used the Nellix System. A single site may not enroll more than 15% of the total enrollment*. Following investigator training, each new infrarenal Principal Investigator (PI) will commence patient screening and enrollment under approval per applicable national and local clinical trial requirements, and complete a minimum of one (1) successful Nellix case (Roll-in) in their training phase. This training period may be modified at the discretion of Endologix. Physicians will be directly supervised in each case by designated Endologix staff.

Total Enrollment: Up to 95 subjects will be enrolled at a maximum of 38 US sites.

Pivotal Cohort: If a site that has used a Nellix device, the first and subsequent subjects will be in the Pivotal Cohort.

Roll-In Cohort: One (1) roll-in subject will be enrolled at each new site (those sites/PIs that have never used a Nellix device) to allow for new infrarenal Investigator(s) to receive additional training on the Nellix System. Any roll-in subjects will be screened, consented, treated and followed identically to the pivotal cohort. However, the roll-in cohort of this study may be evaluated separately.

5.2. INCLUSION CRITERIA

A patient who meets ***all of the following criteria*** potentially ***may be included*** in the study:

1. Male or female at least 18 years old;
2. Informed consent form understood and signed
3. Patient agrees to all follow-up visits;

* Per the protocol revision 12.6, which required a total of 105 subjects, an enrollment cap of 15% limited a site to enroll no more than 15 subjects in order to guard against overrepresentation of higher enrollers. With the enrollment revision from 105 subjects to 95 subjects, this can lead to a slightly higher percentage than 15% (15 subjects is 15.8%) for any site that already reached the enrollment cap. As described in the corresponding Statistical Analysis Plan (SAP) maintained by Endologix, this is not a substantial deviation from the original intent.

4. Abdominal aortic aneurysm with sac diameter ≥ 5.0 cm, or ≥ 4.5 cm which has increased by ≥ 0.5 cm within the last 6 months, or which exceeds 1.5 times the transverse dimension of an adjacent non-aneurysmal aortic segment. No AAA < 4 cm will be included.
5. Anatomically eligible for the Nellix System (per Instructions For Use):
 - a. Adequate iliac/femoral access compatible with the required delivery systems (diameter ≥ 6 mm);
 - b. Aneurysm blood lumen diameter ≤ 60 mm;
 - c. Proximal non-aneurysmal aortic neck: length ≥ 10 mm; diameter 18 to 28mm; angle $\leq 60^\circ$ to the aneurysm sac;
 - d. Most caudal renal artery to each hypogastric artery length ≥ 100 mm;
 - e. Common iliac artery lumen diameter between 9 and 35mm;
 - f. Distal iliac artery seal zone with length of ≥ 10 mm and diameter range of 9 to 25mm;
 - g. Ability to preserve at least one hypogastric artery.
 - h. Ratio of maximum aortic aneurysm diameter to maximum aortic blood lumen diameter < 1.40

5.3. EXCLUSION CRITERIA

A patient who meets *none of the following* criteria potentially *may be included* in the study:

1. Life expectancy < 2 years as judged by the Investigator;
2. Psychiatric or other condition that may interfere with the study;
3. Participating in another clinical study;
4. Known allergy or contraindication to any device material;
5. Coagulopathy or uncontrolled bleeding disorder;
6. Ruptured, leaking or mycotic aneurysm;
7. Serum creatinine (S-Cr) level > 2.0 mg/dL;†
8. CVA or MI within three months of enrollment/treatment;
9. Aneurysmal disease of the descending thoracic aorta;
10. Clinically significant mural thrombus within the proximal landing zone (minimum 10mm) of the infrarenal non-aneurysmal neck (> 5 mm thickness over $> 50\%$ circumference);
11. Connective tissue diseases (e.g., Marfan Syndrome);
12. Unsuitable vascular anatomy that may interfere with device introduction or deployment;
13. Pregnant (female of childbearing potential only).

6. RESPONSE MEASURES

6.1. PRIMARY SAFETY

The safety endpoint is defined as the incidence of Major Adverse Events (MAE) at 30 days, defined as the composite of the following. Event definitions are provided in §8.8.7.

- *All-Cause Mortality;*

†This criterion does not apply to patients on dialysis prior to study screening/enrollment.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

- *Bowel Ischemia;*
- *Myocardial Infarction;*
- *Paraplegia;*
- *Renal Failure;*
- *Respiratory Failure;*
- *Stroke;*
- *Procedural Blood Loss $\geq 1,000cc$*

Refer to the Statistical Analysis Plan for primary safety endpoint analysis details.

6.2. PRIMARY EFFECTIVENESS

The primary effectiveness endpoint is defined as the rate of Treatment Success at year 2. Treatment Success is a composite of outcomes clinically relevant to the endovascular repair of infrarenal AAA as follows. Event and related definitions are provided in §8.8.7.

Treatment Success: Procedural technical success [¥] and absence of:

- Abdominal aortic aneurysm rupture;
- Conversion to open surgical repair;
- Endoleak Type I or III at year 2;
- Clinically significant migration[‡];
- $> 5\text{mm}$ aneurysm sac enlargement of the maximal diameter as measured by the difference from the one month CT time point Secondary endovascular procedure up to year 2 for resolution of
 - Endoleak (Type I or Type III)
 - Device obstruction or occlusion
 - Device migration
 - Abdominal aneurysm sac expansion
 - Device defect.

A summary analysis of this endpoint will also be presented at 1 year. Refer to the Statistical Analysis Plan for primary effectiveness endpoint analysis details.

[¥] Procedural technical success is deployment of the Nellix System in the planned location and without unintentional coverage of both internal iliac arteries or any visceral aortic branches and with the removal of the delivery system.

[‡]Refer to protocol §8.8.7 for detailed definitions.

6.3. ADDITIONAL EVALUATIONS

As outlined in the SAP, additional evaluations include:

Secondary Mortality and Adverse Cardiovascular Event Endpoint, a Kaplan-Meier analysis and descriptive analysis within 30 days (early), 31 to 365 days (late), and annually at years 1 through 5.

Procedural and In-Hospital Evaluations:

- Anesthesia type;
- Vascular access type;
- Volume of contrast media used;
- Estimated blood loss;
- Fluoroscopy time;
- Catheter time;
- Polymer fill volume;
- Polymer fill pressure;
- Adjunctive device placement
- Subjects requiring blood transfusion
- Total procedure time^{*}
- Anesthesia time;[†]
- Concomitant procedures;
- Time in ICU;[‡]
- Time to hospital discharge;^{*}

Mortality, all-cause and aneurysm-related, within 30 days, 31 to 365 days, and annually to 5 years;

MAE Individual Components within 30 days, 31 to 365 days, and annually to 5 years;

Aneurysm Rupture within 30 days, 31 to 365 days, and annually to 5 years;

Conversion to Open Repair within 30 days, 31 to 365 days, and annually to 5 years;

Adverse Events (serious and non-serious) within 30 days, 31 to 365 days, and annually to 5 years;

Renal Function pre-discharge and at 30 days, 6 months, and annually to 5 years, as assessed by the estimated glomerular filtration rate (eGFR) and changes over time;

Device Patency and Integrity within 30 days, at 6 months, and annually to 5 years.

Device Performance (aneurysm sac diameter change from the first post-operative visit; device migration; clinically significant device migration, incidence of endoleak) at 30 days, 6 months, and annually to 5 years;

Secondary Procedure within 30 days, 31 to 365 days, and annually to 5 years for resolution of endoleak, device occlusion, migration, aneurysm sac expansion and/or a device defect.

Gender Analysis of the primary safety and effectiveness endpoints.

Site Comparison Analysis of the primary safety and effectiveness endpoints.

^{*} Elapsed time from the first break of skin to final closure (i.e., skin to skin time).

[†] Elapsed time from the initiation to the end of the anesthesia protocol.

[‡] Elapsed time from the first administration of anesthesia to release from the ICU or post-anesthesia care unit providing ICU-level care. If the patient is not admitted to the ICU, this is defined as 0 hours.

7. STUDY MATERIALS



The Nellix System is comprised of five components as further described below:

1. Nellix Catheters (single use);
2. Nellix Accessory kit (single use);
3. Nellix Dispenser (single-use or reusable);
4. Nellix Polymer (single-use).
5. Ovation iX Iliac Stent Grafts (single use, as needed)

7.1. DEVICE DESCRIPTION


Implanted Devices. The Nellix implant consists of two devices, each advanced from the respective femoral artery via a catheter-based delivery system over a standard .035" guidewire. Each implanted device has the following main components: an expanded polytetrafluoroethylene (ePTFE) covered balloon expandable cobalt chromium alloy stent surrounded by a polyurethane EndoBag with welded seams and a lumen having an inner polyester (PET) sleeve is attached to the stent proximally and distally using polyethylene sutures, and a biostable Polymer formulation which is introduced through the catheter system to the EndoBag and cures in situ.

An illustration of the implant procedure is shown below. Images of the stent (with and without ePTFE graft), EndoBag, and the final implant in a silicone model are shown subsequently.



* Elapsed time from initiation of the procedure to physical discharge from the hospital.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.



The mechanism of action is as follows: The ePTFE covered stents form lumens through which blood flows to the distal limbs and vasculature. The EndoBags with cured polymer exclude the aneurysm from blood flow, preventing aneurysm pressurization and rupture. This seals the entire anatomy from the infrarenal segment to the origin of the hypogastric arteries.

In the case of distal extension, the Ovation iX Iliac Stent Grafts can be used. These iliac limbs and extensions are comprised of a nitinol stent encapsulated in low-permeability PTFE. The iliac limbs are deployed into the distal end of the Nellix stent.

Delivery Systems. The 17Fr profile (OD) delivery systems are sterile, single use, catheter-based systems used to deploy the stents with EndoBags and are compatible with standard 0.035-inch guidewires. All delivery systems are of the same design regardless of stent/EndoBag model.

Three other single use components complete the Nellix System:

- Sterile Polymer dual chamber cartridges requiring storage at or below -20°C until use. Immediately prior to use (1-2 hours), the required Polymer cartridge(s) are thawed per the Instructions For Use (IFU).
 - Sterile accessory kit including a console with dual sets of tubing and quick connects for simultaneous attachment to two delivery catheters; two mixers, and a Centurion Compass™ GP Pressure Transducer.[†] The mixer attaches to the end of the Polymer cartridge and mixes the two chamber solutions during transfer into the EndoBags. A second mixer is provided as a back-up if needed.
- Polymer dispenser

The pre-operative high resolution, contrast enhanced CT scan serves to determine anatomical measures for subject eligibility and device initial sizing determinations. Arteriography with the aid of a marker catheter bilaterally serves to verify the specific device length needed on each

[†]The pressure monitoring device is marketed by Centurion Medical Products, Inc. (Williamston, MI, USA) and is 510(k) cleared through K112203.

side to ensure full coverage of the Nellix device from the infrarenal segment to the supra-hypogastric segments.

After selected delivery system is introduced through femoral arteriotomy under fluoroscopic guidance, each is advanced and positioned across the aneurysm sac at the desired location to place the stents below the most caudal renal artery and aligned at their proximal edge. The outer sheath of each delivery system is retracted to expose each constrained stent/EndoBag. (Note: A tether wire attaches each EndoBag to the catheter. The catheter can be removed only after disengaging the tether wire. This is a safety mechanism designed to hold the implant in place during deployment.)

At this time, the console is attached to both delivery systems using the mating, color-coded quick connects as depicted in the image below. White connectors denote the EndoBag evacuation and Polymer fill line; black connectors denote the stent balloon expansion line; luer connectors denote the angiographic media injection line (exiting the ports in the delivery system tip). The .035" guidewire lumens are on the outermost ends of the delivery catheter handles.



EndoBag evacuation, balloon expansion, Polymer introduction, and angiographic media injection are all performed through the console:

- A vacuum is applied via console port 1 with the switch in the 'open' position to simultaneously evacuate air from both EndoBags. After evacuation, the switch is placed in the 'closed' position.
- An inflation device is attached to control console port 2 to simultaneously balloon expand both ePTFE covered stents with contrast enhanced saline per the IFU. The stents expand from each end toward the center. The balloons are then simultaneously deflated per standard technique.
- The EndoBags are then filled by dispensing Polymer through console port 3 with the pressure transducer placed between the mixer and the port 3 luer. Polymer filling of the EndoBags occurs simultaneously, and is to continue until the pressure transducer displays a reading of 180mmHg. In the case of a patient with systolic blood pressure exceeding 180mmHg, Polymer filling continues to a pressure that is 20mmHg above the patient's systolic pressure. Once Polymer is completely injected into the EndoBags, the physician will allow the polymer to cure for approximately 3 to 5 minutes. Polymer curing is visually

observable through the window on console port 3, which will change in color from red to white when cured.

A final angiographic run is done to confirm the complete seal of the aneurysm sac and absence of endoleaks. If deemed appropriate by the physician, a secondary fill may be performed to add additional Polymer to each EndoBag independently. The delivery systems are then detached from the implant by releasing the tether wires, and are removed from the patient and discarded per hospital standard practice.

7.2. INVESTIGATOR TRAINING AND EXPERIENCE

Investigators will be directly supervised in each case by an Endologix trainer or a certified Proctor until a minimum of one Nellix case have been successfully completed. One Principal Investigator (PI) at each participating site will be responsible for supervision of study conduct. He/she and any authorized sub-investigator must satisfy the following criteria prior to the enrollment of their first patient in the confirmational trial.

1. Complete review of prior clinical safety and effectiveness information regarding the earlier generation device (3SQ), as well as design qualification testing regarding the Nellix System under study, and the scientific literature review.
2. Have prior training and experience in the open surgical or endovascular repair of AAA (approximately ≥ 25 cases as an institutional team, in the prior year).
3. Undergo didactic training on the device anatomy and design, Instructions for Use (IFU) including patient selection criteria, contraindications, warnings, precautions, and proper device use, and troubleshooting methods.
4. Practice using the Nellix System in a simulated use bench top model. Completion of each criterion will be documented for each PI and site *prior* to performance of the first clinical procedure at a given site under this protocol. Sub-Investigators will also be required to meet training requirements. This may be completed at a later date per Endologix discretion.

For procedures lead by a sub-investigator, the PI is expected to be present to provide proper supervision of the case, unless the sub-Investigator has been through investigational device training and approval by Endologix.

For those physicians who have not had commercial access to or experience with Ovation, Endologix will require they undergo standard Ovation device and procedure training prior to use.

7.3. DEVICE ACCOUNTABILITY

Usage of the Nellix catheter, polymer, and polymer dispenser will be documented in the electronic Case Report Forms (eCRF) and within an Investigational Device Accountability Log.

In the case of a device malfunction prior to implantation, the information will be noted in the eCRFs and all device components must be returned to Endologix. Should the device be removed from the body, the device must be returned to Endologix for evaluation per the IFU. Endologix must be informed if there is a circumstance that precludes return of any investigational product/components. Usage of the devices will be documented in the eCRFs and on the site's inventory log.

7.4. PATIENT AND DEVICE PREPARATION

All procedures must be performed in an operating room, or in an endovascular suite having full imaging capabilities, vascular surgery and anesthesia services. The PI will refer to institutional protocols relating to anesthesia and monitoring of vital signs. The appropriate devices will be selected in accordance with the patient eligibility determination. The devices will be prepared and handled in accordance with the IFU, which is provided in *Attachment 1*.[†] Other ancillary devices will be selected by the physician per institutional standards.

8. STUDY METHODS

8.1. GENERAL ENTRY PROCEDURES

Prospective subjects as defined by the criteria in §[5.2](#) and [5.3](#) will be considered for entry into this study.

All site personnel involved in patient screening and data collection procedures will be trained on this protocol and the electronic remote data capture (eDC) system. Following patient consent, inclusion/exclusion criteria information will be entered into the eDC system. This action will notify Endologix that a patient has been consented and is being considered for enrollment. In parallel, the pre-operative CT scan will be uploaded to Core Lab by the site. The Core Lab will subsequently provide the anatomical measurements for confirming eligibility. Endologix will then facilitate a Case Review Board to examine anatomical and medical information for final sign off. Sites will be notified of individual study subject eligibility via the eDC system.

8.2. INFORMED CONSENT

Written informed consent, in accordance with ICH E6 (as applicable), 21 CFR 50 (Informed Consent of Human Subjects), applicable Good Clinical Practice standards, and study center regulations, shall be obtained from each patient, or from their legal representative, prior to the study procedures. The PI will retain a copy of the signed informed consent document in each patient's record, and provide a copy to the patient.

The PI must not request the written informed consent of any patient, and must not allow any potential patient to participate in the investigation before obtaining governing EC/IRB approval.

Prior to starting the study, the PI will provide Endologix with a copy of the sample Informed Consent document approved by the EC/IRB with documented evidence that the EC/IRB has approved the protocol.

8.3. SUBJECT IDENTIFICATION

Upon satisfactory completion of all site start-up training and documentation requirements, Endologix will supply the site with written 'Go' notification that initiation of subject screening/consenting for study enrollment is authorized.

Subject numbers will be automatically assigned by the eDC system.

[†]Ancillary devices recommended for use during the procedure are listed in the IFU.

Confidential. This document and the information contained herein may not be reproduced, used or disclosed without written permission from Endologix, Inc.

A Screening Log will be maintained by the site, either on paper or electronically in the eDC, to document each subject who undergoes the screening process and the determination of eligibility.

The actual procedure date will serve as the “start” date from which follow-up evaluations will be measured. After treatment, each patient will be evaluated prior to hospital discharge and will then be evaluated at one month (defined as 30 ± 14 days), at six months (defined as 180 ± 30 days), and at annual follow-ups at 1 year (defined as 365 ± 60 days), and at 2 to 5 years (± 90 days).

8.4. SCHEDULE OF MEASUREMENTS

A summary of the tests and measurements to be conducted pre-study/at baseline, operatively, prior to discharge, and during follow-up is illustrated in the following chart:

Schedule of Tests:	Screening/ Baseline	Procedure (Day 0)	Pre- Discharge	1 Mo. ± 14 days	6 Mo. ± 30 days	1 Yr. ± 60 days	Every 6 months (Enhanced Follow-up) ^f	2 to 5 Yrs. ± 90 days
Inclusion/Exclusion	x							
Demographics/Medical History	x							
Physical Exam [†]	x		x	x	x	x	x	x
Blood Laboratory Analysis [‡]	x		x	x	x	x		x
Contrast-enhanced CT scan [§]	x			x	x	x	x	x
Procedural Information		x						
Adverse Events		x	x	x	x	x	x	x

[†]The physical exam includes overall health, physical assessment, and vital signs.

[‡]Only serum creatinine will be collected.

[§]The baseline high resolution, contrast-enhanced CT scan performed within the prior three months will be performed to determine anatomical eligibility.

^fSubjects who experience migration, AAA sac enlargement and/or a Type 1A or 1B endoleak or who are at risk of experiencing any of these events (e.g., those subjects who would not meet the new inclusion/exclusion criteria) are recommended to be followed every 6 months

8.5. STUDY PROCEDURES AND EVALUATIONS

8.5.1. Informed Consent

The patient, or his/her legal representative, is to be informed about the study and provide written consent. Confirmation of written consent will be entered in the eCRF by the site for verification by the Sponsor prior to enrollment eligibility determination. Please see *Attachment 2* for the informed consent template and *Attachment 3* for the study eCRFs.

8.5.2. CT Scan Protocol

Patient enrollment into this study is based on the site evaluation of patient conformance with the protocol-specified selection criteria §5.2 and 5.3, including anatomical requirements for the device based on the high resolution (slice thickness *minimum* ≤ 3 mm; axial, coronary, sagittal views), contrast-enhanced CT scans*. To ensure consistency, below are the requirements for CT acquisition:

- Only high resolution, contrast-enhanced spiral CT scans are acceptable.
- Data must be uncompressed.

- **Preferred maximum slice spacing is 2mm. In no case should it exceed 3mm.**
- The preferred protocol, shown below, is easier to attain with a multi-row scanner. If the preferred protocol cannot be used, an alternate protocol is provided.
- Instruct patient not to move during scan. Do not move table height, position, or field of view during scan. If such movement occurs, repeat scan in its entirety.

Parameter	Preferred	Alternate
Scan Mode	Helical/Spiral	
Scan Parameters	140kVp, Auto mA, 0.5sec	140kVp, 280mA (min), 1.0sec
Collimation	0.625 to 2mm	3mm
Slice Spacing	0.625 to 2mm	3mm
Superior Extent	Superior to the celiac artery origin	
Inferior Extent	Lesser trochanter of femur	
Patient Instruction	Single breath hold	1 st hold: above celiac to bifurcation 2 nd hold: bifurcation to lesser trochanter
Contrast*	Standard non-ionic	
Volume and Rate	150mL at 3 to 4 mL/sec or as per institutional standards	
Scan Delay	ROI - threshold 90Hu in aorta	
Field of View	Large body	
Window Level	400/40	

*For those subjects with renal insufficiency, a non-contrast CT and duplex ultrasound will be acceptable.

8.5.3. Pre-Operative Activities

Following informed consent, the **SCREENING AND BASELINE** eCRFs are to be completed. The PI is to proceed with procedure scheduling **only** after receiving acknowledgement from Endologix that the patient meets the enrollment criteria. This determination is based on the documented responses and data provided by the PI, the Core Lab CT assessment and the Case Review Board.

8.5.4. Subject Enrollment

Enrollment in the study only occurs upon advancement of the study device into the subject's vasculature. Therefore, the date of study enrollment is the date of the Nellix System implant, as documented in the Visit Date field in the Index Procedure eCRF. If the device does not contact the patient, the patient will be considered as screen failure and the assigned study number will not be given to another patient. If, during the index procedure, the device is introduced but is not implanted and/or there is conversion to open surgery, the patient will be followed until the 30 day follow-up.

8.5.5. Post-Operative Follow-up Visits

The eCRFs are to be entered for each subject at the post-operative follow-up visits. These visits should be scheduled post-procedurally at **30±14 days** (one month visit),

180±30 days (six month visit), **365±60 days** (one year visit), **730±90 days** (two year visit), **1095±90 days** (three year visit), **1460±90 days** (four year visit), and **1825±90 days** (five year visit).

8.5.6. Enhanced Post-Operative Follow-up Visits

Subjects who experience migration, AAA sac enlargement and/or a Type 1A or 1B endoleak during follow-up are recommended to be followed every 6 months with a contrast enhanced CT as part of the follow-up. For those subjects with renal insufficiency, a non-contrast CT and duplex ultrasound will be acceptable. Please reference the **Schedule of Measurements** in section 8.4 for a summary of tests and measurements to be performed at the enhanced follow-up visit.

All Investigators are trained to best practices and the IFU prior to the index procedure. If any inadequate procedural issues are observed that could result in “at risk” subjects (as observed on the first available post-operative CT), enhanced follow-up visits are recommended.

An inadequate procedure is defined as not meeting any one of the following:

1. (CoreLab Proximal Seal – 4 mm) \geq 10 mm
2. Acquired distal seal \geq 10 mm
3. Lowest stent within 10 mm of the lowest renal artery

Instructions for eCRF completion and the study-required evaluations for each visit will be provided at the site initiation visit.

8.6. RECOMMENDED SECONDARY INTERVENTION TREATMENT OPTIONS

Endologix submitted an Unanticipated Adverse Device Effect (UADE) report to the FDA on October 14, 2016 to report higher than anticipated rates of migration, endoleak (Type IA), and AAA sac enlargement, based on the 2-year follow-up data from the Nellix US IDE trial. As part of the investigation of the higher than anticipated rates of Type 1A endoleaks, migration and AAA sac enlargement, Endologix, in consultation with physicians, reviewed alternative treatment options from clinical studies, commercial use, literature, physician experience and use in similar technologies (e.g., EVAR) to develop a list of recommended treatment options. For many of the treatment options, the clinical data to support the safety and effectiveness of the treatment is limited.

The treating physician should consider multiple factors such as subject anatomy, subject risk profile, hospital standard of care and physician preference when determining the appropriate treatment option. Endologix’s treatment recommendation is based on a subset of those factors, and as such, should be used as a reference for physicians and not a requirement. The recommended treatment options are provided below.

It should be noted conversion to open repair should be used if a less invasive option is not feasible for the subject, based on the physician’s clinical judgment. Even with the other recommended treatment options available, conversion to open repair can be performed at any time, per the physician’s discretion.

8.6.1. Treatment Options for Type 1 Endoleaks

8.6.1.1. Type 1A Endoleaks

The recommendation for treatment of Type 1A endoleaks is based on evaluation of the implant position and proximal seal zone. The Nellix implant is correctly aligned if the bottom of the first stent cell element is aligned with the distal origin of the most caudal renal artery and the second stent not more than one stent strut higher on the opposite side. The Nellix implant has an adequate proximal seal zone if at least 10mm of the Nellix Endobags are in contact with healthy (i.e. minimal tortuosity, occlusive disease, and/or calcification) aortic proximal neck vessel wall. The length of the aortic proximal seal is measured from the lower margin of the most caudal renal artery to the point distally where the lumen diameter change is 10%. Based on evaluation of these factors, two potential treatment options for the treatment of Type 1A endoleaks have been identified:

- Coils (see Appendix A on **page 60** for list of recommended coils) and liquid embolics (Onyx®) with or without proximal covered stent extenders (balloon-expandable or self-expanding)
- Nellix proximal extenders with or without commercially available stent grafts in the visceral arteries (NiNA with polymer/ Secondary ChEVAS)
 - Covered stents are placed in the visceral arteries to allow blood flow and the Nellix implant is placed in a sufficient length of healthy supra-renal aortic neck to enable aneurysm sealing.

The table below summarizes these options, aligning each with evaluation criteria:

Implant and Anatomic Evaluation Criteria	Potential Treatment Options for Type IA Endoleaks
Implant positioned correctly with adequate sealing zone	<ul style="list-style-type: none">- Coils* and liquid embolics (Onyx®)-
Implant positioned incorrectly with adequate sealing zone	<ul style="list-style-type: none">- Coils* and liquid embolics (Onyx) with proximal covered stent extenders (iCAST)
Implant positioned correctly with inadequate sealing zone	<ul style="list-style-type: none">- Nellix Proximal Extenders with commercially available stent grafts in the visceral arteries (iCAST) (Secondary ChEVAS)
Implant positioned incorrectly with inadequate sealing zone	<ul style="list-style-type: none">- Nellix Proximal Extenders (NiNA with Polymer)- Nellix Proximal Extenders with commercially available stent grafts in the visceral arteries (iCAST) (Secondary ChEVAS)

* See Appendix A on **page 60** for a list of recommended coils

8.6.1.2. Type 1B Endoleaks

A potential treatment option for Type 1B endoleaks have been identified, which can be used based on the physician's preference:

- Distal Covered Stents

The table below summarizes this options, aligning each with evaluation criteria:

Implant and Anatomic Evaluation Criteria	Potential Treatment Options for Type IB Endoleaks
--	---

Inadequate sealing in between the Nellix Endobags and the iliac vessel wall	- Distal Covered Stent Extenders (Ovation iX Iliac Stent Graft)
---	---

8.6.2. Treatment Options for Migration

The recommendation for treatment of migration is based on evaluation of the distance the implant has migrated, presence of endoleak and comparison between proximal and distal migration distances. Based on evaluation of these factors, three potential treatment options for the treatment of migration have been identified:

- CT Surveillance
- Nellix proximal extenders with or without commercially available stent grafts in the visceral arteries (NiNA with polymer/ Secondary ChEVAS)
- Nellix Stent Relining (NiNA without Polymer)
 - For subjects in which the proximal migration is more significant than the distal migration of the Nellix implant, stent relining should prevent or slow the rate of migration. Relining the stent will increase the stiffness and column strength of the Nellix implant which will increase its resistance to the hemodynamic forces acting upon it.

The table below summarizes these options, aligning each with evaluation criteria:

Implant and Anatomic Evaluation Criteria	Potential Treatment Options for Migration
Nellix implant has migrated <10mm, in the absence of Type 1A endoleak and a proximal seal zone of at least 10mm remains	- CT Surveillance (Enhanced Follow-up)
Nellix implant has migrated ≥10mm, in the absence of Type 1A endoleak and a proximal seal zone of at least 10mm remains	- Nellix Relining (NiNA without Polymer)
Implant migration associated with Type 1A endoleak and a proximal neck of at least 30mm is available	- Nellix Proximal Extenders (NiNA with Polymer)
Implant migration associated with Type 1A endoleak and a proximal neck of less than 30mm is available	- Nellix Proximal Extenders with commercially available stent grafts (iCAST) in the visceral arteries (Secondary ChEVAS)

8.6.3. Treatment Options for Aneurysm Expansion

The recommendation for treatment of aneurysm expansion is based on the presence of Type 1A/B endoleaks. Based on evaluation of these factors, two potential treatment options for the treatment of aneurysm expansion have been identified:

- Type 1A Endoleak Treatments
- Distal Extension (Ovation iX Iliac Stent Graft)

The table below summarizes these options, aligning each with evaluation criteria:

Implant and Anatomic Evaluation Criteria	Potential Treatment Options for Aneurysm Expansion
--	--

Aneurysm expansion is associated with Type 1A endoleak	Refer to treatment option for Type IA endoleak for further detail
Aneurysm expansion is associated with Type 1B endoleak	- Distal Covered Stent Extenders (Ovation iX Iliac Stent Graft)
Aneurysm expansion in the absence of endoleak	- Distal Covered Stent Extenders (Ovation iX Iliac Stent Graft)

8.7. PROTOCOL DEVIATIONS

A protocol deviation occurs when a clinical investigator and/or study site personnel do not conduct the study according to the clinical investigational plan. All deviations are recorded on a Protocol Deviation Form. United States regulations (21 CFR 812.140) require that investigators maintain accurate, complete, and current records relating to the clinical study. This includes documents showing the dates and reasons for each deviation from the clinical investigational plan. Depending upon the nature of the protocol deviation, expedited reporting and prior approval from Endologix may be required. All deviations will be summarized and submitted in IDE progress reports, annual reports, and the final study report to FDA.

If Endologix finds that an investigator is not complying with the executed study agreements, the investigational plan, FDA regulations, or the requirements of the reviewing IRB, prompt action will be taken to secure compliance. In addition, shipment of the device may be stopped or the participation of the investigator may be terminated.

8.7.1. Deviations with Expedited Reporting Requirements

For the following types of protocol deviations (per 21 CFR 812.150), an investigator is required to notify Endologix and the IRB within 5 business days of the deviation.

- Emergency Deviation from the Investigational Plan (a deviation to protect the life or physical well-being of a subject in an emergency).
- Failure to obtain Informed Consent

Notification to Endologix and/or the IRB should be documented and maintained in the clinical study file at the site and at Endologix.

8.7.2. Deviations Requiring Prior Approval

An investigator is required to obtain prior approval from clinical study management at Endologix and the IRB before initiating deviations from the Investigational Plan that affect the scientific soundness of the plan, or the rights, safety, and welfare of the subjects (non-emergent situation). However, prior approval is not required in situations where unforeseen circumstances are beyond the investigator's control, e.g., subject did not attend scheduled follow-up visit, laboratory test was performed incorrectly, and test equipment did not operate properly.

8.7.3. Non-Urgent Deviations

Protocol deviations which do not have the urgency associated with expedited notification or prior Endologix/ IRB approval (as discussed in the above paragraphs) will be reported upon discovery, such as during completion of eCRFs or a monitoring visit.

8.8. ADVERSE EVENT REPORTING

Adverse events will be entered on the applicable eCRF and in the patient's medical records. The date of onset, severity, and action taken are requested to be identified by the PI.

8.8.1. General Definitions

An adverse event (AE) is any untoward medical occurrence, unintended disease, injury or untoward clinical sign (including an abnormal laboratory finding) in subjects, users or other persons, whether or not related to the investigational medical device (ISO 14155).

A serious adverse event (SAE) is an adverse event that:

- Led to a death.
- Led to a serious deterioration in the health of the subject that:
 - a. Resulted in life threatening illness or injury.
 - b. Resulted in a permanent impairment of a body structure or a body function.
 - c. Required in-patient hospitalization or prolongation of existing hospitalization.
- Resulted in medical or surgical endovascular procedure to prevent permanent impairment to a body structure or a body function.
- Led to fetal distress, fetal death, or a congenital abnormality or birth defect.

An unanticipated adverse device effect (UADE) is any serious adverse effect on health or safety or any life threatening problem or death caused by or associated with a device, if that effect, problem, or death was not previously identified in nature, severity, degree of incidence. Additionally, an unanticipated adverse device effect includes any other unanticipated serious problem associated with a device that relates to the rights, safety, or welfare of subjects (21 CFR 812.3 (s)).

Severity of an AE is a clinical determination of the event intensity. The severity assessment for a clinical AE should be completed using the following definitions as guidelines:

Mild (+1): Awareness of sign or symptom, but easily tolerated.

Moderate (+2): Discomfort enough to cause interference with usual activity.

Severe (+3): Incapacitating with inability to work or do usual activity.

NOTE: An event that is fatal should be recorded as death on the **ADVERSE EVENT** eCRF. The cause of death will be detailed on the **Serious Adverse Events** section of the form.

A relationship to device or procedure of an AE is a judgment determination made by the PI that there is a logical connection between device use (e.g., delivery system manipulation) and the occurrence of the AE.

Device Related – Complication associated with the device design as it relates to placement, efficacy or durability, (these may involve the delivery system).

Procedure Related – Complication associated with the initial placement of the device or any necessary secondary interventions. This includes morbidity associated with either anesthesia or surgical procedure. This also includes inappropriate patient selection and errors attributed to inappropriate operator techniques, measurements or judgement

8.8.2. Reporting of Adverse Events

For all adverse events occurring during the study period, data must be entered in the Adverse Event eCRF in the eDC system. For serious adverse events (SAEs), the section of the eCRF describing the serious nature of the AE must also be completed.

All adverse events will be investigated by Endologix.

The PI should supply to Endologix and the responsible EC/IRB with a complete, written case history (AE forms) and any additional redacted source documentation and information (e.g., other diagnostic testing, discharge reports, autopsy reports, etc.) as it is available.

Deaths

For any death, regardless of cause or timing, it is recommended that the site notify the Endologix Clinical Department via the EDC system within 1 working day of awareness. The PI should supply to Endologix and the responsible EC/IRB with a complete, written case history (AE forms) and any additional redacted source documentation and information (e.g., other diagnostic testing, discharge reports, autopsy reports, etc.).

In all cases, the death certificate is to be uploaded to the eDC system. Every attempt should be made to obtain as much detailed information on the events or conditions leading up to the death as possible. If the patient was hospitalized, a copy of the discharge summary source document is required. Also sites can search Social Security Death Index (SSDI) or CDC National Death Index (NDI) to obtain more information.

If the device is explanted after death, the site should notify Endologix for submission of the explant to Endologix. A device explant kit will be provided to the hospital for processing and shipment to an independent pathologist per established methods. To request a kit, send a fax to Endologix at +1 949-954-7601, or contact Endologix by telephone at +1 (949) 595-7200.

8.8.3. Adverse Events: required reporting timeframes

Note: Site reporting times to Endologix only; responsible IRB/EC reporting requirements must still be followed by the site, in addition to those recommended below.

AE Type	Reporting Recommendations
Non-device or procedure related AE	Recommended monthly at minimum
Device or procedure related AE	Within 10 working days
Non-device or procedure related SAE	Within 10 working days
Device or Procedure related SAE	Within 1 working day
Death as outcome (regardless of cause)	Within 1 working day
Unanticipated Adverse Device Effect	Within 1 working day

8.8.4. Secondary procedures

For subjects that experience any non-diagnostic invasive secondary treatment of the groin access areas or of the abdominal aortoiliac vessels (e.g., hematoma drainage, vascular exploration for bleeding, iliac stenting, thrombectomy, additional device placement for endoleak, embolization for Type II endoleak, conversion to open repair, etc.), it is important to identify the specific procedure performed on the relevant **ADVERSE EVENT eCRF**. A copy of the discharge summary source document should be uploaded to the eDC system.

8.8.5. Anticipated Adverse Events

Adverse events that could potentially occur during this investigation are called anticipated adverse events. These events must be reported if clinically significant. A list of anticipated AEs are listed in alphabetical order:

- Access site complications and sequelae (e.g., dehiscence, infection, pain, hematoma, pseudoaneurysm)
- Allergic reaction to contrast agent (e.g., pruritus, urticaria, bronchospasm, angioedema, hypotension or anaphylaxis that occurs during or post-procedure)
- Amputation
- Anesthetic complications and sequelae (e.g., aspiration)
- Aneurysm enlargement
- Aneurysm rupture
- Arterial damage or trauma (e.g., bleeding, perforation, dissection, rupture)
- Arterial or venous thrombosis and/or pseudoaneurysm
- Arteriovenous fistula
- Bleeding requiring transfusion and/or surgical intervention
- Bowel complications (e.g., ileus, ischemia, infarction, necrosis)
- Cardiac complications and sequelae (e.g., arrhythmia, myocardial infarction, congestive heart failure, hypotension, hypertension)
- Catheter or implant component fragmentation and sequelae (e.g., embolization, vessel trauma)
- Claudication
- Coagulopathy
- Death (due to any cause)
- Edema
- Embolization (micro and macro) with transient or permanent ischemia or infarction
- Endoleak

- Fever and localized inflammation
- Fistula (vascular or aeroenteric)
- Genitourinary complications and sequelae (e.g., ischemia, fistula, incontinence, hematuria, impotence, infection)
- Hepatic failure
- Infection of the aneurysm, device access site, including abscess formation, transient fever and pain.
- Local or systemic neurologic complications and sequelae, transient or permanent (e.g., stroke, transient ischemic attack, paraplegia, paraparesis, paralysis, numbness and/or tingling in legs)
- Lymphatic complications and sequelae (e.g., lymph fistula)
- Nellix implant: improper component placement; incomplete component deployment; component migration; occlusion/thrombosis; infection; stent fracture; EndoBag material wear; dilatation; erosion; puncture and perigraft flow
- Thrombosis or occlusion of stent graft or arterial vessel of the lower extremities
- Pulmonary/respiratory complications and sequelae (e.g., pneumonia, respiratory failure, prolonged intubation, pulmonary embolism)
- Renal complications and sequelae (e.g., artery occlusion, infarction, insufficiency, failure)
- Secondary Intervention
- Surgical conversion to open repair

8.8.6. Unanticipated Adverse Device Effect

The investigator shall submit to Endologix and to the reviewing EC/IRB a report of any unanticipated adverse device effect occurring during an investigation as soon as possible, but in no event later than 24 hours after the PI first learns of the effect. PIs must submit to Endologix documentation of the report made to the EC/IRB via the EDC system. Endologix will report a UADE to the FDA within 5 working days of becoming aware of the event, as required by regulations.

Please note: The following events are associated with endovascular repair and these became UADEs in the prior 3SQ+ protocol due to the higher than anticipated rates at which they were observed. This study is now being executed to evaluate the refined indications; therefore, the event rates have been returned to baseline and would not be considered UADEs without further evaluation:

- Type 1 endoleaks
- Device migration
- Aneurysm sac enlargement

8.8.7. Adverse Event Definitions

The following event definitions will be applied during this study.

- *Death:* Any death occurring during the study period, regardless of cause.
 - *Aneurysm-related death* is defined as any death occurring within 30 days from the date of the procedure, regardless of cause, and death due to aneurysm rupture or following any procedure intended to treat the aneurysm.
 - *Cardiac-related death* is defined as death due to arrhythmia, heart failure (including cardiogenic shock), or myocardial infarction.
 - *Pulmonary-related death* is defined as death due to due to pulmonary edema,

respiratory failure, or pulmonary embolism.

- *Vascular-related death* is defined as death due to stroke, cerebral hemorrhage, or other clear vascular event that is not categorized as cardiac-related or pulmonary-related.
- *Other* is to be used to identify a death due to any event that cannot be clearly categorized as above, but where some information is available.
- *Unknown* is to be used to identify a death where no information is available.
- *Procedural Technical Failure* is defined as a failure of the Nellix system to be delivered and deployed, such that the procedure is not completed, or the device failure results in a serious complication, or a residual endoleak occurs that cannot be resolved during the index procedure and is still present at the 30 day CT as read by the Core Lab.
- *Major Adverse Event*: An event occurring during the study that meets one of the following criteria:
 - *All-Cause Death* (see above);
 - *Bowel Ischemia*: the lack of adequate blood flow to the intestines that requires intensification of medical therapy or surgical/endovascular intervention;
 - *Myocardial Infarction*: the presence of raised levels of one or more cardiac biomarkers in comparison to local laboratory reference ranges;
 - *Paraplegia*: Paralysis of the lower extremities inclusive of the lower trunk;
 - *Renal Failure*: the need for temporary or permanent dialysis or $>0.5\text{mg/dL}$ increase in pre-operative serum creatinine level at two consecutive intervals;
 - *Respiratory Failure*: pneumonia or respiratory failure requiring ventilator support beyond 24 hours post-procedure;
 - *Stroke*: a sudden development of neurological deficit due to vascular lesions of the brain such as hemorrhage, embolism, or thrombosis that persists for >24 hours;
 - *Procedural Blood Loss $>1,000\text{mL}$* : Estimated blood loss during the index procedure $\geq 1,000\text{cc}$.
- *Other Definitions*:
 - *Aneurysm Sac Enlargement*: Aneurysm sac maximum diameter increase of $>5\text{mm}$ in late follow-up as compared to the one month CT time point.
 - *Aneurysm Rupture*: bleeding or leaking of blood from the aneurysm subsequent to the index procedure.
 - *Clinically significant migration*: Core Lab reported stent distal movement $>10\text{mm}$ from the original implant location relative to the center of the distal renal artery resulting in an intervention or in a serious complication.
 - *Conversion to Open Repair*: open surgical repair of the abdominal aortic aneurysm due to unsuccessful delivery or deployment of the stent graft, due to complications or other clinical situations that precluded successful

endovascular treatment, or at any time following initial successful endovascular treatment for any reason.

- *Device Positional Stability*: Core Lab reported position of each stent relative to the center of the L3 vertebrae. Stable positioning is defined as within 5mm of the positioning on the first post-operative CT scan.
- *Luminal Thrombus*: any endovascular surgical intervention after completion of the Nellix System implantation for resolution of endograft thrombosis.
- *Distal Ischemia*: New onset of compromised peripheral blood flow resulting in femoral or peripheral arterial occlusion or stenosis (attributable to the index procedure and not related to natural progression of atherosclerotic disease) causing a threat to the viability of the limb and requiring surgical or percutaneous intervention; or stent graft occlusion requiring any intervention..
- *Endoleak*: Clear evidence of contrast outside of one or both EndoBags which communicates with the aneurysm sac originating proximally at the infrarenal segment (Type IA), distally (Type IB); between components, if an extender is used (Type III); trans-device (Type IV); or from a patent collateral vessel (Type II: e.g., lumbar artery; inferior mesenteric artery). Contrast outside of an EndoBag that does not communicate with the aneurysm sac is not to be reported as an endoleak.
- *Migration*: Core Lab reported stent distal movement >5mm from the original implant location relative to the center of the distal renal artery.
- *Occlusion*: Intervention for stent occlusion or as reported by the Core Lab.
- *Secondary procedure*: any non-diagnostic intervention after the index procedure intended to correct or repair an endoleak, device occlusion, migration, aneurysm sac expansion and/or a device defect (including infection).

9. RISK ANALYSIS

The decision to repair an aortic aneurysm is generally based on the risk of rupture, the risk of complications of surgery, and patient preference. There are currently two methods used to repair aortic aneurysms. The conventional method is an open surgical repair, with the implantation of a synthetic graft to replace the diseased aneurysmal vessel through a large abdominal incision. More recent technological developments have resulted in an alternative, minimally invasive, endovascular aneurysm repair, in which a prosthesis is placed within the aorta through a small incision in the groin. Blood can then flow through the implanted device and is excluded from the aneurysmal portion of the aorta.

The disadvantages of open surgical repair are: general anesthesia is required, it is a major abdominal surgery (large incision), has a significant surgical complication rate, and typically requires a long hospital stay and recovery. EVAR/EVAS enables local or regional anesthesia to be used, uses a minimally invasive groin incision for catheter-based access, and has been reported in multiple studies to offer a lower perioperative complication rate, reduced blood loss and procedure times, and shorter hospital stay. In contrast to open repair, EVAR/EVAS is a relatively new treatment, long term results have not been fully established, and life-long

surveillance is recommended to verify implant integrity and patency and continued aneurysm exclusion. Multiple device systems are CE Marked and marketed in the European Community for endovascular abdominal aortic aneurysm repair. All of these devices require the introduction of catheter-based treatment devices varying in outer diameter profile. Standard vascular exposure is indicated for access. Prospective clinical study results support the safety and effectiveness of these devices through early follow-up (to 30 days) and for some in late follow-up to one year and beyond to up to five years.

As with any procedure there are risks of serious complications, such as death. The inclusion and exclusion criteria for this population have been carefully established to limit the risk of mortality and morbidity in this population. The overall risk will be evaluated on an individual basis and discussed with each patient. All of the potential adverse events outlined previously could cause prolonged illness, permanent impairment of daily function or, in rare cases, death. Possible treatments could include, but are not limited to, emergency cardiac or vascular surgery.

Eligibility criteria that exclude subjects who are at higher risk for experiencing an anticipated adverse event have been selected to reduce the potential risks to subjects who participate in this study. In addition, the assessment of patient computed tomography angiography imaging by an experienced Core Laboratory is also intended to reduce the potential risks to subjects who participate in this study.

Pre-procedural high resolution, contrast-enhanced CT scanning and intraprocedural arteriography will be used to identify and target the aortic anatomy to facilitate the proper introduction, delivery, and deployment of the endovascular repair devices. Physician experience, rigorous application of a common protocol, and careful performance of the procedure with close monitoring of the patient after the procedure will also help to minimize risks.

The alternative to endovascular repair of AAA is open surgical repair.

10. STATISTICAL PROCEDURES

10.1 ANALYSIS POPULATIONS

Refer to the SAP for details.

10.2 PATIENT ACCOUNTABILITY AND MISSING DATA

The handling of missing data and patient compliance is discussed in the Statistical Analysis Plan (SAP).

10.3 GENERAL NOTES

Analyses of MAEs and other adverse events will be conducted at exact time points. Early events are defined as those occurring from the date of the procedure to 30 calendar days post-operatively. Late events are defined as those occurring from after 30 calendar days post-operatively (from day 31 forward). Events within six months and one year, respectively, are those occurring up to and including 180 days and 365 days following the procedure.

10.4 BASELINE COMPARABILITY ANALYSIS

A set of important demographic or prognostic variables will be compared across study sites to determine homogeneity of sites in terms of baseline patient characteristics. Factors found to differ significantly by site will identify that variable or site as a possible covariate in subsequent analyses.

In the analysis of comparability of trial sites, it is anticipated that a small number of sites may have a patient enrollment that is too small to allow meaningful statistical analysis. To allow the inclusion of sites with this condition, small sites will be combined into one pseudo-site the size of which will not exceed the size of the largest trial site. It is not possible to determine the number of subjects below which this aggregation will occur until after enrollment is completed, but generally trial sites with four or fewer subjects may be combined.

10.5 POOLING

Data among the pivotal cohort will be pooled on a clinical basis including harmonized protocol, use of the same Core Lab, screening procedure, endpoints and the same training on the same implantable device and console.

10.6 PRIMARY SAFETY ANALYSIS

10.6.1 Primary Safety Hypothesis

The primary safety endpoint will be analyzed at 30 days. It is defined as the incidence of Major Adverse Events (MAE) at 30 days, defined as the composite of the following:

- *All-Cause Mortality;*
- *Bowel Ischemia;*
- *Myocardial Infarction;*
- *Paraplegia;*
- *Renal Failure;*
- *Respiratory Failure;*
- *Stroke;*
- *Procedural Blood Loss $\geq 1,000\text{cc}$*

Details for this analysis are provided in the SAP.

10.7 SAMPLE SIZE

The sample size for this study is guided by the analysis of the effectiveness endpoint, using a Frequentist exact approach. All studies with a harmonized protocol can be included in the analysis to meet the minimum sample size. Details of power and type I error control are described in the SAP.

10.8 ADDITIONAL EVALUATIONS

- *Procedural and In-Hospital Evaluations:* The estimated blood loss volume, % of subjects receiving a blood transfusion, contrast volume, fluoroscopy time, endovascular device time, Polymer volume used, procedure time, anesthesia time and type, frequency of concomitant procedures, ICU time, time to hospital discharge, and % of subjects discharged to skilled nursing facilities or assisted living facilities who were not prior residents will be presented descriptively. Qualitative variables will be presented with rate. Quantitative variables will be presented with mean, standard deviation, median, minimum and maximum.
- *Mortality:* The proportion of subjects with mortality (all-cause and aneurysm-related) will be presented at 30 days and at each subsequent time point for which data are available (i.e.,

six months, and annually at years 1 through 5). A descriptive analysis and a Kaplan-Meier analysis of all-cause and aneurysm-related mortality through 30 days and for late follow-up including 30 days will be presented.

- *Major Adverse Events (MAE) Individual Components:* The proportion of subjects experiencing each category of MAE within 30 days and after 30 days at each subsequent time point for which data are available will be presented.
- *Composite Late MAEs (after 30 days):* The proportion of subjects experiencing an MAE after 30 days will be presented at each subsequent time point for which data are available. A descriptive analysis and a Kaplan-Meier analysis of the MAE composite will be presented for follow-up through 30 days and for late follow-up including 30 days. The descriptive analysis will present the rates for the composite MAE and each individual adverse event that is part of the composite.
- *Aneurysm Rupture:* The proportion of subjects experiencing a rupture of the aneurysm will be presented at 30 days and at each subsequent time point for which data are available.
- *Conversion to Open Repair:* The proportion of subjects undergoing surgical conversion to open repair will be presented at 30 days and at each subsequent time point for which data are available.
- *Serious Adverse Events (SAEs):* The proportion of subjects experiencing an SAE will be presented descriptively by time point.
- *Non-serious AEs:* The proportion of subjects experiencing a non-serious AE will be presented descriptively by time point.
- *Device Performance:* A descriptive analysis of the following will be done based on Core Lab evaluations at 30 days and at each subsequent time point for which data are available. A determination will be made from the data as to whether any observation was associated with an SAE or device malfunction.
- *Luminal Thrombus:* any endovascular surgical intervention after completion of the Nellix System implant for resolution of endograft thrombosis.
- *Aneurysm Sac Changes:* The changes in aneurysm sac (vessel) diameter from the first post-operative CT evaluation will be presented at each subsequent follow-up. The maximum diameter parameter will be used for the analysis because it is the basis for comparison to other studies with endografts. For each follow-up interval, the number and percent of subjects with increased sac diameter ($>5\text{mm}$), stable sac diameter ($\leq 5\text{mm}$), or decreased sac diameter ($>5\text{mm}$) will be presented.
- *Incidence of Migration:* Core Lab reported stent distal movement $>5\text{mm}$ from the original implant location relative to the center of the distal renal artery.
- *Incidence of Clinically Significant Migration:* Core Lab reported stent distal movement $>10\text{mm}$ from the original implant location relative to the center of the distal renal artery resulting in an intervention or in a serious complication.
- *Incidence of Endoleak:* The occurrence of endoleak (any type or location, new and persistent) will be presented descriptively at each follow-up.
- *Incidence of Stent Occlusion:* The occurrence of stent occlusion will be presented descriptively at each follow-up. Renal Function Evaluations: A descriptive analysis of renal function will be done based on serum creatinine levels and calculated estimated glomerular filtration rate (eGFR) at 30 days and at each subsequent timepoint for which data are

available. Renal dysfunction will be defined as a reduction in eGFR of >30% from the preoperative value.

- *Nellix Implant Patency and Integrity:* A descriptive analysis of stent patent luminal flow, absence of stent fracture, kinking, or occlusion, and absence of device failure will be determined by contrast-enhanced CT scan, as assessed by the independent core laboratory at 30 days and at each subsequent time point for which data are available. A determination will be made from the data as to whether the occurrence was associated with an SAE or device malfunction.
- *Secondary Endovascular procedures:* A descriptive analysis of secondary endovascular procedures for resolution of endoleak (Type I or III), device occlusion, migration, aneurysm sac expansion, and/or a device defect (including infection) will be done at 30 days and at each subsequent time point for which data are available.

11. ADMINISTRATIVE PROCEDURES

11.1 STATEMENT OF COMPLIANCE

This clinical investigation will follow the international standard for Clinical Investigation of Medical Devices for Human Subjects -- Good Clinical Practice (ISO 14155:2011). The Sponsor and Investigator(s) will comply with their responsibilities as defined in 21 CFR 812.

11.2 RESPONSIBILITIES

1. *Sponsor:* Endologix is responsible as the Sponsor to ensure proper site and investigator selection, availability of signed investigator agreements prior to study initiation, availability of regulatory and EC/IRB approval prior to the initiation of the study at any site, obtaining and maintaining appropriate insurance policies for the study, and management and monitoring of the study with special attention to verification of all clinical requirements, adherence to protocol, good clinical practices and compliance with applicable government and institutional regulations. Furthermore, the sponsor is responsible for ensuring proper regulatory approvals are obtained, and reporting to regulatory authorities per applicable regulations.
2. *Investigators:* Each investigator and study site is required to conduct the clinical investigation in accordance with the protocol, the signed investigator agreement, Good Clinical Practices (GCP), all applicable laws and regulations and any conditions or restrictions imposed by the reviewing EC/IRB. This includes compliance with requirements related to EC/IRB approval and reporting, and proper patient informed consent prior to participation in the study. The investigator is also responsible for protecting the rights, safety, and welfare of the subjects under his or her care.
 - a. Each investigator is responsible for supervising all procedures conducted under this protocol at his or her institution.
 - b. Furthermore, the investigator is responsible for ensuring that data are completely, accurately, and promptly entered in each patient's eCRFs and related documents are available to verify the accuracy of the eCRFs, and for ensuring the clinical monitor has access to all necessary records to ensure the integrity of the data.
3. *CT Scan Core Laboratory:* The CT Scan Core Laboratory is a contract supplier having validated processes that is responsible for receipt, de-identification, handling, processing, and

- assessment of all submitted CT scans preoperatively and postoperatively and for reporting of results per a written, approved protocol.
4. *Data Management Group*: The data management group is responsible for electronic data capture database development, validation, control and management of input from study sites/monitored data, maintenance, and reporting for statistical analysis.
 5. *Clinical Events Committee*: The independent clinical events committee will consist of at least three physician members and is responsible for review of events and complications documented on the eCRFs and in source documents by the study PIs during the trial, and for categorization of these events according to the event definitions and primary endpoint criteria in this protocol. The CEC will review all MAEs, Device-Related SAEs and Deaths that occur in the study. Reviews will occur on an ongoing basis throughout the trial as events are reported. The adjudicated events will be reviewed by the Data Safety Monitoring Board.
 6. *Data Safety Monitoring Board (DSMB)*: Consistent with the U.S. FDA guidance document *Establishment and Operation of Clinical Trial Data Monitoring Committees*, Endologix has established a DSMB having pertinent expertise to review on a regular basis adjudicated safety data accumulated and trial progress (i.e., enrollment among groups; completeness and timeliness of data; protocol deviations; etc.) from its ongoing clinical investigations. The DSMB consists of five members, two of which must be physicians with specialty training in endovascular repair. One member of the DSMB is a statistician. The DSMB will be convened to review interim data accumulated during the trial and to render a recommendation for enrollment in the trial: continue; amend; suspend; or, terminate. All action items discussed during the meeting will be documented in the minutes, with agreed upon target completion dates for resolution. The final meeting minutes will be distributed to all DSMB members. Endologix will maintain sole decision-making of implementing recommendations proposed by the DSMB. Endologix will disclose the recommendations of the DSMB to the US FDA in its annual progress reports to the IDE.
 7. *Case Review Board (CRB)*: The Case Review Board will contain experts in endovascular therapy for repair of abdominal aortic and aortoiliac aneurysms. The Case Review Board will be chaired by a physician. The CRB will review all aneurysm-related images prior to subject enrollment into the study. Approval by the CRB will include assessing vascular suitability for treatment with the study device and review of anatomic measurements provided by the Core Lab to determine each patient's eligibility.
 8. *Biostatistician*: The biostatistician is responsible for the development and implementation of the data analysis plan, for conducting the data analysis, and for reporting it per the plan.

11.3 PATIENT PROTECTION

This clinical investigation will be conducted in accordance with the ethical principles that have their origin in the Declaration of Helsinki. Written informed consent, in accordance with applicable international and local study center regulations, must be obtained from each patient, or from their legal representative, prior to the formal screening process as outlined in this protocol. The PI will retain a copy of the signed informed consent document in each subject's record, and provide a copy to the subject. The PI will not request the written informed consent of any patient, and will not allow any patient to participate in the investigation before obtaining EC/IRB approval.

Appropriate precautions will be taken to maintain confidentiality of patient medical records and personal information. However, the patient's name may be disclosed to the sponsor or designee, or any health authorities if they inspect the study records. A report of this study may be published; however the patient's identity will not be disclosed.

11.4 PROTOCOL CHANGES

The PI should not implement any deviation from or changes to the protocol without approval by Endologix and prior review and documented approval from the governing EC/IRB. The only exception to this is where necessary to eliminate immediate hazards to study subjects, or when changes involve only administrative aspects (e.g., change in monitors, telephone numbers, etc.).

A report of withdrawal of EC/ IRB approval must be submitted to the Sponsor within five working days.

11.5 DOCUMENTATION

Source Documents: Source documents may include a patient's medical record, hospital charts, clinic charts, the investigator's study files, questionnaires, as well as the results of diagnostic tests such as laboratory tests, CT scans, angiograms, and the like. Source document worksheets will be provided by Endologix for use by study personnel.

The following information should be included in the patient's medical record:

- The study title, number/name, and sponsor name;
- The patient ID number;
- A statement that written informed consent was obtained;
- Date of procedure and implanted device information;
- Dates of all visits;
- Occurrence of complication and adverse events, hospitalizations, or adverse events;
- Date patient exited the study, and a notation as to whether the patient completed the study or discontinued, with the corresponding reason.

Electronic Case Report Form Completion: The PI who signs the protocol signature page must personally electronically sign the eCRFs to ensure that the observations and findings are entered correctly and completely. The eCRFs are to be entered in a timely manner at the intervals specified by Endologix and this protocol.

All eCRFs must be entered as instructed. An explanation must be provided for any missing data points. Completion instructions and training will be provided to each site for reference.

All eCRFs will be reviewed and monitored for completeness and clarity. Queries for missing or unclear data will be made as necessary throughout the study. Timely resolution of queries is required by institutional staff and the PI.

Study Logs: The following logs will be maintained for the study:

- Investigator and site training to the protocol (Training Log)
- Authorized study site personnel (Site Signature Log)

- Patient consent and screening (Enrollment Log)
- Monitoring visit tracking (Site Visit Log)
- Nellix Inventory eTracking Log (Device Accountability Log)

Document Retention: Study-related correspondence, patient records, consent forms, records of device implant, and source documents are to be maintained by the study site. Endologix requires that it be notified in writing if the PI wishes to relinquish ownership of the data and information so that mutually agreed upon arrangements can be made for transfer of ownership to a qualified entity.

Publication: Endologix, as the sponsor of record, has a proprietary interest in this study. Authorship and manuscript composition will reflect joint cooperation between multiple investigators and sites, core laboratories, and Endologix. Authorship will be established prior to writing of the manuscript. No individual publications will be allowed prior to the completion of the final report for this study and as agreed in writing by Endologix.

11.6 MONITORING PLAN

Written procedures have been established by Endologix for monitoring clinical investigations, to assure the quality of the study and to assure that each person involved in the monitoring process carries out his or her duties. Standardized written procedures, sufficiently detailed to cover the general aspects of clinical investigations, will be used as a basic monitoring plan and will be supplemented by more specific or additional procedures, as required by the clinical investigation. A pre-study monitoring visit or Site Initiation Visit will be conducted to ensure that the PI clearly understands and accepts the obligations incurred in undertaking the clinical investigation as set forth in relevant international standard, and that the facilities are acceptable. Periodic monitoring visits will be conducted with adequate frequency to ensure that the PI's obligations are being fulfilled and that the facilities continue to be acceptable.

Site Termination: If the Sponsor or a clinical monitor becomes aware that a PI is not complying with the signed Investigator's Agreement, the Investigational Plan, the requirements of applicable health authority regulations, or any conditions of approval imposed by the reviewing EC/IRB or health authority, Endologix will immediately either secure compliance or terminate the PI's participation in the study. The final action will be taken with the goal of assuring the rights, safety and welfare of the subjects.

12. REFERENCES

1. AbuRahma A, Campbell J, Stone PA, et al. The correlation of aortic neck length to early and late outcomes in endovascular repair patients. *J Vasc Surg* 2009;50:738-48.
2. Arko FA, Filis KA, Seidel SA, et al. How many patients with infrarenal aneurysms are candidates for endovascular repair? *J Endovasc Ther* 2004;11:33-40.
3. ASTM F1058:2002, Standard Specification for Wrought 40Cobalt-20Chromium-16Iron-15Nickel-7Molybdenum Alloy Wire and Strip for Surgical Implant Applications (UNS R30003 and UNS R30008).
4. Browne TF, Hartley D, Purchas S, et al. A fenestrated covered suprarenal aortic stent. *Eur J Vasc Endovasc Surg* 1999;18:445-9.
5. Carpenter, JP, Cuff, R, Buckley, C et al. One-year pivotal trial outcomes of the Nellix system for endovascular aneurysm sealing. *J Vasc Surg* 2017;65:330-6.
6. Carpenter, JP, Cuff, R, Buckley, C et al. Results of the Nellix system investigational device exemption pivotal trial for endovascular aneurysm sealing. *J Vasc Surg* 2016;63:23-31.
7. Carpenter JP, Baum RA, Barker CF, et al. Impact of exclusion criteria on patient selection for endovascular abdominal aortic aneurysm repair. *J Vasc Surg* 2001;34:1050-4.
8. Fleiss, J. The statistical basis of meta-analysis. *Statistical Methods in Medical Research* 1993;2: 121-45.
9. Friedman L, Furberg C, and DeMets D. *Fundamentals of Clinical Trials*. 1985: Mosby Year Book, St. Louis.
10. Fulton JJ, Farber MA, Marston WA, et al. Endovascular stent-graft repair of pararenal and type IV thoracoabdominal aortic aneurysms with adjunctive visceral reconstruction. *J Vasc Surg* 2005;41:191-8.
11. Harrell F, Lee K, and Mark D. Multivariable prognostic models: issues in developing models, evaluating assumptions, and adequacy, and measuring and reducing errors. *Statistics in Medicine* 1996;15:361-87
12. Hoyert DL, Arias E, Smith B L, et al. Deaths: final data for 1999. *Natl Vital Stat Rep* 2001; 49(8):1-113.
13. Jeyabalan G, Park T, Rhee RY, et al. Comparison of modern open infrarenal and pararenal abdominal aortic aneurysm repair on early outcomes and renal dysfunction at one year. *J Vasc Surg* 2011;54:654-9.
14. ISO 14155:2011, *Clinical Investigations of Medical Devices for Human Patients: Good Clinical Practice*
15. Leurs LJ, Kievit J, Dagnelie PC, et al. Influence of infrarenal neck length on outcome of endovascular abdominal aortic aneurysm repair. *J Endovasc Ther* 2006;13:640-8.
16. Meinert C. *Clinical Trials: Design, Conduct, and Analysis*. 1986; Oxford University Press. New York, NY
17. Park JH, Chung JW, Cho IW, et al. Fenestrated stent grafts for preserving visceral branches in the treatment of abdominal aortic aneurysms: preliminary results. *J Vasc Interventional Radiology* 1996;7:819-23.
18. Pocock S. *Clinical Trials (A Practical Approach)*. 1983: John Wiley and Sons, Chichester.
19. Rubin D. *Multiple Imputation for Nonresponse in Surveys*. 1987: John Wiley and Sons, New York.
20. Turnbull I, Criado FJ, Sanchez L, et al. Five year results for the Talent enhanced low profile system abdominal stent graft pivotal trial including early and long term safety and efficacy. *J Vasc Surg* 2010; 51:537-44.
21. U.S. Department of Health and Human Services, *Guidance For Industry. E9 Statistical Principles for Clinical Trials*, September 1998.
22. U.S. Department of Health and Human Services, *Guidance For Industry. ICH E6, Good Clinical Practice: Consolidated Guidance*, April 1996.

**13. APPENDIX A: LIST OF COILS INVESTIGATORS MAY USE ADJUNCTIVELY
WITH ONYX®**

Coil Name	Manufacturer
Interlock	Boston Scientific
Nester	Cook Medical
Concerto	Medtronic
POD	Penumbra
Ruby	Penumbra
Azur	Terumo