

Title

Clinical Evaluation of a Daily Wear Frequent Replacement Silicone Hydrogel Lens

Protocol Number: CLL949-C009 / NCT03920280

Development Stage of Project:
[REDACTED]Sponsor Name and Address:
Alcon Research, LLC. and its affiliates ("Alcon")
6201 South Freeway
Fort Worth, Texas 76134-2099

Test Product: LID015385

*Property of Alcon
Confidential*

*May not be used, divulged, published, or otherwise disclosed without the consent of
Alcon*

Investigator Agreement:

- I have read the clinical study described herein, recognize its confidentiality, and agree to conduct the described trial in compliance with Good Clinical Practice (GCP), the ethical principles contained within the Declaration of Helsinki, this protocol, all applicable regulatory authority regulations, and conditions of approval imposed by the reviewing IRB or regulatory authority.
- I will supervise all testing of the device involving human subjects and ensure that the requirements relating to obtaining informed consent and IRB review and approval are met in accordance with applicable local and governmental regulations.
- I have read and understand the appropriate use of the investigational product(s) as described in the protocol, current Investigator's Brochure, product information, or other sources provided by the Sponsor.
- I understand the potential risks and side effects of the investigational product(s).
- I agree to maintain adequate and accurate records in accordance with government regulations and to make those records available for inspection.
- I agree to comply with all other requirements regarding the obligations of clinical Investigators and all other pertinent requirements of the Sponsor and government agencies.
- I agree to ensure that all associates, colleagues, and employees assisting in the conduct of the study are informed of their obligations in meeting the above commitments.

Have you ever been disqualified as an Investigator by any Regulatory Authority?

No Yes

Have you ever been involved in a study or other research that was terminated?

No Yes

If yes, please explain here:

Principal Investigator:

Signature

Date

Name and professional position:

Address:

Table of Contents

Clinical Evaluation of a Daily Wear Frequent Replacement Silicone Hydrogel Lens	1
Table of Contents	3
List of Tables	5
List of Figures	6
1 GLOSSARY OF TERMS	7
2 LIST OF ACRONYMS AND ABBREVIATIONS	11
3 PROTOCOL SUMMARY	13
4 [REDACTED]	
5 [REDACTED]	
5 INTRODUCTION	22
5.1 Rationale and Background	22
5.2 Purpose of the Study	23
5.3 Risks and Benefits	23
6 STUDY OBJECTIVES	24
6.1 Primary Objective(s)	24
6.2 Secondary Objective(s)	24
6.3 [REDACTED]	
6.4 Safety Objective(s)	24
7 INVESTIGATIONAL PLAN	25
7.1 Study Design	25
7.2 [REDACTED]	
7.3 [REDACTED]	
7.4 Rationale for Choice of Control Product	26
7.5 Data Monitoring Committee	27
8 STUDY POPULATION	27
8.1 Inclusion Criteria	27
8.2 Exclusion Criteria	28
8.3 Rescreening of Subjects	29
9 TREATMENTS ADMINISTERED	29
9.1 Investigational Product(s)	29

9.3	Treatment Assignment / Randomization	35
9.4	Treatment masking	35
9.5	Accountability Procedures.....	37
9.6	Changes to concomitant medications, treatments/ procedures.....	38
10	STUDY PROCEDURES AND ASSESSMENTS	38
10.1	Informed Consent and Screening	39
10.2	Description of Study Procedures and Assessments	40
10.2.1	Demographics.....	40
10.2.2	Medical History and Concomitant Medications.....	40
10.2.3	Investigational Product Compliance.....	40
10.2.4	Adverse Event Collection: Safety Assessment.....	40
10.2.5	Slit-Lamp Biomicroscopy: Safety Assessment	40
10.2.6	Device Deficiencies: Safety Assessment.....	40
10.5	Discontinued Subjects	42
10.5.1	Screen Failures	42
10.5.2	Discontinuations	42
10.5.3	Schedule of Procedures and Assessments for Subjects Discontinued from Investigational Product	43
10.6	Clinical Study Termination.....	43
10.6.1	Follow-up of subjects after study participation has ended	44
11	ADVERSE EVENTS AND DEVICE DEFICIENCIES	44
11.1	General Information	44
11.2	Monitoring for AEs.....	47
11.3	Procedures for Recording and Reporting	47
11.5	Unmasking of the Study Treatment.....	50
11.6	Follow-Up of Subjects with AEs.....	50
11.7	Pregnancy in the Clinical Study	50
12	ANALYSIS PLAN	51
12.1	Subject Evaluability.....	51
12.2	Analysis Sets	51

12.3	Demographic and Baseline Characteristics	51
12.4	Effectiveness Analyses	51
12.4.1	Analysis of Primary Effectiveness Endpoint(s).....	52
12.4.1.1	Statistical Hypotheses	52
12.4.1.2	Analysis Methods	52
12.5	Handling of Missing Data	53
12.6	Safety Analyses	53
13	DATA HANDLING AND ADMINISTRATIVE REQUIREMENTS	54
13.1	Subject Confidentiality.....	54
13.2	Completion of Source Documents and Case Report Forms	55
13.3	Data Review and Clarifications.....	56
13.4	Sponsor and Monitoring Responsibilities	56
13.5	Regulatory Documentation and Records Retention	56
13.6	Quality Assurance and Quality Control.....	57
14	ETHICS	57
15	REFERENCES	58
15.1	References applicable for all clinical studies	58
15.1.1	US references applicable for clinical studies.....	58

List of Tables

Table 2-1	List of Acronyms and Abbreviations Used in This Protocol	11
Table 3-1	Schedule of Study Procedures and Assessments	18
Table 6-1	Primary Objective(s).....	24
Table 6-2	Safety Objective(s).....	24
Table 9-1	Test Product	30

Table 9–2	Control Product.....	32
-----------	----------------------	----



List of Figures

Figure 7-1	Flowchart of Study Visits.....	26
Figure 11-1	Categorization of All AEs	44
Figure 11-2	Categorization of All SAE	45

1 GLOSSARY OF TERMS

Names of test product(s)	Throughout this document, test product(s) will be referred to as [REDACTED] soft contact lenses [REDACTED] [REDACTED]
Name of Control Product(s)	CooperVision® BIORFINITY® soft contact lenses (Biofinity contact lenses or Biofinity)
Adverse Device Effect (ADE)	Adverse event related to the use of an investigational medical device (test product) or control product. <i>Note: This definition includes adverse events resulting from insufficient or inadequate instructions for use, deployment, implantation, installation, or operation; any malfunction; and use error or intentional misuse of the test product or control product.</i>
Adverse Event (AE)	Any untoward medical occurrence, unintended disease or injury, or untoward clinical signs (including abnormal laboratory findings) in subjects, users or other persons, whether or not related to the investigational medical device (test product). <i>Note: For subjects, this definition includes events related to the test product, the control product, or the procedures involved. For users or other persons, this definition is restricted to events related to the test product.</i> Requirements for reporting Adverse Events in the study can be found in Section 11.
Anticipated Serious Adverse Device Effect	Serious adverse device effect which by its nature, incidence, severity, or outcome has been identified in the risk management file.
Device Deficiency	Inadequacy of a medical device with respect to its identity, quality, durability, reliability, safety, or performance. <i>Note: This definition includes malfunctions, use errors, and inadequate labeling.</i>

	Requirements for reporting Device Deficiencies in the study can be found in Section 11.
Enrolled Subject	Any subject who signs an informed consent form for participation in the study.
Interventional Clinical Trial	A research trial that prospectively assigns, whether randomly or not, human participants or groups of humans to one or more health-related interventions to evaluate the effects on health outcomes, and/or a research trial in which diagnostic or monitoring procedures beyond standard of care are conducted and generate outcomes for use in analysis of data.
Investigational Product	Is defined as a preventative (vaccine), a therapeutic (drug or biologic), device, diagnostic, or palliative used as a test or control product in a clinical trial, including a product with a marketing authorization when used or assembled (formulated or packaged) in a way different from the authorized form, or when used for an unauthorized indication, or when used to gain further information about the authorized form.
Malfunction	Failure of a medical device to perform in accordance with its intended purpose when used in accordance with the instructions for use or clinical investigation plan.
Non-serious Adverse Event	Adverse event that does not meet the criteria for a serious adverse event.
Randomized Subjects	Any subject who is assigned a randomized treatment.
Serious Adverse Device Effect (SADE)	Adverse device effect that has resulted in any of the consequences characteristic of a serious adverse event.
Serious Adverse Event (SAE)	Adverse event that led to any of the following: <ul style="list-style-type: none"> • Death.

	<ul style="list-style-type: none">• A serious deterioration in the health of the subject that either resulted in:<ol style="list-style-type: none">a. a life-threatening illness or injury. <i>Note: Life-threatening means that the individual was at immediate risk of death from the event as it occurred, ie, it does not include an event which hypothetically might have caused death had it occurred in a more severe form.</i>b. any potentially sight-threatening event or permanent impairment to a body structure or a body function.c. in-patient hospitalization or prolonged hospitalization. <i>Note: Planned hospitalization for a pre-existing condition, without serious deterioration in health, is not considered a serious adverse event. In general, hospitalization signifies that the individual remained at the hospital or emergency ward for observation and/or treatment (usually involving an overnight stay) that would not have been appropriate in the physician's office or an out-patient setting. Complications that occur during hospitalization are adverse events. If a complication prolongs hospitalization or fulfills any other serious criteria, the event is serious. When in doubt as to whether "hospitalization" occurred, the event should be considered serious.</i>d. a medical or surgical intervention to prevent a) or b).e. any indirect harm as a consequence of incorrect diagnostic test results when used within manufacturer's instructions for use.
--	---

	<ul style="list-style-type: none">• Fetal distress, fetal death, or a congenital abnormality or birth defect. <p><i>Refer to Section 11 for additional SAEs.</i></p>
Significant Non-Serious Adverse Event	<p>Is a symptomatic, device-related, non-sight threatening adverse event that warrants discontinuation of any contact lens wear for greater than or equal to 2 weeks.</p> <p><i>Refer to Section 11 for additional Significant Non-Serious AEs.</i></p>
Unanticipated Serious Adverse Device Effect	Serious adverse device effect which by its nature, incidence, severity or outcome has not been identified in the risk management file.
Use Error	Act or omission of an act that results in a different medical device response than intended by manufacturer or expected by user. <i>Note: This definition includes slips, lapses, and mistakes. An unexpected physiological response of the subject does not in itself constitute a use error.</i>

2 LIST OF ACRONYMS AND ABBREVIATIONS

Table 2-1

List of Acronyms and Abbreviations Used in This Protocol

Abbreviation	Definition
ADE	Adverse device effect
AE	Adverse event
BCVA	Best corrected visual acuity
Biofinity contact lens or Biofinity	CooperVision BIOFINITY (comfilcon A) soft contact lenses
CFR	Code of Federal Regulations
CLEAR CARE	CLEAR CARE Cleaning & Disinfecting Solution
CRF	Case report form
CSM	Clinical site manager
CTT	Clinical trial team
D	Diopter(s)
D/C	Discontinue
eCRF	Electronic case report form
EDC	Electronic data capture
FDA	US Food and Drug Administration
GCP	Good Clinical Practice
IB	Investigator's brochure
ICF	Informed consent form
ICH	International Council for Harmonisation of Technical Requirements for Pharmaceuticals for Human Use
IEC	Independent ethics committee
IP	Investigational product
IRB	Institutional review board
ISO	International Organization for Standardization
LID	Lens identification
logMAR	Logarithm of the minimum angle of resolution
mm	Millimeter
N/A	Not applicable
OD	Right eye
OS	Left eye
OU	Both eyes
SAE	Serious adverse event

Abbreviation	Definition
SADE	Serious adverse device effect
SiHy	Silicone hydrogel
SOP	Standard operating procedure
[REDACTED]	[REDACTED]
US / USA	United States of America
USAN	United States Adopted Name
USV	Unscheduled visit
VA	Visual acuity
vs	Versus

3 PROTOCOL SUMMARY

This will be a prospective, randomized [REDACTED] controlled, double-masked, parallel-group clinical study.

Approximately 8 sites in the US will enroll approximately 120 subjects. Subjects will be expected to attend 6 office visits: Screening / Baseline / Dispense, Week 1 Follow up, Week 2 Follow-up, Month 1 Follow-up, Month 2 Follow-up and Month 3 Follow up/Exit. The total expected duration of participation for each subject is approximately 3 months in this daily wear clinical study. Approximately 80 subjects will be assigned to wear the test lenses and 40 subjects will be assigned to wear the control lenses, [REDACTED]

Subjects and the study personnel conducting the study evaluations will be masked to treatment. Subjects who meet the inclusion and exclusion criteria will be randomized to wear either the test contact lenses [REDACTED] in both eyes or the control contact lenses (Biofinity) in both eyes for 3 months of daily wear exposure. [REDACTED]

[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED].

At the Screening / Baseline / Dispense Visit, study lenses will be dispensed to qualified subjects [REDACTED]

All study lenses will be worn for at least 5 days per week and 8 hours per day in a daily wear modality (eg, will not be worn while sleeping). [REDACTED]

[REDACTED]
[REDACTED]
[REDACTED]

[REDACTED] Regardless of whether the subject is randomized to the test or the control group, CLEAR CARE® Cleaning & Disinfecting Solution (CLEAR CARE) must be used for cleaning and disinfection. [REDACTED]

Investigational product type	Medical Device
Study type	Interventional
Investigational products	Test Product: [REDACTED] soft contact lens Control Product: Biofinity soft contact lens
Purpose and rationale	The purpose of this clinical study is to demonstrate the performance of the investigational [REDACTED] soft contact lens compared to the commercially available Biofinity soft contact lens when worn in a daily wear modality, by assessing visual acuity as the primary variable.
Objective(s)	The primary objective is to demonstrate safety and effectiveness of the [REDACTED] soft contact lens when worn for daily wear as compared to the Biofinity soft contact lens.
Endpoint(s)	Primary Effectiveness <ul style="list-style-type: none">Distance VA (Snellen) [REDACTED] [REDACTED] [REDACTED] [REDACTED] [REDACTED]

	<p>■ [REDACTED]</p> <p>Safety</p> <ul style="list-style-type: none"> • AEs • Biomicroscopy • Device deficiencies
Study Design	<p>This will be a prospective, randomized, [REDACTED] controlled, double-masked, parallel-group, daily wear clinical study. Subject participation in the study will be approximately 14 weeks with approximately 3 months of exposure to study lenses.</p>
Subject population	<p>Volunteer subjects aged 18 or over who are adapted daily wear frequent replacement soft contact lens wearers, excluding Biofinity habitual wearers, have at least 3 months of soft contact lens wearing experience, and who wear their habitual lenses at least 5 days per week and at least 8 hours per day.</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p>
Key inclusion criteria (See Section 8.1 for a	<ul style="list-style-type: none"> • Successful wear of spherical daily wear frequent replacement soft contact lenses for distance correction in both eyes during the past 3 months for a minimum of 5 days per week and 8 hours per day.

complete list of inclusion criteria)	<ul style="list-style-type: none"> Manifest cylinder \leq 0.75 D in each eye
Key exclusion criteria (See Section 8.2 for a complete list of exclusion criteria)	<ul style="list-style-type: none"> Wearing habitual contact lenses in an extended wear modality (routinely sleeping in lenses for at least 1 night per week) over the last 3 months prior to enrollment. Monovision contact lens wearers Habitually wearing Biofinity lenses
Data analysis and sample size justification	<p>[REDACTED]</p> <p>[REDACTED]</p> <p>effectiveness and safety data will be presented separately based upon subject study status of Completed or Discontinued.</p> <p>No formal hypotheses are formulated for the primary effectiveness endpoint, hence no inferential testing will be performed.</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p>
Key words	[REDACTED] Biofinity, daily wear, [REDACTED]
Associated materials	<p>CLEAR CARE Cleaning & Disinfecting Solution will be used for daily cleaning and disinfection.</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p> <p>[REDACTED]</p>

Table 3-1 Schedule of Study Procedures and Assessments

Procedure/ Assessment	Visit 1 Screening/ Baseline/ Dispense	Visit 2 Week 1 Follow-up	Visit 3 Week 2 Follow-up	Visit 4 Month 1 Follow-up	Visit 5 Month 2 Follow-up	Visit 6 Month 3 Follow-up/Exit	Early Exit	USV
	Day 1	Day 7	Day 15	Day 30	Day 60	Day 95		
Informed Consent	✓	-	-	-	-	-	-	-
Demographics	✓	-	-	-	-	-	-	-
Medical History	✓	✓	✓	✓	✓	✓	✓	(✓)
Concomitant Medications	✓	✓	✓	✓	✓	✓	✓	(✓)
Inclusion/ Exclusion	✓	-	-	-	-	-	-	-
Habitual lens information (brand / manufacturer, modality, power, wear success, habitual lens care brand)	✓	-	-	-	-	-	-	-
[REDACTED]	█	█	█	█	█	█	█	█
[REDACTED]	█	█	█	█	█	█	█	█
[REDACTED]	█	█	█	█	█	█	█	█
[REDACTED]	█	█	█	█	█	█	█	█

Procedure/ Assessment	Visit 1 Screening/ Baseline/ Dispense	Visit 2 Week 1 Follow-up	Visit 3 Week 2 Follow-up	Visit 4 Month 1 Follow-up	Visit 5 Month 2 Follow-up	Visit 6 Month 3 Follow-up/Exit	Early Exit	USV
	Day 1	Day 7	Day 15	Day 30	Day 60	Day 95		
Biomicroscopy	✓	✓	✓	✓	✓	✓	✓	(✓)
Randomize	✓	-	-	-	-	-	-	-
IP Dispense	✓	✓	(✓)	(✓)	(✓)	-	-	(✓)
VA w/ study lenses (OD, OS Snellen distance) ¹	✓	✓	✓	✓	✓	✓	✓	(✓)

Printed By:

Print Date:

Procedure/ Assessment	Visit 1	Visit 2	Visit 3	Visit 4	Visit 5	Visit 6	Early Exit	USV
	Screening/ Baseline/ Dispense	Week 1 Follow-up	Week 2 Follow-up	Month 1 Follow-up	Month 2 Follow-up	Month 3 Follow-up/Exit		
	Day 1	Day 7	Day 15	Day 30	Day 60	Day 95		
1								
2								
3								
4								
5								
6								
7								
8								
9								
10								
11								
12								
13								
14								
15								
16								
17								
18								
19								
20								
21								
22								
23								
24								
25								
26								
27								
28								
29								
30								
31								
32								
33								
34								
35								
36								
37								
38								
39								
40								
41								
42								
43								
44								
45								
46								
47								
48								
49								
50								
51								
52								
53								
54								
55								
56								
57								
58								
59								
60								
61								
62								
63								
64								
65								
66								
67								
68								
69								
70								
71								
72								
73								
74								
75								
76								
77								
78								
79								
80								
81								
82								
83								
84								
85								
86								
87								
88								
89								
90								
91								
92								
93								
94								
95								
96								
97								
98								
99								
100								
101								
102								
103								
104								
105								
106								
107								
108								
109								
110								
111								
112								
113								
114								
115								
116								
117								
118								
119								
120								
121								
122								
123								
124								
125								
126								
127								
128								
129								
130								
131								
132								
133								
134								
135								
136								
137								
138								
139								
140								
141								
142								
143								
144								
145								
146								
147								
148								
149								
150								
151								
152								
153								
154								
155								
156								
157								
158								
159								
160								
161								
162								
163								
164								
165								
166								
167								
168								
169								
170								
171								
172								
173								
174								
175								
176								
177								
178								
179								
180								
181								
182								
183								
184								
185								
186								
187								
188								
189								
190								
191								
192								
193								
194								
195								
196								
197								
198								
199								
200								
201								
202								
203								
204								
205								
206								
207								
208								
209								
210								
211								
212								
213								
214								
215								
216								
217								
218								
219								
220								
221								
222								
223								
224								
225								
226								
227								
228								
229								
230								
231								
232								
233								
234								
235								
236								
237								
238								
239								
240								
241								
242								
243								
244								
245								
246								
247								
248								
249								
250								
251								
252								
253								
254								
255								
256								
257								
258								
259								
260								
261								
262								
263								
264								
265								
266								
267								
268								

5 INTRODUCTION

5.1 Rationale and Background

In this clinical study, the performance of the investigational [REDACTED] contact lens will be compared to the commercially available Biofinity contact lens in a parallel group design to demonstrate the safety and effectiveness of the [REDACTED] contact lens, worn daily [REDACTED]

[REDACTED] as compared to Biofinity contact lens, worn daily [REDACTED]
[REDACTED] The intended use of this contact lens is for vision correction. Therefore, the measurement of distance VA is planned as the primary effectiveness variable.

[REDACTED]

[REDACTED]

[REDACTED]

5.2 Purpose of the Study

The purpose of this clinical study is to demonstrate the safety and performance of the investigational [REDACTED] soft contact lens compared to the commercially available Biofinity soft contact lens when worn in a daily wear modality, by assessing visual acuity as the primary variable.

At the end of the study, a clinical study report will be prepared in accordance with applicable regulatory requirements and standards.

[REDACTED]

[REDACTED]

5.3 Risks and Benefits

Contact lenses may offer improved peripheral vision and the convenience of not wearing spectacles. Material properties and design characteristics of [REDACTED] contact lenses are features consistent with successful contact lens wear. Based upon nonclinical testing and documented rationale for applicability of test results, [REDACTED] contact lenses are assessed to be non-toxic and biocompatible for on-eye use.

In the US, Biofinity contact lenses have approved indications for use for both daily wear and extended wear for up to 6 continuous nights. Further details on any known potential risks and benefits can be found in the product package insert.

A summary of the known potential risks and benefits associated with [REDACTED] contact lenses can be found in the IB. [REDACTED]

[REDACTED]

[REDACTED]

In

[REDACTED]

[REDACTED]

[REDACTED]



Refer to the IB for additional information.

6 STUDY OBJECTIVES

6.1 Primary Objective(s)

Table 6–1 Primary Objective(s)

<u>Objective(s)</u>	<u>Endpoint(s)</u>
Demonstrate safety and effectiveness of [REDACTED] soft contact lens when worn for daily wear as compared to Biofinity soft contact lens.	Primary Effectiveness <ul style="list-style-type: none">• Distance VA (Snellen)

6.2 Secondary Objective(s)

Not Applicable



6.4 Safety Objective(s)

Table 6–2 Safety Objective(s)

<u>Objective(s)</u>	<u>Endpoint(s)</u>
Duty of care and evaluation of safety profile of the investigational products.	AEs Biomicroscopy findings Device deficiencies

7 INVESTIGATIONAL PLAN

7.1 Study Design

This will be a prospective, randomized, [REDACTED] controlled, double-masked, parallel-group, daily wear clinical study.

This clinical study will engage approximately 8 clinic sites to enroll approximately 120 subjects with approximately 15 subjects per site.

Subjects will be expected to attend 6 office visits: Screening / Baseline / Dispense , Week 1 Follow-up, Week 2 Follow-up, Month 1 Follow-up, Month 2 Follow-up, and Month 3 Follow up/Exit. The total expected duration of participation for each subject is approximately 3 months in this daily wear study. Subjects will be randomized to wear either the test [REDACTED] contact lenses in both eyes or the control Biofinity contact lenses in both eyes.


Following randomization, study lenses will be dispensed to the subject [REDACTED]

[REDACTED]
[REDACTED]
[REDACTED]

Subjects will wear the lenses while awake in a daily wear modality. CLEAR CARE will be used for daily cleaning and disinfection. [REDACTED]

[REDACTED]
[REDACTED]

Figure 7-1 **Flowchart of Study Visits**



7.4 Rationale for Choice of Control Product

The Biofinity contact lens was chosen as the control product [REDACTED] to compare to [REDACTED] contact lens with regard to effectiveness and safety. Both the [REDACTED] contact lens and Biofinity contact lens are frequent replacement SiHy lenses and are to be prescribed for daily wear. The Biofinity contact lenses are indicated for the optical correction of refractive ametropia (myopia and hyperopia) in phakic or aphakic

persons with non-diseased eyes. [REDACTED]

[REDACTED]

7.5 Data Monitoring Committee

Not Applicable

8 STUDY POPULATION

The study population consists of adult male or female subjects (aged 18 or over), with non-diseased eyes, who require optical correction for refractive ametropia. [REDACTED]

[REDACTED]

The intended study population consists of volunteer subjects who are frequent replacement daily wear soft contact lens wearers, excluding Biofinity habitual wearers, who have at least 3 months of contact lens wearing experience, and who wear their habitual lenses at least 5 days per week and 8 hours per day. [REDACTED]

[REDACTED]

[REDACTED]

8.1 Inclusion Criteria

Written informed consent must be obtained before any study specific assessment is performed. Upon signing informed consent, the subject will be considered enrolled in the study.

Subjects eligible for inclusion in this study must fulfill **all** of the following criteria:

1. At least 18 years of age and be able to understand and sign an IRB/IEC approved Informed Consent form.
2. Willing and able to attend all scheduled study visits as required per protocol.
3. Successful wear of spherical daily wear frequent replacement soft contact lenses for distance correction in both eyes during the past 3 months for a minimum of 5 days per week and 8 hours per day.
4. Manifest cylinder ≤ 0.75 D in each eye.

8.2 Exclusion Criteria

Subjects fulfilling **any** of the following criteria are not eligible for participation in this study.

11. Wearing habitual contact lenses in an extended wear modality (routinely sleeping in lenses for at least 1 night per week) over the last 3 months prior to enrollment.

13. Monovision wearers.

16. Any habitual wear of Biofinity lenses.

Women of childbearing potential or women who are pregnant at the time of study entry are not excluded from participation. [REDACTED]

8.3 Rescreening of Subjects

Rescreening of subjects is not allowed in this study.

9 TREATMENTS ADMINISTERED

9.1 Investigational Product(s)

Test Product(s): [REDACTED] soft contact lenses

Control Product(s): [REDACTED] Biofinity (comfilcon A) soft contact lenses

Table 9–1 Test Product

Test Product	soft contact lenses (contact lens) (LID015385)
Manufacturer	Alcon Vision, LLC. 6201 South Freeway Fort Worth, Texas 76134-2099 USA
Indication for use and intended purpose in the current study	The intended use of this contact lens is for vision correction.
Product description and parameters available for this study	<ul style="list-style-type: none">■ [REDACTED]■ [REDACTED]■ [REDACTED]■ [REDACTED]■ [REDACTED] <ul style="list-style-type: none">■ [REDACTED]■ [REDACTED]
Usage	<ul style="list-style-type: none">• Wear:<ul style="list-style-type: none">○ Daily Wear<ul style="list-style-type: none">▪ During waking hours only.■ [REDACTED][REDACTED]○ Bilateral• Exposure: At least ~8 hours per day and ~5 days per week over a ~3 month exposure period. <ul style="list-style-type: none">■ [REDACTED]■ [REDACTED]


	<ul style="list-style-type: none">• Lens Care: CLEAR CARE
Packaging description	Blister foil pack

Table 9–2 **Control Product**

Control Product(s)	Biofinity (comfilcon A) soft contact lenses (Biofinity contact lens) (LID010221)
Manufacturer	CooperVision
Indication for Use	The intended use of this contact lens is for vision correction.
Product description and parameters available for this study	<ul style="list-style-type: none">• Material: comfilcon A• Water content: 48%• Power range: -1.00 to -6.00 D (0.25 D steps)• Base curve: 8.6 mm (target)• Diameter: 14.0 mm (target)
Formulation	Silicone Hydrogel. Additional details can be found in the Biofinity package insert.
Usage	<ul style="list-style-type: none">• Wear:<ul style="list-style-type: none">○ Daily Wear

- During waking hours only.
 - [REDACTED]
 - [REDACTED]
 - [REDACTED]
- Bilateral
- Exposure: At least ~8 hours per day and ~5 days per week over a ~3 month exposure period.
 - [REDACTED]
 - [REDACTED]
 - [REDACTED]
 - [REDACTED]
- Lens Care: CLEAR CARE [REDACTED]
 - [REDACTED]
 - [REDACTED]
 - [REDACTED]

Packaging description	Blister foil pack.
-----------------------	--------------------



More information on the test product can be found in the IB; information on the control product can be found in the Package Insert.



9.3 Treatment Assignment / Randomization

Subjects will be randomized in a 2:1 manner to receive either the [REDACTED] or Biofinity contact lenses, respectively. [REDACTED]
[REDACTED]

Only after signing the ICF, a subject will be assigned a subject number by the electronic data capture system.


[REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]



At Visit 1, all eligible subjects will be randomized [REDACTED] to one of the treatment arms. [REDACTED]
[REDACTED]
[REDACTED]
[REDACTED]

9.4 Treatment masking

This study is double-masked, with subjects randomized to use [REDACTED] or Biofinity contact lenses for the duration of the 3-month treatment period. [REDACTED]
[REDACTED]

[REDACTED]





9.5 Accountability Procedures



It is the Investigator's responsibility to ensure that:

- All study products are accounted for and not used in any unauthorized manner
- [REDACTED]

- All unused products are available for return to the Study Sponsor, as directed
- Any study lenses associated with a device deficiency or with any product-related AE (ie, ADE or SADE) are returned to the Study Sponsor for investigation, unless otherwise directed by the Sponsor. Refer to Section 11 of this protocol for additional information on the reporting of device deficiencies and AEs [REDACTED]
[REDACTED]

The Investigator is responsible for proper disposition of all unused IPs at the conclusion of the study, [REDACTED]

9.6 Changes to concomitant medications, treatments/ procedures

After the subject is enrolled into the study, the Investigator must instruct the subject to notify the study site about:

- Any new medications
- Alterations in dose or dose schedules for current medications,
- Any medical procedure or hospitalization that occurred or is planned
- Any non-drug therapies (including physical therapy and blood transfusions).

The Investigator must document this information in the subject's case history source documents.

10 STUDY PROCEDURES AND ASSESSMENTS

Subjects will be expected to attend 6 office visits, as shown below.

Visit #	Visit Type	Visit Day	[REDACTED]
Visit 1	Screening/Baseline/Randomization/Dispense	Day 1	[REDACTED]
Visit 2	Week 1 Follow-up Visit	Day 7	[REDACTED]
Visit 3	Week 2 Follow-up Visit	Day 15	[REDACTED]
Visit 4	Month 1 Follow-up Visit	Day 30	[REDACTED]
Visit 5	Month 2 Follow-up Visit	Day 60	[REDACTED]
Visit 6	Month 3 Follow-up / Exit Visit	Day 95	[REDACTED]

Country	Percentage (%)
Argentina	21.0
Australia	22.0
Austria	23.0
Belgium	23.0
Brazil	21.0
Canada	22.0
Chile	21.0
China	21.0
Costa Rica	21.0
Czech Republic	23.0
Denmark	23.0
Finland	23.0
France	23.0
Germany	23.0
Greece	21.0
Hungary	21.0
Ireland	21.0
Italy	21.0
Japan	21.0
Mexico	21.0
Netherlands	23.0
New Zealand	23.0
Norway	23.0
Portugal	21.0
Spain	21.0
Sweden	23.0
Switzerland	23.0
United Kingdom	23.0
United States	21.0

At the Screening / Baseline / Dispense Visit, study lenses will be dispensed to the subject. All subjects will wear the study lenses in a daily wear modality, only during waking hours.

Term	Percentage
GMOs	85%
Organic	80%
Natural	75%
Artificial	65%

10.1 Informed Consent and Screening

The Investigator or delegate must explain the purpose and nature of the study, and have the subject read, sign, and date the IRB/IEC-approved informed consent document. The subject must sign the ICF BEFORE any study-specific procedures or assessments can be performed, including study-specific screening procedures. Additionally, have the individual obtaining consent from the subject and a witness, if applicable, sign and date the informed consent document.

The Investigator or delegate must provide a copy of the signed document to the subject and place the original signed document in the subject's chart, or provide documentation as required by local regulations.

10.2 Description of Study Procedures and Assessments

[REDACTED] The Investigator is responsible for ensuring responsibilities for all procedures and assessments are delegated to appropriately qualified site personnel.

10.2.1 Demographics

Obtain demographic information including age, race, ethnicity, and sex.

10.2.2 Medical History and Concomitant Medications

Collect medical history information, including information on all medications used within the past 30 days. Include herbal therapies, vitamins, and all over-the-counter as well as prescription medications. Throughout the subject's participation, obtain information on any changes in medical health and/or the use of concomitant medications.

10.2.3 Investigational Product Compliance

Review subject compliance with the study lenses usage and adjunct product usage [REDACTED]

10.2.4 Adverse Event Collection: Safety Assessment

Assess and record any AEs that are observed or reported, including those associated with changes in concomitant medication dosing since the previous visit. Requirements for reporting AEs in the study can be found in Section 11.

10.2.5 Slit-Lamp Biomicroscopy: Safety Assessment

Slit-lamp examination of the cornea, iris/anterior chamber, and lens must be performed in both eyes before instillation of any diagnostic eye drops.

10.2.6 Device Deficiencies: Safety Assessment

Assess and record any device deficiencies that are reported or observed since the previous visit. Requirements for reporting device deficiencies in the study can be found in Section 11.



10.5 Discontinued Subjects

10.5.1 Screen Failures

Screen failures are subjects who were excluded from the study after signing the informed consent, not meeting the inclusion/exclusion criteria, and prior to randomization to product/dispense of study product.

The Investigator must document the reason for screen failure in the subject's case history source documents.

Subject numbers must not be re-used.

10.5.2 Discontinuations

Discontinued subjects are individuals who voluntarily withdraw or are withdrawn from the study by the Investigator after signing the informed consent.

Subject numbers of discontinued subjects must not be re-used.

Subjects may discontinue from study or study treatment at any time for any reason. Subjects may also be discontinued from study treatment at any time if, in the opinion of the Investigator, continued treatment poses a risk to their health.

Topic	Percentage
The concept of a 'smart city'	95
Smart city technologies	85
Smart city infrastructure	80
Smart city governance	40
Smart city data	75
Smart city planning	90
Smart city implementation	95

10.6 Clinical Study Termination

The Study Sponsor reserves the right to close the investigational site or terminate the study in its entirety at any time.

If the clinical study is prematurely terminated or suspended by the Study Sponsor:

- The Study Sponsor must:
 - Immediately notify the Investigator(s) and subsequently provide instructions for study termination.
 - Inform the Investigator and the regulatory authorities of the termination/suspension and the reason(s) for the termination/suspension.
- The Investigator must:
 - Promptly notify the IRB/IEC of the termination or suspension and of the reasons.
 - Provide subjects with recommendations for post-study treatment options as needed.

The Investigator may terminate the site's participation in the study for reasonable cause.

10.6.1 Follow-up of subjects after study participation has ended

Following this study, the subject will return to their eye care professional for their routine eye care.

11 ADVERSE EVENTS AND DEVICE DEFICIENCIES

11.1 General Information

An AE is any untoward medical occurrence, unintended disease or injury, or untoward clinical signs (including abnormal laboratory findings) in subjects, users, or other persons, whether or not related to the investigational medical device (test article). Refer to the Glossary of Terms and figures below for categories of AEs and SAEs.

Figure 11-1 Categorization of All AEs





Figure 11-2 Categorization of All SAE



SAEs

In addition to reporting all AEs (serious and non-serious) meeting the definitions, the Investigator must report any occurrence of the following as an SAE:

- An ocular infection including a presumed infectious ulcer with any of the following characteristics*:
 - Central or paracentral location
 - Penetration of Bowman's membrane
 - Infiltrates > 2 mm diameter
 - Iritis
 - Increase in intraocular pressure
 - Culture positive for microorganisms
 - Increasing size or severity at subsequent visits
- Any central or paracentral corneal event (such as neovascularization) that results in permanent opacification

- Hypopyon
- Hyphema
- Neovascularization within the central 6 mm of the cornea
- Permanent vision loss as defined by loss of 2 or more lines of BCVA from enrollment visit that fails to resolve
- Uveitis (anterior, intermediate, or posterior)
- Corneal abrasion affecting \geq 50% of corneal surface area

Significant Non-Serious AE

A significant non-serious AE is a device-related, non-sight threatening AE that warrants discontinuation of any contact lens wear for greater than or equal to 2 weeks. In addition, the Investigator must report any occurrence of the following as a Significant Non-Serious Adverse Event:

- Peripheral non-progressive non-infectious ulcers
- All symptomatic corneal infiltrative events
- Corneal staining score greater than or equal to Grade 3
- Temporary vision loss as defined by loss of 2 or more lines of BCVA from enrollment visit that persists for 2 or more weeks
- Neovascularization score greater than or equal to Grade 2

Device Deficiencies

A device deficiency is inadequacy of a medical device with respect to its identity, quality, durability, reliability, safety, or performance. A device deficiency may or may not be associated with subject harm (ie, ADE or SADE); however, not all ADEs or SADEs are due

to a device deficiency. The Investigator should determine the applicable category listed in the Device Deficiency eCRF for the identified or suspect device deficiency and report any subject harm separately. Examples of device deficiencies include the following:

- Failure to meet product specifications (eg, incorrect lens power/diameter/base curve/color)
- Lens cloudy
- Lens surface/edge defect
- Torn lens during handling/in pack
- Packaging deficit (eg, mislabeled product)
- Suspect product contamination
- Lack of performance

11.2 Monitoring for AEs

At each visit, after the subject has had the opportunity to spontaneously mention any problems, the Investigator should inquire about AEs by asking the standard questions:

- “Have you had any health problems since your last study visit?”
- “Have there been any changes in the medicines you take since your last study visit?”




11.3 Procedures for Recording and Reporting

AEs are collected from the time of informed consent. Any pre-existing medical conditions or signs/symptoms present in a subject prior to the start of the study (ie, before informed consent is signed) are not considered AEs in the study and should be recorded in the Medical History section of the eCRF.



- ADEs or SAEs are documented on the *Serious Adverse Event and Adverse Device Effect* eCRF within 24 hours of the Investigator's or site's awareness.
- Device deficiencies are documented on the *Device Deficiency* eCRF within 24 hours of the Investigator's or site's awareness.
- A printed copy of the completed *Serious Adverse Event and Adverse Device Effect* and/or *Device Deficiency* eCRF must be included with product returns.
- Additional relevant information after initial reporting must be entered into the eCRF as soon as the data become available.
- Document any changes to concomitant medications on the appropriate eCRFs.
- Document all relevant information from Discharge Summary, Autopsy Report, Certificate of Death, etc, if applicable, in narrative section of the *Serious Adverse Event and Adverse Device Effect* eCRF.



Intensity and Causality Assessments

Where appropriate, the Investigator must assess the intensity (severity) of the AE based upon medical judgment with consideration of any subjective symptom(s), as defined below:

Intensity (Severity)

Mild An AE is mild if the subject is aware of but can easily tolerate the sign or symptom.

Moderate An AE is moderate if the sign or symptom results in discomfort significant enough to cause interference with the subject's usual activities.


Severe An AE is severe if the sign or symptom is incapacitating and results in the subject's inability to work or engage in their usual activities.

For every AE in the study, the Investigator must assess the causality (Related or Not Related to the medical device or study procedure). An assessment of causality will also be performed by Study Sponsor utilizing the same definitions, as shown below:

Causality

Related An AE classified as related may be either definitely related or possibly related where a direct cause and effect relationship with the medical device or study procedure has not been demonstrated, but there is a reasonable possibility that the AE was caused by the medical device or study procedure.

Not Related An AE classified as not related may either be definitely unrelated or simply unlikely to be related (ie, there are other more likely causes for the AE).



11.5 Unmasking of the Study Treatment

Masked information on the identity of the assigned medical device should not be disclosed during the study. [REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

11.6 Follow-Up of Subjects with AEs

The Investigator is responsible for adequate and safe medical care of subjects during the study and for ensuring that appropriate medical care and relevant follow-up procedures are maintained after the study.

[REDACTED]

11.7 Pregnancy in the Clinical Study

Women of childbearing potential or women who are pregnant at the time of study entry are not excluded from participation. [REDACTED]

[REDACTED]

[REDACTED]

[REDACTED]

12 ANALYSIS PLAN

Any deviations to the analysis plan will be updated during the course of the study as part of a protocol amendment or will be detailed in the clinical study report.

All analyses will be conducted according to the applicable statistical analysis plan.

12.1 Subject Evaluability

Final subject evaluability must be determined prior to breaking the code for masked treatment (lens) assignment and locking the database, based upon the Deviations and Evaluability Plan.

12.2 Analysis Sets

Five analysis sets will be defined:

- a) All Enrolled – all subjects signing the informed consent form
- b) Enrolled Dispensed – subjects/eyes from All Enrolled that have been exposed to study lenses [REDACTED]
- c) Enrolled Not Dispensed – subjects/eyes from All Enrolled that have not been exposed to study lenses [REDACTED]
- d) Completed – Enrolled Dispensed subjects/eyes completing the study
- e) Discontinued – Enrolled Dispensed subjects/eyes not complete the study

12.3 Demographic and Baseline Characteristics

Demographic information, [REDACTED] will be presented by lens group and overall [REDACTED]
[REDACTED]
[REDACTED]

12.4 Effectiveness Analyses

For [REDACTED] the primary [REDACTED], separate summary tables will be prepared, when applicable, for the Completed and the Discontinued analysis sets as follows:

- Completed Control (eyes/subjects)
- Completed Test (eyes/subjects)

- Discontinued Control (eyes/subjects)
- Discontinued Test (eyes/subjects)

No inferential testing will be performed on effectiveness endpoints,

12.4.1 Analysis of Primary Effectiveness Endpoint(s)


The primary objective of this study is to demonstrate safety and effectiveness of [REDACTED] soft contact lens when worn for daily wear as compared to Biofinity soft contact lens.

The primary effectiveness endpoint is distance VA with study lenses, collected in Snellen, for each eye. Conversion will be made to the logMAR scale.

12.4.1.1 Statistical Hypotheses

No hypothesis testing of the primary effectiveness endpoint is planned.

12.4.1.2 Analysis Methods

Summary statistics will be provided. 

Topic	Percentage
The Internet	95
Smartphones	92
Cloud Computing	88
Big Data	85
Machine Learning	82
Blockchain	78
Artificial Intelligence	75
Quantum Computing	72
The Internet	95
Smartphones	92
Cloud Computing	88
Big Data	85
Machine Learning	82
Blockchain	78
Artificial Intelligence	75
Quantum Computing	72
The Internet	95
Smartphones	92
Cloud Computing	88
Big Data	85
Machine Learning	82
Blockchain	78
Artificial Intelligence	75
Quantum Computing	72
The Internet	95
Smartphones	92
Cloud Computing	88
Big Data	85
Machine Learning	82
Blockchain	78
Artificial Intelligence	75
Quantum Computing	72



12.5 Handling of Missing Data

All data obtained in evaluable subjects/eyes will be used. No imputation for missing values will be carried out.

Incidence and reasons for discontinuation by lens group will be tabulated at each visit and overall.

12.6 Safety Analyses

The safety endpoints are:

- AEs
- Biomicroscopy Findings
- Device Deficiencies

There are no safety hypotheses planned in this study. The focus of the safety analysis will be a comprehensive descriptive assessment of occurrence of AE as well as the other listed parameters.

Descriptive summaries (counts and percentages) for ocular and nonocular AEs will be presented by Medical Dictionary for Regulatory Activities Preferred Terms, for Completed and Discontinued sets. A listing containing details of the AEs will also be provided. [REDACTED]

Term	Percentage
GMOs	100%
Organic	~85%
Natural	~75%
Artificial	~55%

Term	Percentage
GMOs	~95%
Organic	~98%
Natural	~98%
Artificial	~75%
Organic	~98%
Natural	~98%
Artificial	~98%
Organic	~98%
Natural	~98%
Artificial	~98%

13 DATA HANDLING AND ADMINISTRATIVE REQUIREMENTS

13.1 Subject Confidentiality


The Investigator must ensure that the subject's anonymity is maintained throughout the course of the study. In particular, the Investigator must keep an enrollment log with confidential identifying information that corresponds to the subject numbers and initials of each study participant. [REDACTED]

Education Level	Percentage
Less than primary	35.0
Primary	32.0
Secondary	25.0
Postsecondary	18.0



13.2 Completion of Source Documents and Case Report Forms

The nature and location of all source documents will be identified to ensure that original data required to complete the CRFs exist and are accessible for verification by the site monitor, and all discrepancies shall be appropriately documented via the query resolution process. Site monitors are appointed by the Study Sponsor and are independent of study site staff.



13.3 Data Review and Clarifications

A review of CRF data to the subject's source data will be completed by the site monitor to ensure completeness and accuracy. After the CRFs have been completed, additional data clarifications and/or additions may be needed as a result of the data cleaning process. Data clarifications are documented and are part of each subject's CRF.

13.4 Sponsor and Monitoring Responsibilities

The Study Sponsor will designate a monitor to conduct the appropriate site visits at the appropriate intervals according to the study monitoring plan. The clinical investigation will be monitored to ensure that the rights and well-being of the subjects are protected, the reported data are accurate, complete, and verifiable from the source documents, and the study is conducted in compliance with the current approved protocol (and amendments[s], if applicable), with current GCP, and with applicable regulatory requirements.

The site may not screen subjects or perform the informed consent process on any subject until it receives a notification from an appropriate Study Sponsor representative that the site may commence conducting study activities. Monitoring will be conducted periodically while the clinical study is ongoing. Monitoring methods may include site visits, telephone, written and fax correspondence. Close-out visits will take place after the last visit of the last subject at the site.

A Coordinating Investigator may be identified by the Study Sponsor to review and endorse the final study report. In cases where a Coordinating Investigator is engaged, the Study Sponsor will select the Coordinating Investigator based upon their experience, qualifications, active study participation, and their willingness and availability to take on this role.

13.5 Regulatory Documentation and Records Retention

The Investigator is required to maintain up-to-date, complete regulatory documentation as indicated by the Study Sponsor and the Investigator's files will be reviewed as part of the ongoing study monitoring.

13.6 Quality Assurance and Quality Control

The Study Sponsor will secure agreement from all involved parties to ensure direct access to all study related sites, source data and documents, and reports for the purpose of monitoring and auditing by the Study Sponsor, and inspection by domestic and foreign regulatory authorities. Quality control will be applied to each stage of data handling to ensure that all data are reliable and have been processed correctly. Agreements made by the Study Sponsor with the Investigator/Institution and any other parties involved in the clinical study will be provided in writing as part of the protocol or as a separate agreement.

14 ETHICS

The Investigator must ensure that all personnel involved in the conduct of the study are qualified to perform their assigned responsibilities through relevant education, training, and experience. The Investigator and all clinical study staff must conduct the clinical study in compliance with the protocol. Deviations from this protocol, regulatory requirements and/or GCP must be recorded and reported to the Sponsor prior to database lock. If needed, corrective and preventive action should be identified, implemented, and documented within the study records. Use of waivers to deviate from the clinical protocol is prohibited.

Before clinical study initiation, this protocol, the informed consent form, any other written information given to subjects, and any advertisements planned for subject recruitment must be approved by an IRB/IEC.

Voluntary informed consent must be obtained in writing from every subject and the process shall be documented before any procedure specific to the clinical investigation is applied to the subject. The Investigator must have a defined process for obtaining consent. [REDACTED]

15 REFERENCES

15.1 References applicable for all clinical studies

- ISO 11980:2012 Ophthalmic optics - Contact lenses and contact lens care products - Guidance for clinical investigations
- ISO 14155:2011 Clinical investigation of medical devices for human subjects - Good clinical practice

15.1.1 US references applicable for clinical studies

- 21 CFR Part 11 - Electronic Records; Electronic Signatures

- 21 CFR Part 50 - Protection of Human Subjects
- 21 CFR Part 56 - Institutional Review Boards
- 21 CFR Part 812 - Investigational Device Exemptions
- 21 CFR Part 54 - Financial Disclosure by Clinical Investigators
- The California Bill of Rights